A-A-53176
28 August 1987

COMMERCIAL ITEM DESCRIPTION

STOCKINET, SURGICAL (DRAPE)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers sterile, disposable surgical stockinets providing a sterile barrier between the patient's non-sterile extremity and the sterile gloves of the surgeon or nurse to permit aseptic manipulation of the extremity.

Salient Characteristics:

Shall be medium or large size sterile, disposable, surgical waterproof cover stockinet as specified, specifically designed for draping extremities for surgical procedures. The waterproof cover stockinet shall be sealed at one end.

The overall dimensions for the medium size shall be 48 ± 2 inches by 9 ± 1/2 inch and 48 ± 2 inches by 12 ± 1/2 inch for the large size.

The outer covering of the stockinet shall be blue-gray and made from a highly stretchable thermoplastic rubber film, pin hole free, that provides protection against bacterial migration (air-borne or liquid) during the surgical procedure. The stockinet shall be impervious to blood and fluids.

The inner tube liner of the stockinet shall be natural or bleached white and shall be made from a highly stretchable tubular knit fabric resistant to limiting and pilling. The inner tube liner fabric shall be made of 100 percent polyester, or an an alternate shall be made of a cotton/polyester blend.

The assembled structure shall be folded in a telescopic configuration or shall be rolled in a cuff configuration from the inside out, starting with the open end and ending near the closed end. Both configurations shall facilitate the application of the stockinet to the extremity.

Each stockinet shall be overwrapped in a white nonwoven fabric. The nonwoven fabric shall be soft and have minimal memory for ease in opening.

AMSC/NA FSC 6510

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

The individually folded and wrapped stockinet shall be packaged in a white kraft heat sealable bag which is tear resistant and which will allow aseptic handling and maintain sterility of the contents until opened or damaged. Three individual stockinets shall be further bagged in an exterior dust covering.

The finished product (film tube and stockinet liner) shall have a maximum surface resistivity of 1.0 x 10" OHMs, or a maximum time of 0.5 seconds for 90 percent of induced charge to decay.

Surface resistivity or static decay test. Testing shall be conducted as described in NFPA 99 (National Fire Codes), a copy of which is available from the National Fire Protection Association, International, 470 Atlantic Avenue, Boston, MA 02210.

Test for Pin Holes. Cut stockinet lengthwise and remove stockinet liner. Make a solution of 0.60 grams of powdered red fabric dye (such as RIT) in 30 ml of water and add 20 ml of 100 percent isopropyl alcohol. Wet out a 3 foot length of the thermoplastic rubber film. After the solution dries, look on the other side of film for evidence of pin holes. There shall be no pin holes.

Workmanship. The Stockinet shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG).

For Size Medium, one package containing 18 stockinets, as specified, constitutes one unit.

For Size Large, one package containing 15 stockinets, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Testing and examination shall be conducted to determine compliance with all specification requirements. Test method shall be suitable, accurate and reproducible.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for examination shall be one stockinet. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5 maximum percent defective.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6510-01-004-6623	Stockinet,Surgical, Stretchable, Rubber Coated Bacterial Barrier, Medium, 9 by 48 inches, Disposable, 18s
6510-01-053-6259	Stockinet,Surgical, Stretchable, Rubber Coated Bacterial Barrier, Large, 12 by 48 inches, Disposable, 15s

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 06082

Project No. 6510-0717

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any p. tion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4986 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:16Z
Check the source to verify that this is the currentversion before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53176	STOCKINET, SURGICAL (DRAPE)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426

PREVIOUS EDITION IS OBSOLETE.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:16Z
Check the source to verify that this is the current version before use.