A-A-53835
28 February 1989

COMMERCIAL ITEM DESCRIPTION

TEST SET, AUDIOMETER CALIBRATION, PORTABLE

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a test set suitable for testing and calibrating air conduction audiometers.

Salient characteristics.

The unit shall be a portable test and calibration set suitable for testing and calibrating audiometers to ANSI S3.6, Specifications for Audiometers. The unit shall be capable of measuring test tone frequencies of 125, 250, 500, 750, 1000, 1500, 2000, 3000, 4000, 6000, and 8000 Hz.

The unit shall operate on a nominal 115 volt, 60 Hz power source.

The unit shall employ a standard artificial ear coupler with a type "L" standard condenser microphone to sense acoustic output of the audiometer earphone. The micropnone output shall be amplified, filtered by a 1/3 octave filter and indicated on a meter as a deviation from the prescribed value.

The meter shall be calibrated from -10dB to +6dB with zero at the mid-scale. Amplifier gain shall be capable of adjustment in 10dB increments to correspond to audiometer hearing level setting. The unit shall be capable of measuring distortion in any 10 dB increment of hearing level for frequencies of up to 4 KHz. The unit shall also be capable of displaying 2nd harmonic distortion.

A 4 digit frequency counter shall indicate test tone frequency to within 1 Hz. An oscilloscope connector shall be provided for measuring the audiometer decay rate and for measurement of transients.

The unit shall be capable of being calibrated using a built-in reference oscillator.

AMSC/NA FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

The unit shall be contained in a rigid carrying case. Overall dimensions, including case, shall be approximately 8 inches high by 14 inches deep by 15 inches wide.

The unit shall be provided with instructions for operation and repair. Instructions shall include schematics, wiring diagrams and repair parts lists.

Workmanship. The test set shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One test set with carrying case, as specified, constitutes one unit.

Quality Assurance Provisions

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defines as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests (dimensional), shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of II. The unit of product for test purposes shall be one test set.

Tests. Testing shall include but not be limited to the following:

Using an audiometer of known calibration and accuracy, verify the following:

Amplifier gain is capable of zero adjustment.

Test set is capable of being calibrated using internal reference oscillator.

Test set is capable of measuring frequency and harmonic distortion for the test tone frequencies and accuracy specified.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting of rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-111-7338	TEST SET, AUDIOMETER, CALIBRATION, Portable, 115 volt, 60 Hz, AC.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4485

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53835	TEST SET, AUDIOMETER CALIBRATION, PORTABLE

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street City State ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *Last First MI* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *Street City State ZIP Code* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, is too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4968 WASHINGTON D.C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.