A-A-53837
30 May 1989

COMMERCIAL ITEM DESCRIPTION

ELECTROSURGICAL APPARATUS

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a lightweight, portable, AC, solid state electrosurgical apparatus.

Salient characteristics:

The electrosurgical apparatus shall be compact, solid state, lightweight and shall not exceed 45 pounds.

The unit shall be portable, suitable for performing urological, and general surgical procedures.

The outside dimensions of the unit shall not exceed 18 inches by 14 inches by 7 inches.

The unit shall be equipped with an appropriately rated three-conductor line cord terminated in accordance with Underwriters' Laboratories, Inc. (UL).

The unit shall be equipped with approved hospital-grade 3-prong plug.

The unit shall be suitable for operation on nominal 115 volt, 50/60 Hz. electric power (230 volt, 50/60 Hz. when so specified in the procurement document). The output power for the unit shall not exceed 1500 VA for full operation.

The unit shall have a minimum of the following connections for accessories:

a. A connector or jack for the patient return electrode.

b. Connectors or jacks for simultaneous attachment of one standard type active electrode, one forceps type coagulating electrode with built-in control switch for selection of cut or coagulation modes.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

c. A connector or jack for attaching an explosion-proof or intrinsically safe foot switch to enable selection of the cut or coagulation modes.

d. A common return.

The unit shall be supplied with the following accessories:

a. A 1/16 inch ball electrode.
b. A flat cutting electrode.
c. A straight 1/2 inch blade
d. Two reusable active electrode chuck handles with ten foot cords.
e. A 45 degree angle 1/2 inch cutting electrode.
f. A straight needle point electrode.
g. An explosion-proof or intrinsically safe (low voltage current limited) foot switch.
h. A dummy load.
i. Two patient plates with cords.

The unit shall have a minimum of the following modes of operation with controls:

a. Cut mode: Cutting mode power shall be in range, which allows precise speed and depth of cutting effect.

b. Coagulation mode: Coagulation motor power shall be in range, which allows exact hemostatic effect with minimum tissue destruction and charring.

c. Modified cut mode: The modified cut mode shall cause the application of intermittent radio frequency (IRF) power to the active electrode to allow control of the degree of hemostasis.

Output power, as measured at the active electrode, shall remain constant within 10 percent, with typical 105-130 volt line voltage fluctuations.

The unit shall have isolated output circuits., i.e., output not referenced to ground.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

The unit shall generate minimal radio frequency interference (RFI), both radiated and conducted, through the power line. The active cable of the ESU shall be properly shielded to limit radiated losses and prevent capacitive coupling into nearby wires and metal equipment.

The unit shall be provided with a positive on/off switch. An indicator light shall signal that the ESU has been turned on. Separate visual indicators shall be incorporated to indicate when the ESU output has been activated. In addition, separate distinct audible tones shall indicate cut and coag output activation. An adjustable volume control which cannot reduce the sound level below 45 dB shall be provided.

Overcurrent protection for the ESU, preferably in the form of a circuit breaker, is required. If a fuse is used, a spare must be provided and clearly labeled as to function and rating. Circuit breakers must be protected from other objects impinging against it.

Unit shall meet NFPA 493, UL 544, and AAMI HF-18 Electrical Safety Standards. All items shall conform to the applicable requirements of nationally recognized standards for the safety of medical equipment of the issue in effect on the date of solicitation. All items shall be approved as conforming to such a standard by an independent, nationally recognized testing laboratory.

Minimum length of the cords and cables supplied shall be:

Cable Type	Length
Line Cord	3.0 meters (9.8 feet)
Active	2.5 meters (8.2 feet)
Dispersive	2.4 meters (7.9 feet)
Footswitch	3.0 meters (9.8 feet)

Workmanship. The electrosurgical apparatus shall be free from defects which detract from its appearance or impair its serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Unit. Each (ea). One electrosurgical apparatus along with accessories, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

First article inspection. First article inspection shall be performed on the electrosurgical unit when a first article sample is required. This inspection shall include examination and tests.

Quality conformance inspection. Quality conformance inspection shall include the examination and tests specified herein.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the item identification, the name of the component/material, the lot number and the lot size.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 with an AQL of one (percent defective) and an inspection level of S-2. The unit of product for sampling purposes shall be one electrosurgical apparatus.

DEFECTS

Unit not complete with all components, equipment items, and accessories.

Any component, equipment item or accessory broken.

Any switch and/or control position not identifiable when viewed at a front distance of 2.5 meters (8 feet) by a viewer with corrected 20/20 vision.

Any cooling fan exhausts air over a sterile field.

Footswitch does not have suitable mechanical protection against accidental depression.

Manual, as specified, not included.

Unit not clean.

Any rivet not properly installed, screw not tight.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of one (percent defective) and an inspection level of S-2. The unit of product for sampling purposes shall be one electrosurgical apparatus.

Tests. Tests shall be conducted to determine compliance with this specification.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:25Z
Check the source to verify that this is the current version before use.

a. First article. Where a first sample is required, all tests shall be performed on that unit.

b. Acceptance testing. When a first article sample has not been required, one unit shall be subjected to the performance test specified herein. If a first article sample has successfully passed all tests, performance test shall not be required at acceptance.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-00-137-6511

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4189

Location: CID/ESA/H16-30

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426

PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:25Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

Fold along this line

Fold along this line

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4968 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.