A-A-53842
24 March 1989

COMMERCIAL ITEM DESCRIPTION

TEST PATTERN AND PENETROMETER ASSEMBLY, X-RAY

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers an x-ray test pattern and penetrometer assembly suitable for testing the parameters for acceptance inspection, performance checking, and calibration of x-ray equipment.

Salient characteristics:

The x-ray test pattern and penetrometer assembly shall be complete with stand, interchangeable filters and assorted radiographic phantoms. The assembly shall be suitable for performing acceptance inspection, performance checking, and calibration of x-ray equipment.

As a minimum, the assembly shall consist of the following:

1. Two approximately 7 by 7 by 3/4 inch thick blocks of high purity aluminum capable of simulating a heavy set patient at 90 kVp.

2. One approximately 7 by 7 by 1/8 inch thick lead beam-stop plate.

3. One approximately 7 by 7 by 1/32 inch thick aluminum penetrometer plate with central pairs of holes of 1/16 inch diameter, 1/8 inch diameter, 3/16 inch diameter, and 1/4 inch diameter.

4. A suitable stand, or legs, to hold the test plates.

Each assembly shall be supplied with written instructions for use.

The assembly shall be free of oil, grease, dirt, or other contamination.

AMSC N/A FSC 6525

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Workmanship. The x-ray test pattern and penetrometer assembly shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One x-ray test pattern and penetrometer assembly, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one x-ray test pattern and penetrometer assembly as specified in this description.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Defects.

Assembly is not complete with all components.

Penetrometer plate does not have correct number or size of holes.

Written instructions for use are not provided.

Assembly is not free of oil, grease, dirt, or other contamination.

Sampling for tests. Sampling for tests (dimensional) shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-3. The unit of product for test purposes shall be one x-ray test pattern and penetrometer assembly.

a. First article. Where a first article sample is required, all tests shall be performed on that unit.

b. Acceptance testing. When a first article sample has not been required, one unit shall be subjected to the performance test specified herein. If a test article sample has successfully passed all tests, performance test shall not be required at acceptance.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6525-00-149-1598

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6525-1218

Location: CID/TEST/H13-32

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53842	TEST PATTERN AND PENETROMETER ASSEMBLY, X-RAY

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for u
ubmitting comments and suggestions for improvements. All users of military standardization documents are invited to pro
suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPL*
mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretat
too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the
problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an
acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being
considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of
specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorizati
to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POST
NECESS
IF MAIL
IN TH
UNITED ST

Source: https://assist.dla.mil -- Downloaded: 2026-10-25T09:25Z
Check the source to verify that this is the current version before use.