A-A-53847
24 March 1989

COMMERCIAL ITEM DESCRIPTION

SPECIMEN KIT, URINE

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-S-37498 which is cancelled.

This Commercial Item Description covers an integrated set of compatible, non-sterile components suitable for collecting, transporting, and testing of urine specimens.

Salient characteristics:

The urine specimen kit shall consist of the following integrated set of compatible non-sterile components for urine specimen collection, transport, and analysis:

500 - Clear plastic calibrated urine specimen tubes.
500 - Snap caps for tubes.
500 - Adhesive labels for tubes.
500 - Urine specimen collection cups.
4 minimum quantity - Holders for urine specimen tubes.

Urine Specimen Tube. The urine specimen tube shall be made of clear plastic and shall have a capacity of 15 plus or minus 1 milliliters (ml). The top of the tube shall be circular and it shall have a lip to accommodate the snap cap. The tube may terminate in a point (may be slightly rounded) at the bottom to facilitate the collection of centrifuged precipitate. The walls of the tube shall be straight, or shall taper inward toward the bottom. The tube shall have the following approximate dimensions:

Diameter of lip	- 7/8 inches
Outside diameter of tube below lip	- 5/8 inches
Length of tube	- 4-1/8 inches

The tube shall have graduated markings at the 1/4 ml, 1/2 ml or 1 ml level and the 12 ml level. Additional graduation marks shall be acceptable. The tube shall have no breaks, holes, cracks, or sharp edges.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Snap Cap. The cap shall be an easy-on, easy-off plastic snap cap. It may have an integral lift tab or rim. It shall snap on and off the lip of the tube without undue effort and shall fit securely over the opening of the tube. The cap shall have no breaks, tears, or sharp edges, and shall protect contents of the tube from dust, other contamination, and spillage. The tube and cap shall show no evidence of leakage when tested.

Label. The label shall be white and shall measure 1-3/4 plus or minus 1/4 inches by 1-1/4 plus or minus 1/4 inches. The label shall be imprinted with at least three lines for recording data. It shall accept and retain writing from a standard ball-point pen. The label shall be coated on one side with an all-purpose, permanent pressure-sensitive adhesive. The adhesive surface shall be protected by a strip of suitable removable backing material. The label shall peel freely from the backing and shall adhere firmly to the specimen tube. The label shall not lift or separate from the tube when tested.

Cup. The cup shall have a capacity of not less than 3 ounces and shall be fabricated of a material suitable for the collection of urine. The cup shall have a flat bottom so that it can stand on a level surface when filled, without rock or wobble. Cups may be constructed so that they can be nested to conserve storage space. The cup shall contain no tears or holes. The cup shall show no evidence of leakage when tested.

Tube Holder. The tube holder shall be fabricated of cardboard, plastic, or other suitable material, free of tears or other deformations. It shall have inserts, approximately 1-1/2 inches deep, or at least two levels of concentric holes, to hold 10 or more specimen tubes in an upright position. It shall be supplied fully assembled, or shall be capable of being assembled, without the use of tools, fasteners, or adhesive in three minutes or less. The holder shall show no evidence of spillage, deformation, or other failure when tested.

All components of the kit shall be free of dirt, oil, grease, or other contamination.

Workmanship. The urine specimen kit shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package consisting of one specimen kit, containing 500 tubes, 500 caps, 500 labels, 500 cups, and four holders as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:25Z
Check the source to verify that this is the currentversion before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one urine specimen kit as specified in this description.

Defects

Complete Kit

Any component is not provided.
Count is less than specified.
Marking of components and packages is not complete or not present.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Defects (continued)

Cap

Not free of grease, oil, and dirt.
Not free of sharp edges.
Not free of breaks and tears.
Does not fit securely on specimen tube.

Tube

Not free of grease, oil, and dirt.
Not free of sharp edges.
Not free of breaks, holes, and cracks.
Plastic cloudy; not clear.
Tube does not have specified graduation marks.
Tube is not free of waviness, water marks, excessive scratches, or surface cracks greater than 1/4 inch long.

Label

Backing material cannot be peeled freely or is otherwise inadequate.
Does not adhere firmly to tube.

Cup

Not free of grease, oil and dirt.
Cup is not free from tears or holes.
Cup crushed or deformed.

Holder

Does not support tubes in an upright position.
Not free of tears.
Not capable of being assembled.
Does not accept tubes.

Sampling for tests. Sampling for tests including dimensional test shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-3. The unit of product for test purposes shall be one urine specimen kit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

a. First article. Where a first article sample is required, all tests shall be performed on that unit.

b. Acceptance testing. When a first article sample has not been required, one unit shall be subjected to the performance test specified herein. If a test article sample has successfully passed all tests, performance test shall not be required at acceptance.

Centrifuge test. Tubes shall be filled with water to the 12 ml. graduation and placed in a centrifuge. The centrifuge shall be operated at 2200 revolutions per minute for a 10 minute period. The tubes shall then be removed from the centrifuge and examined for evidence of leakage or breakage. Any evidence of leakage or breakage shall constitute a defect.

Tube leakage test. Tubes shall be filled with water to the 12 ml. level. The caps shall be fastened on and all traces of water removed from the exterior of the tubes and caps. The tubes shall be held upside down for two minutes, and then placed lengthwise on a horizontal surface in a vacuum chamber and subjected to a negative pressure of 6 inches of mercury for one hour at normal room temperature (65° F to 90° F). At the end of the hour the tubes shall be examined for any evidence of leakage. Any evidence of leakage shall constitute a defect.

Cup leakage test. Cups shall be filled with 2-1/2 fluid ounces of water and left standing upon a level surface for 2 hours at room temperature. Cups shall then be examined for any evidence of leakage. Any evidence of leakage shall constitute a defect.

Tube holder test. Ten tubes shall be filled to the 12 ml. level and placed in the holder. Each tube shall be removed and replaced twice, and allowed to remain in the holder for 2 hours. After 2 hours, the tube holder shall be examined for any evidence of failure to support the tubes upright without spillage. Any evidence of failure shall constitute a defect.

Label adhesion. Specimen tubes shall be filled with 12 ml. of water at room temperature and capped. A label shall be applied firmly to each tube. The tubes shall be refrigerated at 40° F or less for a period of at least 10 hours. Tubes shall then be centrifuged at 2200 revolutions per minute for a 10 minute period. The tubes shall then be removed from the centrifuge and examined for any evidence of labels lifting from the tube surface. Any evidence of label lifting from the tube surface shall constitute a defect.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:25Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-00-075-6636

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-2008

Location: CID/SPECIMEN/CUP/H13-32

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER (Last First MI) – Optional	b WORK TELEPHONE NUMBER (Include Area Code) – Optional
c MAILING ADDRESS (Street City State ZIP Code) – Optional	8 DATE OF SUBMISSION (YYMMDD)

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53847	SPECIMEN KIT, URINE

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION (Mark one)
	☐ VENDOR
	☐ USER
b ADDRESS (Street City State ZIP Code)	☐ MANUFACTURER
	☐ OTHER (Specify)

5 PROBLEM AREAS

a Paragraph Number and Wording:

b Recommended Wording

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:25Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (DO NOT STAPLE) and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSIE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:25Z
Check the source to verify that this is the current version before use.