A-A-53852
24 March 1989

COMMERCIAL ITEM DESCRIPTION

CASSETTE, RADIOGRAPHIC FILM

Lightweight

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a 24 cm by 24 cm radiographic film cassette with two green-emitting "rare earth" intensifying screens.

Salient characteristics:

The radiographic film cassette shall be lightweight and suitable for use with medical x-ray film. The cassette shall be furnished with two green-emitting "rare earth" intensifying screens.

The cassette shall consist of two aluminum covers mounted in an aluminum, American Iron and Steel Institute (AISI) Series 300 corrosion-resisting steel, or polyurethane frame joined together on the long dimension by a suitable hinge, or hinges. The cassette tube side (front side) shall shall have a scratch resistant vinyl covering. On the side opposite the hinge, there shall be a suitable fail-safe, flush mounted, latching system.

The cassette shall be designed to insure the proper tightness and contact of the film against the screens. One cover shall have a foam backing to enhance this tightness. The back cover shall contain sufficient lead to preclude excess scattered radiation.

The back cover of the cassette shall have a window with a latched cover in order to perform identification procedures in daylight areas.

The intensifying screens shall be in accordance with the latest revision of ANSI Standard PH1.50, Dimensions for Radiographic Film Cassette Intensifying Screens. This commercial item description shall govern in event it deviates from the ANSI Standard.

AMSC N/A FSC 6525

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

The cassette shall be suitable for use with stock listed radiographic film, 24 cm by 24 cm in size.

Each cassette shall be furnished with appropriate instructions for use and identification stickers.

Other specifications for the medical radiographic film cassette shall conform to ANSI Standard PH1.49. This commercial item description shall govern in event it deviates from the ANSI Standard.

The cassette finish shall be free of chips, cracks, scratches, and other appearance defects.

The cassette shall be furnished clean and free of dirt, oil, and grease.

Workmanship. The radiographic film cassette shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One cassette, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one radiographic film cassette as specified in this description.

Defects.

Cassette is not complete with all components.

Cassette covers are not aluminum.

Cassette frame is not aluminum, AISI Series 300 corrosion-resisting steel, or polyurethane.

Cassette latching system is not flush mounted.

Cover does not have foam backing.

Cassette back cover does not contain sufficient lead to preclude excess scattered radiation.

Back cover of cassette does not have window with a latched cover.

Intensifying screens are not green-emitting "rare earth" type.

Intensifying screens are not correctly installed.

Cassette tube side does not have a scratch resistant vinyl covering.

Cassette door, hinges, and springlocks, do not operate as specified.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Defects. (continued)

Cassette is not furnished with appropriate instructions for use and identification stickers.

Cassette is not free of burrs and sharp edges.

Cassette finish is not free of chips, cracks, scratches, and other appearance defects.

Cassette is not clean and free of dirt, oil, and grease.

Sampling for tests. Sampling for tests including dimensional test shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one radiographic film cassette.

a. First article. Where a first article sample is required, all tests shall be performed on that unit.

b. Acceptance testing. When a first article sample has not been required, one unit shall be subjected to the performance test specified herein. If a test article sample has successfully passed all tests, performance test shall not be required at acceptance.

Tests.

Tests shall be conducted to determine compliance with specification requirements. Test radiographs shall be made at a focal spot to film distance of 40 plus or minus 1 inch, at 80 Kvp, with screen-type, medium speed x-ray film. The x-ray tube shall be perpendicular to the center of the cassette. The nominal focal spot size of the x-ray tube shall not be more than 1.0 mm. A minimum of 2.3 mm of aluminum filtration shall be in the useful beam. The exposed x-ray film shall be developed in accordance with the manufacturer's instructions. The developed x-ray film shall be viewed with an unaided eye using a fluorescent type x-ray film illuminator.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Drop test

The selected sample shall be dropped on each corner from a height of 30 inches onto a concrete or tile floor. The cassette shall be dropped so that the plane of the cassette is perpendicular to the floor and the center of gravity of the cassette is directly over the corner which strikes the floor. Except for deformation at the impacted corner area, the cassette shall show no other sign of damage. Any other kind of damage, will constitute a defect. After the drop test, the sample cassette shall be subjected to, and pass, the flatness test, lightproof test, and film contact test.

Flatness test

The test for flatness shall be conducted on the open and closed cassette. The cassette frame shall rest on a flat test surface and any height variation shall be measured by passing feeler gages between the flat surface and the frame. The deviation from flatness of the front panel shall be determined by measuring the maximum distance between the front face and a straight edge placed at various angles on the front panel. The damaged corners resulting from the drop test shall not be placed on the flat test surface, and shall not be measured. Both frame and front panel shall be flat within 1/16 inch. Any deviation will constitute a defect.

X-Ray opacity of the front panel test

A non-screen type radiographic film shall be exposed in a cardboard, or plastic film holder, through the front panel with an exposure time to produce a diffuse transmission density of 1.0 plus or minus 0.2. The developed film shall be examined for uniformity of opacity and for evidence of included foreign matter. Visible shadows or non-uniformity shall be cause for rejection.

Lightproof test

A high speed radiographic film, of the applicable size, shall be placed in the cassette, and all sides and edges of the cassette shall be exposed to a 100 watt incandescent lamp, operated at its rated voltage, at a distance of 3.3 feet (1 meter) from the cassette for a period of ten minutes. The developed test film shall be compared with a developed, unexposed film of the same lot. A density on the test film greater than that on unexposed film shall be cause for rejection. When viewed against the illuminator, the test film shall show no density variations, except that fog within 1/8 inch of film border or within 1/4 inch square at the film corners shall not be considered a defect.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Cassette liners test

A radiographic film in the sample cassette shall be exposed with an exposure time to produce a diffuse transmission density of 1.0 plus or minus 0.2. Shadows and artifacts on the test film shall be cause for rejection of the cassette.

Film contact

Contact shall be determined by using a flat sheet of non-magnetic, bare metal, .039 plus or minus .004 inch thick, continuously perforated with holes of .098 + 0, -02 inch diameter, at intervals of .16 + 0, -.02 inch. A flat screen of 1/4 inch mesh, made up of not more than 0.032 inch diameter, non-magnetic, bare, unsoldered wire may be used as an alternate. To facilitate the measurement of density of the film exposed during the test, the test object may have a clear area, 1/2 to 1 inch diameter, near its center. The test object shall be placed on the cassette using a cone which covers the film area but equally cuts the four film corners. Exposure shall be made to produce a diffuse transmission density of 2 to 3. Uniformity of film contact shall be indicated by a uniformly sharp exposure through the mesh. The test shall be repeated after exposure of the cassette to a minimum temperature of 120°F and a relative humidity of 80 plus or minus 2 percent for 48 hours. The intensifying screens may be removed during the conditioning period.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6525-01-282-5367

NOTE: The intensifying screen shall be Kodak Lanex Regular Screen, or equal.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6525-1249

Location: CID/CASSETTE/H13-21

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-53852	CASSETTE, RADIOGRAPHIC FILM, Lightweight

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, w too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.