A-A-53858
30 May 1989

# COMMERCIAL ITEM DESCRIPTION

## TURNING FRAME, ORTHOPEDIC BED

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for MIL-T-36310(DM) which is cancelled.

This Commercial Item Description covers an orthopedic bed turning frame suitable for use in orthopedic, neurosurgical, and gynecological services.

Salient characteristics:

The turning frame shall be for use where immobilization of the spine, with or without hyperextension, is required, and where frequent turning of the patient about the longitudinal axis of the patient's body is necessary for proper nursing care.

The turning frame shall have nominal overall dimensions of 97 inches (247 cm) in length, 30 inches (76 cm) in width, and 50 inches (127 cm) in height (standing), with a hydraulic height range of 25-1/2 to 37-1/2 inches (65 to 95 cm).

The turning frame shall be provided with anterior and posterior frames, turning assembly, and necessary covers to support a patient on the frames.

Foot control hydraulics shall activate the turning frame and include trendelenberg and reverse trendelenberg positions. The anterior and posterior frames shall not distort nor develop a permanent set, and the force required to rotate the frames shall not exceed 16 pounds lineal pressure when the frames are tested as specified herein.

The turning frame shall be supplied with two nylon forehead support bands, a 2 inch (5 cm) conductive pad (anterior frame) with a foam mattress (posterior frame), 8 inch (20.3 cm) conductive casters with 4-wheel brakes, adjustable side rails and arm boards, three gel adjustable face pieces, a sheet set, a plastic dust cover, and a storage tray.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Each turning frame shall have a metal nameplate securely attached and permanently marked with the manufacturer's name, contract or purchase order number, and the National Stock Number.

All welds shall be free of cracks, pits, porosity, and other defects. All welds shall be ground smooth.

The turning frame shall be free of sharp edges, burrs, pits, dents, cracks, ripples, and deformity. The frame shall be free of grease, oil, dirt, chips, or other foreign matter.

Each turning frame shall be supplied with a manual having assembly and operation instructions and a replacement parts list.

Workmanship. The turning frame shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One turning frame, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one turning frame, as specified in this description.

Defects.

Turning frame not as specified, incomplete.

Moving parts do not move freely.

Nameplate not supplied, incomplete, or illegible.

Metal parts not free of cracks, tears, ruptures or deformed surfaces.

Instruction manual missing or incomplete.

Plated finish not free of blisters, pits, peeling, or visible porosity.

Welds not continuous, not clean, not free of pitting, spatter and discoloration.

Exposed welds not ground smooth to blend with adjacent surfaces.

Welds not free of cracks; welds broken or not sound.

Exposed or accessible edges, corners, or surfaces not free of burrs, snarpness, and sharp projections.

Sampling for tests. Sampling for tests (dimensional) shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-3. The unit of product for test purposes shall be one turning frame.

a. First article. Where a first article sample is required, all tests shall be performed on that unit.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

b. Acceptance testing. When a first article sample has not been required, one unit shall be subjected to the performance test specified herein. If a test article sample has successfully passed all tests, performance test shall not be required at acceptance.

Tests.

With the posterior frame in its normal horizontal position, sand bags totaling 250 pounds shall be equally distributed and secured to the frame. After placing the anterior frame over the sand bags, the assembled turning frame shall be rotated through three complete cycles of all possible positions while testing the rotating force, after which the frame shall be examined for distortion. A lineal force exceeding 16 pounds or any evidence of distortion or permanent set shall be cause for rejection.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-01-252-5591

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

A-A-53858

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-2086

Location:
CID/TURNING/FRAME/H-11-33



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

**INSTRUCTIONS** In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

*NOTE* This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

*(Fold along this line)*

*(Fold along this line)*

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
PHILADELPHIA,PENNSYLVANIA 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

*(See Instructions – Reverse Side)*

| 1 DOCUMENT NUMBER | 2 DOCUMENT TITLE |
|---|---|
| A-A-53858 | TURNING FRAME, ORTHOPEDIC BED |

| 3a NAME OF SUBMITTING ORGANIZATION | 4 TYPE OF ORGANIZATION *(Mark one)* |
|---|---|
| | ☐ VENDOR |
| b ADDRESS *(Street City State ZIP Code)* | ☐ USER |
| | ☐ MANUFACTURER |
| | ☐ OTHER *(Specify)* |

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason Rationale for Recommendation

6 REMARKS

| 7a NAME OF SUBMITTER *(Last First, MI)* – Optional | b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional |
|---|---|
| c MAILING ADDRESS *(Street City, State, ZIP Code)* – Optional | 8 DATE OF SUBMISSION *(YYMMDD)* |

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.