INCH-POUND

MIL-DTL-52701C
17 November 2009
SUPERSEDING
MIL-I-52701B (ME)
12 March 1991

DETAIL SPECIFICATION

INDICATOR SOLUTION, pH DETERMINATION (RANGE 3.0 TO 11.0)

Inactive for new design after 28 May 1998.

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers an indicator solution for the determination of pH values of water in the range of 3.0 to 11.0.

2. APPLICABLE DOCUMENTS

2.1 General. The documents listed in this section are specified in sections 3, 4, or 5 of this specification. This section does not include documents cited in other sections of this specification or recommended for additional information or as examples. While every effort has been made to ensure the completeness of this list, document users are cautioned that they must meet all specified requirements of documents cited in sections 3, 4, or 5 of this specification, whether or not they are listed.

2.2 Government documents.

2.2.1 Specifications, standards, and handbooks. The following specifications, standards, and handbooks form a part of this document to the extent specified herein. Unless otherwise specified, the issues of these documents are those cited in the solicitation or contract.

Comments, suggestions, or questions on this document should be addressed to Defense Supply Center Richmond, ATTN: DSCR-VEB, 8000 Jefferson Davis Highway, Richmond, VA 23297-5616 or e-mailed to STDZNMGT@dla.mil. Since contact information can change, you may want to verify the currency of this address information using the ASSIST database at https://assist.daps.dla.mil.

AMSC N/A FSC 6810

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T21:30Z
Check the source to verify that this is the current version before use.

DEPARTMENT OF DEFENSE SPECIFICATIONS

MIL-D-537 - Disk, Color Standards

DEPARTMENT OF DEFENSE STANDARDS

MIL-STD-1916 - DoD Preferred Methods for Acceptance of Products

(Copies of these documents are available online at https://assist.daps.dla.mil/quicksearch/ or from the Standardization Document Order Desk, 700 Robbins Avenue, Building 4D, Philadelphia, PA 19111-5094.)

2.3 Non-Government publications. The following documents form a part of this document to the extent specified herein. Unless otherwise specified, the issues of these documents are those cited in the solicitation or contract.

AMERICAN PUBLIC HEALTH ASSOCIATION (APHA)

APHA SME - Standard Methods for the Exam of Water and Wastewater

(Application for copies should be addressed to American Public Health Association, 800 I Street NW, Washington, DC 20001 or on-line at http://www.apha.org.)

NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY

Research Paper 2619 - pH Values of the Clark and Lubs Buffer Solutions at 25ºC, by Vincent E. Bower and Roger G. Bates

(Research Paper 2619 was published in the Journal of Research of the National Bureau of Standards, Volume 55, No. 4, October 1955, pgs. 197-200. Copies of the paper may be obtained from National Institute of Standards and Technology, 100 Bureau Drive, Stop 2500, Gaithersburg, MD 20899-2500 or by e-mail from Library@nist.gov.)

2.4 Order of precedence. Unless otherwise noted herein or in the contract, in the event of a conflict between the text of this document and the references cited herein, the text of this document takes precedence. Nothing in this document, however, supersedes applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 Description. The indicator solution shall produce a color for specific pH values of water that visually match color standards of a master color disk. The solution shall be suitable for use in determining pH values of water from 3.0 to 11.0 in increments of 1 pH when tested in accordance with 4.2.

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T21:30Z
Check the source to verify that this is the current version before use.

3.2 Standard sample. A standard sample of indicator solution will be supplied by the Government on request by a potential bidder. The standard sample will be furnished only to show the potential bidder the color (obtained when 0.5 milliliter (mL) of standard sample is added to 15 mL of each buffer solution prepared in accordance with 4.2.2) that the indicator solution he proposes to furnish must produce (see 6.4).

3.3 Material. Material shall be as specified herein. Material not specified shall be selected by the contractor and shall be subject to all provisions of this specification.

3.3.1 Suspended matter and foreign material. Total suspended matter and foreign material shall not exceed 10 mg/L when tested per 4.2.1. When specified (see 6.2), the contractor may certify that the indicator solution has been filtered through Whatman grade 934AH and 984H, Gelmen type A/E, Millipore type AP40, or equivalent glass fiber filters prior to packaging.

3.4 Government-loaned property. Unless otherwise specified (see 6.2), the following property conforming to MIL-D-537, type I, class 8, in the quantities indicated will be loaned by the Government (see 6.3):

Item No.	Description	Quantity
1	Disk, master color standard, medium values, pH 3.0 to 11.0	1
2	Disk, master color standard, low values, pH 3.0 to 11.0	1
3	Disk, master color standard, high values, pH 3.0 to 11.0	1
4	Disk evaluator	1

4. VERIFICATION

4.1 Sampling. Sampling shall be accordance with MIL-STD-1916, verification level II or as specified in the acquisition document (see 6.2).

4.2 Tests.

4.2.1 Suspended matter foreign material. Unless otherwise specified, samples selected in accordance with 4.1 shall be tested for suspended matter and foreign material in the solution in accordance with APHA “Standard Methods for the Examination of Water and Wastewater”, section 209c “Total Suspended Solids dried at 103 - 105ºC”. See 3.3.1 for limits of suspended matter and foreign material.

4.2.2 Color. Buffer solutions of pH 3.0 to 11.0 inclusive in increments of 1 pH shall be prepared in accordance with NBS Research Paper 2619. 0.50 mL of indicator shall be added to 15 mL of each buffer solution. The mixtures shall be compared for color with the Government-loaned low, medium and high master color standard disks. The color of each pH mixture shall be visually in the range of color of the low and high master color standard disks as per 3.1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T21:30Z
Check the source to verify that this is the current version before use.

5. PACKAGING

5.1 Packaging. For acquisition purposes, the packaging requirements shall be as specified in the contract or order (see 6.2). When packaging of materiel is to be performed by DoD or in-house contractor personnel, these personnel need to contact the responsible packaging activity to ascertain packaging requirements. Packaging requirements are maintained by the Inventory Control Point's packaging activities within the Military Service or Defense Agency, or within the military service's system commands. Packaging data retrieval is available from the managing Military Department's or Defense Agency's automated packaging files, CD-ROM products, or by contacting the responsible packaging activity.

6. NOTES

(This section contains information of a general or explanatory nature that may be helpful, but is not mandatory.)

6.1 Intended use. The indicator solution covered by this specification is intended for use in determining the pH of water by comparison with color disks conforming to MIL-D-537, type I, class 8.

6.2 Acquisition requirements. Acquisition documents should specify the following:

a. Title, number, and date of this specification.
b. When Government-loaned property is not required (see 3.4)
c. Alternate sampling requirements (see 4.1)
d. When certification is acceptable (see 3.3.1)
e. Unit size required (4 or 16 ounces)
f. Special packaging requirements
g. Special marking requirements

6.3 Government-loaned property. The contracting officer should arrange to loan the property specified in 3.4, when applicable (see 6.2).

6.4 Standard sample. The contracting officer should arrange to furnish potential bidders a 16 ounce bottle of indicator solution, pH wide range, NSN 6810-00-087-2340.

6.5 Material Safety Data Sheets. Contracting officers will identify those activities requiring copies of completed Material Safety Data Sheets prepared in accordance with FED-STD-313. The pertinent Government mailing addresses for submission of data are listed in FED-STD-313. 29 CFR 1910.1200 requires the material Safety Data Sheet for each hazardous chemical used in an operation must be readily available to personnel using the material. Contracting officers will identify the activities requiring copies of the Material Safety Data Sheet.

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T21:30Z
Check the source to verify that this is the current version before use.

6.6 Subject term (key word) listing.

Chemical solution

6.7 Changes from previous issue. Marginal notations are not used in this revision to identify changes with respect to the previous issue due to the extent of the changes.

Custodians:
Army – AT
DLA – GS

Preparing Activity:
DLA-GS3

(Project 6810-2009-022)

Review Activity:
Army – MD1

NOTE: The activities listed above were interested in this document as of the date of this document. Since organizations and responsibilities can change, you should verify the currency of the information above using the ASSIST database at https://assist.daps.dla.mil.

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T21:30Z
Check the source to verify that this is the current version before use.