DISTRIBUTION STATEMENT A. APPROVED FOR PUBLIC RELEASE, DISTRIBUTION IS UNLIMITED.

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

USER ACTIVITIES: ARMY - ME

REVIEWER ACTIVITIES: ARMY - ME
NAVY - MS
AIR FORCE - 03

FED. SUP CLASS
6640



MS36001-1
Fixed Jaw



MS36001-2
Swivel Jaw



MS36001-3
All-Position Jaw

MATERIAL: Clamps shall be aluminum-base alloy in accordance with Federal Specification QQ-A-591, Alloy No. A380, or zinc-base alloy in accordance with Federal Specification QQ-Z-363, Composition A.

Fittings shall be forged aluminum in accordance with Federal Specification QQ-A-367, Composition 2017, or chemical resisting zinc alloy. Zinc-base alloy clamps and clamp holders may have copper alloy fittings in accordance with Federal Specification QQ-C-523 or QQ-C-525.

DESIGN: Fitting threads shall be in accordance with Federal Handbook H28 Unified National Coarse Thread Series. Thumbscrew shall be either 5/16-18UNC-2 or 1/4-20UNC-2 and have one to one and one-half inches of thread length.

Swivel type clamp shall be adjustable through an angle of 360 degrees and position fixed by means of a set screw.

All-position type clamp shall be adjustable to any angle within each plane.

Clamps shall be capable of direct attachment to rods up to 0.500 inches diameter without being passed over rod ends. When passed over rod ends, clamps shall accommodate a maximum diameter of 0.687 inches.

Overall length of clamps is approximately 5 to 6 inches when thumbscrews are retracted to accommodate maximum diameter rod.

Clamps shall be constructed in accordance with Federal Specification NNN-C-440 or ASTM E 784.

NOTES:
1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.
2. For design feature purposes, this standard takes precedence over procurement document referenced herein.
3. Referenced documents shall be of the issue in effect on date of invitations for bid.
4. FSCM: 08455, 11273, 22527, 64484.

APPROVED 23 DEC 1959 REVISED (A) 15 DEC 1967 (B) 7 NOV 1975 (C) 26 SEP 1985

(C) ENTIRE STANDARD REVISED

| | | | |
|---|---|---|---|
| P.A. EA<br>Other Cust MS 99 | INTERNATIONAL INTEREST | TITLE<br>CLAMP, LABORATORY SUPPORT ROD | MILITARY STANDARD<br>MS36001 |
| Procurement Specifications<br>NNN-C-440 | SUPERSEDES: | | SHEET 1 OF 1 |

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-1308

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:32Z
Check the source to verify that this is the current version before use.