
FED. SUP CLASS
6640



PART NUMBER	CAP	HEIGHT (CM)
MS3628[illegible]		45.7
MS36284		45.7

CANCELLED
CANCELLED AFTER 3 SEP 1991
NO SUPERSEDING STANDARD

MATERIAL: All [illegible] shall be non-toxic and [illegible] ade of thermal-insulating glass[illegible] for construction that will enable[illegible]

DESIGN: MS36282-[illegible] x 26.7 cm viewing window and MS36282-2 shall have thr[illegible]m viewing windows. The inside shall be coated with ocean blu[illegible] epoxy and the exterior antique white-colored high-impact laminated [illegible]estic on marine plywood. The aquarium shall have two independent water circulation systems, each with its own pump. A refrigeration system with a self-lubricating, hermetically sealed compressor shall thermostatically control the temperature range from 3°C to ambient +1°C (lower limit may vary with room temperature). Also, a cover plate, hydrometer/thermometer and operating manual shall be included.

The power requirements are 110 volts AC, 60 Hz, 6 amp (grounded plug). The electrical wiring shall conform to all the requirements of the Underwriters Laboratories, Code Number UL73 and UL1262. The approximate filled weight of MS36282-1 is 195Kg and MS36282-2 is 347Kg.

NOTE: Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

Ⓐ CANCELLATION

P.A.	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
EA		AQUARIUM, LABORATORY	MS36282
Other Cust MS 99			
Procurement Specifications: NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

CANCELLATION

Project Number 6640-1732-01

APPROVED 5 JAN 1983 REVISED Ⓐ 3 SEP 1991

USER ACTIVITIES: ARMY - MD, CM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 03

AMSC N/A
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

Source: https://assist.dla.mil -- Downloaded: 2020-10-23T17:37Z
Check the source to verify that this is the current version before use.