
FED. SUP CLASS
6640



MATERIAL: Rack shall be made of polyethylene plastic.

DESIGN: The rack shall hold twelve pipettes, straight or large bulb type, six on each side. The overall dimensions are 241 mm long, 178 mm wide and 292 mm high with 213 mm between the end plates. The rack shall have a 19 mm diameter carrying bar at the top.

NOTE: 1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

APPROVED 26 JUL 1982 REVISED

USER ACTIVITIES: ARMY - MD, OM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 03

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		RACK, PIPET	MS36401
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-0797
STANDARD DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23 17:40Z
Check the source to verify that this is the current version before use.