INTERIM COMMERCIAL ITEM DESCRIPTION

SPONGE, LAPAROTOMY, RADIOPAQUE, 8 by 36 Inches

The General Services Administration has authorized the use of this Interim Commercial Item Description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This Interim Commercial Item Description covers a Sterile, Disposable Radiopaque, gauze abdominal pack, suitable for use in gynecological procedures.

Salient Characteristics:

The sponge shall be a sterile, prewashed, disposable, radiopaque, four-ply gauze abdominal pack, 8 inches (+/- 1 inch) by 36 inches (+/- 2 inches) suitable for use in gynecological procedures.

The gauze shall be either Type III or Type IV Absorbent Gauze, USP. An inert nontoxic radiopaque insert tab shall be securely stitched between layers of gauze and shall be capable of withstanding normal steam sterilization after packs are removed from their immediate container.

The sponge shall be free from exposed raw edges, loose threads, puckering and other defects that detract from serviceability. All four plys of the finished sponge shall be joined together along all four edges and the sponge shall show no tendency to bunch or migrate. The sponge shall have a suitable tape loop sewn at one corner of the pack and if the loop is dyed, material shall be colorfast and shall contain no known toxic elements. The sponge shall be packaged so as to guarantee sterility until manually opened.

FSC 6510

DISTRIBUTION STATEMENT A. Approved for public release, distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:55Z
Check the source to verify that this is the current version before use.

Testing and examination shall be conducted to determine compliance with all specification requirements. Test methods shall be suitable, accurate and reproducible.

Workmanship. The sponge shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). Five packs shall be supplied in an immediate container capable of maintaining sterility of contents unless damaged or opened. One package containing 20 immediate containers (100 packs), as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6510-01-164-2724	SPONGE, LAPAROTOMY, Radiopaque, 8 by 36 inches (20.3 by 91.4 cm), 100s

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be as specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Project No. 6510-0731

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:55Z
Check the source to verify that this is the current version before use.