A-A-51385
9 APRIL 1986

COMMERCIAL ITEM DESCRIPTION

CONNECTOR, CANNULA, HEMODIALYSIS

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers one size of sterile, disposable connectors suitable for the connection of the extracutaneous end of two hemodialysis cannulae.

Salient characteristics:

Style and design. The connector shall be a tube 38 mm ± 2 mm in length. The tube shall have an inside diameter of 2.8 mm ± 0.2 mm and outside diameter of 3.4 mm ± 0.2 mm. The finish of the connector shall be smooth and non-etched. The connector shall be provided with an obturator. The obturator shall fit the connector securely and shall be color-coded green. The obturator shaft shall be 31 to 36 mm in length.

Material. The connector shall be manufactured from polytetrafluoroethylene. The obturator shall be manufactured from a rigid plastic.

Sterility. The connector and obturator shall be sterile when tested in accordance with the United States Pharmacopeia (USP).

Toxicity. The connector shall meet the requirements of the USP for Class IV plastics.

Workmanship. The item shall be processed in such a manner as to be uniform in quality and free from defects that will affect life, serviceability, or appearance.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:55Z
Check the source to verify that this is the currentversion before use.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Package (PG). One package containing fifty connectors, each connector individually packaged sterile in a peel-open type container capable of maintaining sterility of contents, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-165-1136	CONNECTOR, CANNULA, HEMODIALYSIS CATHETER

Ordering data (Intermediate/exterior packaging quantities and labeling and marking must be specified in the contract and/or order.)

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

50978

Project No. 6515-3564

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.