A-A-51390A
22 August 1986
SUPERSEDING
A-A-0051390(DM)
6 May 1986

COMMERCIAL ITEM DESCRIPTION

WIRE, ELECTRODE, PATIENT

The General Services Administration has authorized the use of this commercial item description in preference to A-A-0051390(DM), 6 May 1986.

This commercial item description covers a patient electrode wire suitable for use with push pin receiver cable for electrocardiograph monitoring equipment and electrode pads.

Salient characteristics:

The color coded set of three disposable patient electrode wires shall have a snap connector on the distal end and a "push pin" type connector on the proximal end. The wires shall be for use with a "push pin" receiver cable for electrocardiograph monitoring equipment and electrocardiograph electrode pads.

The electrode set shall be a minimum of 18 inches (45.72 cm) in length.

The electrode set shall contain one black, one white, and one green coded snap connector for easy placement identification.

Workmanship. The patient electrode wire shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing ten sets, each set containing three disposable wires, as specified, constitutes one unit.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contact or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and per table I. The unit of product for test purposes shall be one set of patient electrode wires.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred units)
Major Defects	II	4.0
Minor Defects	II	4.0

Tests. Tests shall be conducted to determine compliance with specification requirements and as per table II. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

TABLE II Categories and Defects *

Categories	Defects
Major	
101	Connector(s) detached from wire(s).
102	Distal connector not snap type.
103	Proximal connector not push pin type.
104	Set not a minimum of 18 inches long.
105	Snap connectors not one each black, white, and green.
Minor	
201	One package does not contain ten sets.
202	One set does not contain three wires.
203	Set package not sealed.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-139-7701

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-3850

Location: CID/WIRE-ELECTRODE-PAT/H17

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.