A-A-0051393(DM)
3 April 1986

INTERIM COMMERCIAL ITEM DESCRIPTION

DRESSING, BURN, 24 by 36 Inches

The General Services Administration has authorized the use of this Interim Commercial Item Description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This Interim Commercial Item Description covers a Sterile Burn Dressing suitable for treating large burns.

Salient Characteristics:

The sterile dressing shall be suitable for use in the treatment of burns involving large areas of the body. The dressing shall be approximately 24 by 36 inches in size and made of the following:

(a) Facing - Absorbent Cotton Gauze, USP, Type I.
(b) Wadding - Absorbent Cellulose (30 plies).
(c) Backing - Color free polyethylene, approximately 3 mil in thickness.
(d) Thread - White cotton, Ticket A. Type ICI, 3 ply, Federal Specification V-T-276.
(e) Finishing - The gauze facing shall be wrapped around the wadding prior to stitching. The gauze ends shall be folded in, to eliminate raw edges. Stitching shall include the gauze facing, absorbent cellulose, and polyethylene backing and run along all edges. The polyethylene backing shall cover all folded in edges of the gauze. Stitches shall number 6 to 12 stitches per inch. The stitch line shall be located in from the edges at 1(± 1/2) inch. The stitches shall be a lock stitch.

FSC 6510

DISTRIBUTION STATEMENT A. Approved for public release, distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

or

(a) Facing - Two (2) layers of 3 mil, high density, perforated polyethylene sheeting.
(b) Wadding - Compressed, absorbent cellulose, capable of absorbing at least ten times its weight in water.
(c) Backing - Color free polyethylene sheeting, at least 3 mils thick.
(d) Finishing - The polyethylene burn pad dressing components shall be heat-seal bound on all edges. The heat seal shall be smooth with even edges, and the cellulose filler shall not bunch, separate or wad when wet.

The dressing shall be white excluding backing and shall be sterile when tested with a biological indicator. Each pad shall appear neatly folded and be free of foreign debris, i.e., dirt, grease, etc., and loose threads shall be kept at a minimum. Each dressing shall have sufficient absorbency to retain approximately three liters of liquid.

Testing and examination shall be conducted to determine compliance with all specification requirements. Test methods shall be suitable, accurate, and reproducible.

Workmanship. The dressing shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). Dressing shall be individually packaged in a convenient tear-open or peel-open package capable of maintaining sterility of contents unless damaged or opened. One package containing 15 individually packaged dressings, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6510-01-153-2857	DRESSING, BURN, 24 by 36 inches (60.96 by 91.44 cm), 15s

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Ordering data. (Intermediate/exterior quantities, labeling, and marking must be as specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Project No. 6510-0729

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.