A-A-51397
11 APRIL 1986

COMMERCIAL ITEM DESCRIPTION

TUBING, SURGICAL
Plastic, 3/16 Inch Inside Diameter

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers three types of 3/16 inch or 5 millimeter inside diameter plastic surgical tubing suitable for suction in surgical procedures or for oxygen or anesthetic gas delivery.

Salient characteristics:

Sizes. Tubing shall be of the following lengths, as specified:

Size 1 - 180cm ± 8 cm or 6 feet ± 0.4 feet
Size 2 - 360cm ± 20 cm or 12.0 feet ± 0.7 feet
Size 3 - 3050cm ± 61 cm or 100 feet ± 2 feet

Style and Design. The tubing shall be nonconductive, flexible, and transparent. The tubing shall have an inside diameter of 3/16 ± 1/32 inch or 0.5cm ± 0.07cm. Tubing of 0.5cm inside diameter shall be provided in the centimeter length specified above. Tubing of 3/16 inch inside diameter shall be provided in the English length specified above. Labeling shall state the unit and size supplied. The tubing shall be compatible for use with oxygen and anesthesia gases, and shall not react with or deteriorate in the presence of the gases. The size 1 and size 2 tubing may have a smooth or corregated outside surface. The size 1 and size 2 tubing shall have two permanently attached female connectors and a nonconductive rigid male adapter provided unattached. The size 1 and size 2 female tubing connectors shall make a mechanically secure and leakproof connection with tubing connectors of outside diameter of 0.9 to 1.1cm. The size 3 tubing shall have an integral bubble dilation at each 91.4cm (36 inch) ± 3 cm interval. The container of the size 3 tubing shall have instructions for forming a taper or funnel connector from the bubble. The taper connector formed from the size 3 tubing bubble shall make a mechanically secure and leakproof connection with 0.56 to 0.95cm inside diameter tubing. The funnel connector formed from the size 3 tubing bubble shall make a mechanically secure and leakproof connection with 0.56 to 0.8cm outside diameter tubing.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Material. Tubing shall be manufactured from flexible transparent polyvinyl chloride (PVC) plastic. Female connectors shall be manufactured from flexible plastic. Male connectors shall be manufactured from rigid PVC.

Vacuum test. Finished tubing shall not collapse when subjected to an internal vacuum of 40 cm of mercury (min) for 30 seconds. Tubing shall be tested for vacuum resistance. Acceptance number shall be zero (0).

Connector test (sizes 1 and 2 only). Attach the tubing to a standard male adaptor ranging in outside diameter in steps from 0.9 to 1.1cm. A five pound tensile force shall be applied to the tubing for 30 seconds. The tubing shall not separate from the connector while the force is being applied.

Sterility (sizes 1 and 2 only). Tubes shall be tested for sterility in accordance with the USP.

Pyrogenicity (sizes 1 and 2 only). Tubing shall be free from pyrogens when tested in accordance with the USP.

Biocompatibility. Material used in the manufacture of tubes shall shall meet the requirements of the USP for Class IV plastic.

Workmanship. The tubing shall be free from foreign matter, cracks, and kinks. There shall be no defects which detract from the appearance or impair the serviceability of the tubing.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Package (PG) for size 1 and 2 and Coil (CL, for size 3. Size 1 and 2 tubing shall be individually packaged sterile in a sealed peel-open type container. Fifty individually packaged size 1 or twenty individually packaged size 2 tubing constitutes one unit or one hundred foot coil of size 3 tubing packaged in a dispenser type box constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for examination puposes shall be one tube. Sampling shall be Level II with an AQL of 2.5.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one tube. Sampling shall be S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-101-1948	TUBING, SURGICAL, 3/16 Inch, 100 feet long
6515-01-101-1949	TUBING, SURGICAL, 3/16 Inch, 12 feet long, 20's
6515-01-104-0043	TUBING, SURGICAL, 3/16 Inch, 6 feet long, 50"s

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PGC-05984

Project No. 6515-3560

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.