A-A-51398
14 April 1986

COMMERCIAL ITEM DESCRIPTION

BOUGIE, URETHRAL, OTIS

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers Otis design urethral bougies suitable for dilating the urethra.

Salient characteristics:

Material. Bougies shall be fabricated from one of the austenitic (Class 3) steels specified in ASTM F899.

Hardness. Bougies shall have a Rockwell hardness of B80 to B100 when tested as specified in GG-I-526.

Corrosion resistance. Bougies shall pass the boil test specified in GG-I-526.

Style and design. Bougies shall conform to the Otis design. Bougies shall consist of a distal olive tip, a straight shaft, and a pear-shaped proximal finger-grip.

Dimensions. Overall length of bougies shall be 9.875 ± 0.005 inches. Diameter of the olive tip shall correspond to the French size specified.

Finish. Bougies shall have a mirror finish. Bougies shall be properly passivated.

Workmanship. Bougies shall be free from cracks, pits, dents, and sharp edges. Bougie finger-grips shall have a uniform thickness. Bougie finger-grip edges shall be suitable dulled or rounded.

Unit. Each (EA). One bougie, of the size specified, constitutes one unit.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one bougie.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-139-8228	Bougie, Urethral, Otis, 8 Fr, Straight, Fixed Olive tip
6515-01-164-3169	Bougie, Urethral, Otis, 14 Fr, Straight, Fixed Olive tip
6515-01-165-6749	Bougie, Urethral, Otis, 10 Fr, Straight, Fixed Olive tip
6515-01-165-6750	Bougie, Urethral, Otis, 12 Fr, Straight, Fixed Olive tip
6515-01-166-9029	Bougie, Urethral, Otis, 22 Fr, Straight, Fixed Olive tip
6515-01-166-9030	Bougie, Urethral, Otis, 24 Fr, Straight, Fixed Olive tip
6515-01-166-9031	Bougie, Urethral, Otis, 28 Fr, Straight, Fixed Olive tip
6515-01-166-3621	Bougie, Urethral, Otis, 16 Fr, Straight, Fixed Olive tip

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:55Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-166-3622	Bougie, Urethral, Otis, 18 Fr, Straight, Fixed Olive tip
6515-01-167-3623	Bougie, Urethral, Otis, 20 Fr, Straight, Fixed Olive tip
6515-01-167-6676	Bougie, Urethral, Otis, 30 Fr, Straight, Fixed Olive tip
6515-01-167-6677	Bougie, Urethral, Otis, 26 Fr, Straight, Fixed Olive tip

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order)

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

06711

Project Order No. 6515-3514

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:55Z
Check the source to verify that this is the current version before use.