A-A-51400A
22 September 1986
SUPERSEDING
A-A-0051400(DM)
7 April 1986

COMMERCIAL ITEM DESCRIPTION

TEST KIT, FIBRIN AND FIBRINOGEN DEGRADATION PRODUCTS, DETERMINATION

The General Services Administration has authorized the use of this Commercial Item Description.

Salient Characteristics:

Shall be a latex agglutination slide test kit for the semi-quantitative determination of fibrin/fibrinogen degradation products.

Shall be suitable for use with human serum or urine samples.

Shall consist of the following components:

(1) One dispenser bottle containing not less than 2.5 ml of not less than 0.5 percent suspension of polystyrene latex particles coated with sheep anti-FDP globulin in glycine saline buffer containing 0.1 percent sodium azide and 0.1 percent thimerosal.

(2) One dispenser bottle containing one ml of positive control serum (human), diluted in glycine saline buffer containing 0.1 percent sodium azide to give an FDP concentration of 5-10 mcg per ml.

(3) One dispenser bottle containing one ml of negative control serum (human), diluted in glycine saline buffer containing 0.1 percent sodium azide to give an FDP concentration with less than 2 mcg per ml.

AMSC/NA FSC 6550

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

(4) Twenty evacuated sample collection tubes each containing 3600 NF units soybean trypsin inhibitor and 20 NIH units bovine thrombin for the collection of 2 ml whole blood or urine.

(5) Two 25 ml bottles of glycine saline buffer, pH 8.2.

(6) Not less than 20 each disposable pipettes and mixing rods.

(7) Rubber bulb.

(8) Glass test slide.

(9) Directions for use.

Workmanship. Product shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One test kit containing sufficient quantity of components for 20 tests, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurdisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The unrefrigerated shipping time for this item shall be not more than 3 days (72 hours).

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6550-01-141-0644	TEST KIT, Fibrin/Fibrinogen Degradation Products, Determination, 20 Tests

Ordering data. Expiration dating required, intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 51958

Project No. 6550-0053

U.S. GOVERNMENT PRINTING OFFICE: 1986 - 704-036/45591

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.