MIL-F-14452A(MU)
30 November 1967
SUPERSEDING
MIL-F-14452(CmlC)
21 December 1956

MILITARY SPECIFICATION

FILTER UNIT, GAS-PARTICULATE, HOSPITAL

SIX-MAN, 12 CFM, M7A1

1. SCOPE

1.1 This specification covers one type of hospital filter unit.

2. APPLICABLE DOCUMENTS

2.1 Government documents. The following documents of the issue in effect on the date of invitation for bids or request for proposal form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

NN-P-530	- Plywood, Flat Panel.
QQ-S-781	- Steel Strapping, Flat.
PPP-B-601	- Box, Wood; Cleated-Plywood.
PPP-B-636	- Box, Fiberboard.

MILITARY

MIL-P-116	- Preservation, Methods of.
MIL-B-121	- Barrier Material, Greaseproofed, Waterproofed, Flexible.
MIL-P-130	- Paper, Wrapping, Laminated and Creped.
MIL-P-14604	- Paper, Glassine.
MIL-T-14605	- Transformer Unit.
MIL-P-51070	- Purifier, Air, M2A2.

FSC 4240

Source: https://assist.dla.mil -- Downloaded: 202610-18T03:19Z
Check the source to verify that this is the current version before use.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

DRAWINGS

US ARMY MUNITIONS COMMAND

DL5-66-273 - Filter Unit, Gas-Particulate, Hospital, Six-Man, 12 CFM, M7A1.
D150-2-1 - Marking Diagram for Shipping Containers.

(Copies of specifications, standards, drawings, and publications required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on the date of invitation for bids or request for proposal shall apply.

UNIFORM CLASSIFICATION COMMITTEE

Uniform Freight Classification

(Application for copies of these ratings, rules and regulations should be addressed to Uniform Classification Committee, 202 Union Station, 516 West Jackson Boulevard, Chicago, Illinois 60606.)

3. REQUIREMENTS

3.1 Materials and components.

3.1.1 Materials. All materials cited on Drawing DL5-66-273 or on the subsidiary drawings shall conform to the specifications listed thereon or to the specific characteristic set forth on the drawings.

3.1.2 Components. All components of the headpiece shall conform to the specifications and drawings listed on Drawing DL5-66-273 and subsidiary drawings.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T03:19Z
Check the source to verify that this is the current version before use.

3.2 Manufacture and assembly. The components of the filter unit shall be manufactured and assembled in accordance with their respective assembly drawings which are listed on DL5-66-273.

3.3 Workmanship. The filter unit shall be free from damage such as bent, chipped or burred metal parts and cracked, kinked or torn rubber parts, and foreign matter (dirt, grease, oil or viscous material).

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection.

4.1.1 Supplier's responsibility. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.2 Objective evidence. The supplier shall provide objective evidence acceptable to the contracting officer that the requirements of 3.1 and section 5 for which specific inspection has not been provided in this specification have been satisfied.

4.2 Regular production.

4.2.1 Lotting. A lot shall consist of the filter units produced by one manufacturer, at one plant, from the same materials, under essentially the same manufacturing conditions. However, no more than one lot of air purifiers shall be represented in any one lot of finished filter units.

4.2.2 Sampling for examination. Sampling shall be conducted in accordance with MIL-STD-105.

4.2.3 Inspection.

4.2.3.1 For examination. Filter units shall be examined and tested in accordance with the classification of defects and MIL-STD-105.

Source: https://assist.dla.mil -- Downloaded: 202640-18T03:19Z
Check the source to verify that this is the currentversion before use.

4.2.3.2 Classification of defects.

(a) Filter unit, gas-particulate, hospital, six-man, 12 cfm, M7A1 (Dwg DL5-66-273).

Categories	Defects
Critical:	None defined
Major:	AQL 1.0 percent defective
101	Component missing
102	Component incorrectly assembled
103	Workmanship (3.3)

5. PREPARATION FOR DELIVERY

5.1 Cleaning and drying. The filter unit and its components shall be cleaned in accordance with method C-1 of MIL-P-116 and dried in accordance with any of the drying methods listed in MIL-P-116.

5.2 Preservation and packaging, level A.

5.2.1 Unit preservation and packaging.

5.2.1.1 Hose assemblies. Fittings attached to hose assemblies shall be cushion-wrapped with greaseproof carrier material conforming to type II, grade A, class 2 of MIL-B-121 or wrapping paper conforming to type II of MIL-P-130. The wrap shall be secured in place by tying or taping.

5.2.1.2 Accessory kit. Washer, machine screws, wood screws, and nuts shall be packed in accordance with method IC1 of MIL-P-116.

5.2.1.3 Technical publications and drawings. Technical publications and drawings shall be packaged in accordance with method IC3 of MIL-P-116.

5.2.1.4 Headpiece and carrier assembly. Each headpiece shall be folded as specified on Drawing C5-70-30 and placed in its carrier. To prevent adhesion, glassine paper conforming to MIL-P-14604 shall be used between folds of the headpiece and components.

Source: https://assist.dla.mil -- Downloaded: 202610-18T03:19Z
Check the source to verify that this is the current version before use.

5.2.2 Intermediate packaging.

5.2.2.1 Hose assembly. The hose assembly (M5) consisting of two each hose assemblies Drawing C5-66-269 and plug and socket Y-fitting, the accessory kit, and the technical publications and drawings, processed as specified in 5.2.1.1, 5.2.1.2 and 5.2.1.3, shall be packaged in a style RSC, type SF, class weather-resistant, grade W5s fiberboard box conforming to PPP-B-636, having approximate inside dimensions of 12-1/4 by 10-1/4 by 3-1/4 inches. The box shall be closed and sealed in accordance with applicable requirements in the appendix of the box specification.

5.2.2.2 Hose assembly. The hose assembly (M4) consisting of two each hose assemblies Drawing C5-66-269 and one each hose assembly Drawing C5-66-230, processed as specified in 5.2.1.1 shall be packaged in a style RSC, type SF, class weather-resistant, grade W5s fiberboard box conforming to PPP-B-636, having approximate inside dimensions of 12-1/4 by 10-1/4 by 5 inches.

5.2.2.3 Headpiece and carrier assembly. The headpiece and carrier assembly processed as specified in 5.2.1.4 shall be packaged in a style RSC, type SF, class weather-resistant, grade W5a fiberboard box which has approximate inside dimensions of 12-1/4 by 10-1/4 by 5 inches and conforms to PPP-B-636.

5.2.2.4 Purifier, air, M2A2. The air purifier shall be packed in accordance with the level A requirements specified in MIL-P-51070.

5.2.2.5 Transformer unit. The transformer unit shall be packed in accordance with the level A requirements specified in MIL-T-14605.

5.3 Packing. Level shall be as specified (see 6.2).

5.3.1 Level A. The filter unit, consisting of items preserved and packaged as specified in 5.2, shall be packed in a style A or B cleated plywood box conforming to PPP-B-601. The plywood used in the construction of the boxes shall conform to group A of NN-P-530 or type I, grade 3 or 4 of CS-35. Plywood shall not have continuous holes or openings extending through the entire panel from face to back. The box shall be closed and strapped in accordance with the appendix to PPP-B-601. Strapping shall conform to type I, class B of QQ-S-781.

5.3.2 Level C. The filter unit, preserved and packaged as specified in 5.2 shall be packed to insure carrier acceptance and safe delivery to the first domestic destination. The containers shall be in compliance with Freight Classification

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T03:19Z
Check the source to verify that this is the current version before use.

rules or the regulations of other common carriers applicable to the mode of transportation.

5.4 Marking. In addition to any special markings required by the contract or order, unit and intermediate packages shall be marked in accordance with MIL-STD-129. Shipping containers shall be marked in accordance with Drawing D150-2-1 (Marking diagram for shipping container.)

6. NOTES

6.1 Intended use. The filter unit covered by this specification is used to remove toxic gases and aerosols from the air and to supply purified air to six or fewer bedridden hospitatl patients suffering from wounds of the face, head or neck.

6.2 Ordering data. Procurement documents should specify the title, number, date of this specification and level of packaging and packing required.

6.3 Component specifications.

Purifier, Air, M2A2	- MIL-P-51070
Precleaner, Air Purifier, M1A1	- MIL-P-50085
Centrifugal Fan and Motor Assembly, 12 CFM	- MIL-C-50084
Separator, Dust	- MIL-S-50087
Filter, Particulate, 12 CFM, M13	- MIL-F-52011
Filter, Gas, 12 CFM, M12A1	- MIL-F-14512
Transformer Unit	- MIL-T-14605
Headpiece, Hospital Collective Protector, M13A1	- MIL-H-51182
Valve, Outlet, C-19	- MIL-V-51195
Clamps, Wire, Gas Mask	- MIL-C-10083
Hose; For Filter Units	- MIL-H-14506
Carrier, Hospital Collective Protector, M12	- MIL-C-14041

6.4 International interest. Certain provisions of this specification are the subject of international standardization agreement ABCA-ARMY-STD-154. When amendment, revision, or cancellation of this specification is proposed which will affect or violate

Source: https://assist.dla.mil -- Downloaded: 202610-18T03:19Z
Check the source to verify that this is the current version before use.

the international agreement concerned, the preparing activity will take appropriate reconciliation action through international standardization channels including departmental standardization offices, if required.

Custodian:

Army - MU

International interest (see Section 6.4)

Preparing activity

Army - MU (EA)

Project No. 4240-A352

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T03:19Z
Check the source to verify that this is the current version before use.