INCH-POUND

A-A-54658
30 May 1992

COMMERCIAL ITEM DESCRIPTION

CRUTCH, ALUMINUM, CANE TYPE, ADJUSTABLE

The General Services Administration has authorized the use of this commercial item description as a replacement for MIL-C-16671B which is cancelled.

This commercial item description covers an aluminum adjustable, cane type crutch for use in ambulation of patients.

Salient characteristics:

Materiel. The tubing shall be drawn seamless aluminum alloy conforming to 2024-T3 or as an alternate aluminum alloy 6061-T6. The handle casting shall be aluminum alloy casting of temper T6 or as an alternate, an aluminum die casting. The handle forging shall be aluminum alloy composition 2014-T6 or equal. The crutch tip shall be fabricated of a compound of natural rubber, synthetic rubber or a mixture of the two. Color of the tips shall be maroon, natural or other light color. The plastic hand grip cover shall be plastic polyvinyl chloride with color to match the tip. Corrosion resisting steel shall be Class 3 in accordance with ASTM Designation F899 for the arm support or as an alternate material, aluminum alloy 5052-H32.

Corrosion resistance. The corrosion resisting components shall show no evidence of corrosion or plating of copper when tested in accordance with ASTM Designation F1089.

Construction. The crutch shall be of single, telescoping aluminum tube construction with projecting plastic grip covered handle, and a C-shaped corrosion-resisting steel or aluminum arm support. Bottom of crutch shall be fitted with a heavy rubber crutch tip. Construction details shall conform to Defense Personnel Support Center (DPSC) Drawing No. 1367. Starting above hand grip, tubing shall bend slightly to provide comfortable positioning of arm support, which shall be attached by means of an anchor block to an adjusting post telescoping

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

into the upper end of tubing. Adjusting post shall be adjustable to four different heights. The lower portion of the crutch shall also be adjustable in length, to eight positions. A 300 pound weight shall be borne by the crutch without damage.

Upper tube section of crutch. Upper tube section of crutch shall be formed cold, without wrinkling interior of bend. Holes for tapered head screws must be jig drilled and reamed to match taper of mating parts to be not less than a Class 2 fit.

Lower tube section. Lower tube section shall have a smooth telescopic fit. Tube shall be jig drilled for interchangeability from side to side and between holes, and shall be jig reamed to be not less than a Class 2 fit with matching tapered holes in upper tube.

Locking screws and tapered sleeves. Locking screws and tapered sleeves shall be machined to a Class 2 or better fit with tapered holes, plus tolerance only on tapers to compensate for wear of the aluminum tubing.

Elastic stop nuts. Corrosion-resisting steel stop nuts for locking screws shall be the size shown on Drawing 1367. Countersunk flathead bolt for hand grip and stop nuts shall be machined to countersink taper as shown on the drawing.

Arm support. Arm support shall be cold formed with all sharp edges broken as shown on the drawing. Dimples for countersinks shall be of sufficient depth to allow heads of mating screws to set nearly flush with interior curvature of arm support when screwed into anchor block.

Anchor block. Countersinks in block shall be machined to fit dimples in arm support band as closely as possible without breaking through side wall of block.

Arm support adjusting post. Arm support adjusting post shall be drilled and reamed to match upper tube section mating holes, side for side, to at least a Class 2 fit. Post shall be machined for a sliding fit in upper tube section. Very slight countersink is indicated for this reason, and heading allowance shall be kept at a minimum to prevent cracking of the aluminum alloy rivets. Riveting shall be accomplished so that little or no resistance is offered by shank of rivet to the rotation of the arm support block.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Hand grip handle. The hand grip metal shall be free of pits, blowholes and other casting or forging defects. The metal hand grip shall be machined so that all hand grips will be interchangeable with all crutches and shall have a sliding fit on upper tube section for replacement purposes. Grip handle shall present a smooth surface before sliding on the plastic hand grip cover. The hand grip cover shall cover handle as shown on the drawing and shall fit sufficiently tight to prevent slippage in use.

Bolt. The bolt used to secure the hand grip handle to the crutch tubing shall be a flat-head bolt. The bolt shall be machined as shown on the drawing to a Class 2 fit for this thread type. Shank shall be capable of bridging upper tube section and hand grip without contact with bolt threads. The bolt shall be secured with the elastic stop nut.

Crutch tip. The crutch tip shall be standard 7/8 inch with hole reamed to fit tubing, and shall conform to details of the drawing. A metal bearing plate shall be molded into the tip. Hardness of the rubber tip shall be 40 to 55, Shore Durometer "A". As an alternate to the rubber, the crutch tip shall be fabricated of a plastic material, suitable for the purpose intended, firmly bonded to the metal. The crutch tip shall be abrasion resistant, and non marking.

Aluminum. Parts shall be anodized to a uniformly dense oxide coating. As an alternate, hand grip casting or forging shall be machined smooth and given a suitable chemical coating.

Identification marking and labeling. Each crutch shall be permanently and legibly marked with the contractor's name or registered trademark, placed in an inconspicuous location.

Workmanship. The crutch shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One crutch, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one crutch as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for test shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one crutch. Sampling shall be inspection level S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-00-777-7315	CRUTCH, ALUMINUM, CANE TYPE, ADJUSTABLE

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

PGC 46252

Project Number: 6515-5404
Location: c:\word\nish\crutchcd

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54658	30 May 1992

3. DOCUMENT TITLE

CRUTCH, ALUMINUM, CANE TYPE, ADJUSTABLE

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center ATTN: DPSC-MSE	215-737-2870	444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.