INCH-POUND

A-A-54670
15 June 1992

COMMERCIAL ITEM DESCRIPTION

PLUGGER, COMPOSITE RESIN, (Dr. Felt, Double-Ended)

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers double-ended, composite resin pluggers, of the Dr. Felt design for use in placement of resin restorative materials.

Salient characteristics:

Classification. The pluggers shall be classified as follows:

Dr. Felt, No. 4
Dr. Felt, No. 5

Material. The pluggers shall be fabricated of 6061-T6 aluminum. The aluminum shall conform to ASTM B221M for composition only. The aluminum surface shall be hardcoat anodized.

Style, design and dimensions. Style, design and dimensions of the Dr. Felt No. 4 and No. 5, double-ended, composite resin pluggers shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 23512. Each plugger shall be of one-piece construction. The diamond knurl shall be 0.004 to 0.007 inch above rod diameter, 29-40 TPI. The blade edges shall be slightly radiused at .010 inch.

Finish. The pluggers shall have a hardcoat anodized, dyed black finish. The blade finish shall be smooth. The pluggers shall have a sealed anodized coating which is colorfast and not sticky to the hand. The pluggers shall have no rough, sharp or burred edges. The pluggers shall have a non-porous, non-stick surface which allows composite, which has been hardened on the blade, to flick off easily without sticking.

Corrosion resistance. The pluggers shall show no evidence of corrosion after being subjected to 5 autoclave cycles of 270^{0}F $\pm$ 2^{0}F, and 27 $\pm$ 1 psi for 20 minutes each.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:37Z
Check the source to verify that this is the current version before use.

Workmanship. The pluggers shall be free from defects which detracts from their appearance or impairs their serviceability.

Unit. Each (EA). One Dr. Felt composite plugger, either No. 4 or No. 5, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one plugger, No.4 or No. 5, as specified in this description.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one plugger, No. 4 or No. 5, as specified in this description.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Tests. Test shall be conducted to determine compliance with specification requirements.

Performance test. The pluggers shall be scraped across cured composite, with an applied pressure of 0.5 ± 0.1 pounds, using 3-5 mm long strokes. Neither the composite nor the instrument shall show signs of streaking after the instrument has been scraped across it, as specified above.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

National Stock Number	Item Identification
6520-01-328-5192	Plugger, Composite Resin, Dr. Felt No. 4, Double-Ended
6520-01-328-5193	Plugger, Composite Resin, Dr. Felt No. 5, Double-Ended

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6520-2520
Location: D.C. Disk #2
(PLUGGER)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE. This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

1. RECOMMEND A CHANGE:	1 DOCUMENT NUMBER: A-A-54670	2 DOCUMENT DATE (YYMMDD): 15 June 1992

3 DOCUMENT TITLE: PLUGGER, COMPOSITE RESIN, (Dr. Felt, Double-Ended)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5 REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable) / e. DATE SUBMITTED (YYMMDD)

7 PREPARING ACTIVITY

a NAME	b TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	(215)737-2870	444-2870
c ADDRESS (Include Zip Code) ATTN: DPSC-MSE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403 Falls Church VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete — 198/288

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:37Z
Check the source to verify that this is the current version before use.