INCH-POUND

A-A-54672
15 June 1992

COMMERCIAL ITEM DESCRIPTION

ANESTHESIA KIT, LARYNGOTRACHEAL, DISPOSABLE

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-A-37842 (DLA-DM) which is cancelled.

This commercial item description covers a disposable laryngotracheal anesthesia kit for the administration of topical anesthesia to the interior of the larynx and trachea.

Salient characteristics:

General. The laryngotracheal anesthesia kit shall consist of a sterile, plastic dispensing injector (vial injector) with integral stainless steel cannula, a 5 ml glass vial containing 4 ml of sterile Lidocaine Hydrochloride Solution, United States Pharmacopeia (USP) 4 percent and instructions for use. The stainless cannula shall be ground to a point at the proximal end to facilitate penetration of the rubber stopper.

Material. Material of the metal barrel cannula shall be Class 3 corrosion resistant steel in accordance with ASTM Designation F899. The barrel plastic cannula shall be fabricated of polycarbonate resin and the barrel shall be fabricated of polypropylene. The vial shall be made of USP Type I borosilicate glass and the rubber stopper (plunger) shall be made of rubber inert to the contained medication.

Style, design and dimensions. The anesthesia kit shall be in accordance with Figure 1 for style, design and dimensions.

Passivation. Corrosion resistant steel shall be properly passivated.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Corrosion resistance. Corrosion resistant steel cannula shall show no evidence of corrosion or plating of copper when tested in accordance with ASTM Designation F1089.

Sterility. The Lidocaine and vial injector shall be sterile in accordance with the USP.

Security of assembly. There shall be no separation of the plastic cannula from the vial injector when a tensile load of 25 pounds is applied for one minute to the injector.

Ease of injection. The cap shall be removed from the vial and the vial screwed into vial injector. The vial shall then be forcefully compressed into the vial injector to discharge the Lidocaine. The Lidocaine shall be completely discharged in three (3) seconds.

Instructions. Each kit shall be furnished with a list of the kit contents, assembly instructions and procedural instructions for use.

Workmanship. The anesthesia kit shall be free from defects which detract from its appearance or impair its serviceability.

Expiration dating period. The anesthesia kit expiration dating period shall be not less than 24 months. Not less than 20 months of the expiration dating period shall remain at time of delivery to the Government.

Unit. Package (PG). One package containing twenty-four individually packaged sterile anesthesia kits, as specified, constitutes one unit. The individual package shall be a sealed peel-open type container capable of maintaining sterility of contents unless the package is damaged or opened.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one anesthesia kit. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one anesthesia kit. Sampling shall be inspection level S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number and its respective Item Identification covered by this document:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-003-9783	ANESTHESIA KIT, LARYNGOTRACHEAL, Disposable, 24's

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:


DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5405
Location: c:\word\nish\anesthcd

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:37Z
Check the source to verify that this is the current version before use.



ANESTHESIA KIT, LARYNGOTRACHEAL

FIGURE 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54672	15 June 1992

3. DOCUMENT TITLE
ANESTHESIA KIT, LARYNGOTRACHEAL, DISPOSABLE

4. NATURE OF CHANGE (*Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.*)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (*Last, First, Middle Initial*)	b. ORGANIZATION	
c. ADDRESS (*Include Zip Code*)	d. TELEPHONE (*Include Area Code*) (1) Commercial (2) AUTOVON (*If applicable*)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME Defense Personnel Support ATTN: DPSC-MSE	b. TELEPHONE (*Include Area Code*) (1) Commercial 215-737-2870 (2) AUTOVON 444-2870
c. ADDRESS (*Include Zip Code*) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil - Downloaded: 202640-26T09:37Z.
Check the source to verify that this is the current version before use.