INCH-POUND

A-A-54675
19 June 1992

COMMERCIAL ITEM DESCRIPTION

BAND AND BAND SET, MATRIX, DENTAL, Tofflemire

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specification GG-B-95E which is cancelled.

This commercial item description covers matrix bands suitable for use for restorations in individual teeth.

Salient characteristics:

Classification. The matrix bands shall be of the following types and sizes, as specified:

Type I. Bands, Junior size, Tofflemire Sizes

Universal, No. 13
Narrow M.O.D., No. 14
Wide M.O.D., No. 15

Type II. Bands, Adult size, Tofflemire Sizes

Universal, No. 1
Wide M.O.D., No. 2
Narrow M.O.D., No. 3
Window-Opening Contra Angle, No. 10
Universal Contra Angle, No. 11

Material. The matrix bands shall be fabricated from austenitic stainless steel type 301, 302, or 304 in accordance with ASTM Designation F899.

Hardness. The matrix bands shall have a Knoop hardness of 351 to 402, when tested as specified below, or a Rockwell hardness of C35 to C40, when tested in accordance with ASTM E18.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Style, design and dimensions. The matrix bands shall be capable of being fitted around a tooth to supply a missing wall or contour, against which the filling material may be placed and condensed.

Bands shall be furnished in pre-stamped designs corresponding to the applicable design and size (number) specified. Junior size bands shall be suitable for use with junior size Tofflemire retainers, and adult size bands shall be suitable for use with adult size Tofflemire retainers. All matrix bands shall conform to the design, dimensions, and size (number) shown in Defense Personnel Support Center (DPSC) Drawing No. 23355. Bands shall be furnished in the applicable size number, thickness, and quantity shown in Table I. The thickness tolerance for all bands shall be ± 0.0002 inch.

Table I. Matrix Band Size (number), Thickness and Quantity.

Type	Size (number)	Thickness, inches	Quantity	Total Quantity	Unit of issue
1. Junior Size Tofflemire	Universal #13	0.0020	6		Set
	Narrow MOD #14		2	12	
	Wide MOD #15		4		
2. Adult Size Tofflemire	Universal #1	0.0015	12	12	Package
	Universal #1	0.0020	12	12	Package
	Wide MOD #2	0.0015	12	12	Package
	Narrow MOD #3	0.0015	12	12	Package
	Window-Opening Contra-Angle, #10	0.0015	4		
	Universal Contra-Angle #11	0.0015	8	12	Set

Corrosion resistance. Bands shall show no signs of corrosion when subjected to the boil test in accordance with ASTM F1089.

Flexibility. Bands shall withstand a minimum of two bends before breaking, when tested as specified below. A bend shall consist of one flattening followed by straightening.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Finish. The bands shall have a commercial number 2B finish or better. The bands shall be cleanly trimmed and smoothly finished. All bands shall be put through a tumbling procedure to provide smooth peripheral edges free from burrs or other imperfections. The surface roughness of bands shall be no more than the root-mean-square average deviation of 15 microinches when the bands are tested with a suitable surface roughness analyzer.

Workmanship. The finish of the bands shall be uniform and free of burrs, sharp edges, pit marks, cracks, wrinkles, folds, tears or other imperfections. The matrix bands shall be free from any imperfections which may detract from their appearance or impair their serviceability, proper functioning, and intended use.

Unit.

Package (PG). One package containing twelve matrix bands of one size and thickness specified, constitutes one unit.

Set. (SE). One set containing twelve junior or twelve adult matrix bands of assorted sizes, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the component or materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one package or one set of matrix bands. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, except for dimensions, which shall be conducted at inspection level S-1, AQL 2.5. Sampling for tests shall be conducted at inspection level S-2 and the acceptance number shall be zero (0) for all sample sizes.

Examination. The matrix bands shall be examined to determine compliance with specification requirements.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Hardness. Knoop hardness shall be determined on a Tukon tester under a 500-gram load, using a rhombus-shaped diamond indenting tool with a longitudinal angle of 172.5 degrees and a traverse angle of 130 degrees.

Flexibility test. The matrix band shall be bent back on itself by creasing it in the middle (so that one half is superimposed on the other) and returning it to its original flat position. The flattening and straightening of the 180 degree bend shall be done using a rubber-covered metal roller weighing $4\text{-}1/2$ plus or minus $1/2$ pounds. Subsequent bends shall be performed in the same way with the inside surface of the crease the same for all bends.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

National Stock Number	Item Identification
6520-00-682-6530	BAND SET, MATRIX, DENTAL, Assorted, Tofflemire, Junior Size, No. 13, 14, and 15, 0.0020 in. thick, set of 12
6520-00-721-9351	BAND, MATRIX, DENTAL, Tofflemire, Adult Size, No. 3, 0.0015 in. thick, 12's
6520-00-721-9355	BAND, MATRIX, DENTAL, Tofflemire, Adult Size, No. 1, 0.0015 in. thick, 12's
6520-00-721-9356	BAND, MATRIX, DENTAL, Tofflemire, Adult Size, No. 2, 0.0015 in. thick, 12's
6520-00-817-2297	BAND SET, MATRIX, DENTAL, Assorted, Tofflemire, Adult Size, No. 10 and 11, 0.0015 in. thick, set of 12
6520-01-079-6363	BAND, MATRIX, DENTAL, Tofflemire, Adult Size, No. 1, 0.0020 in. thick, 12's

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6520-2299
Location: PL Disk #2
(BANDS)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54675	19 June 1992

3. DOCUMENT TITLE
BAND AND BAND SET, MATRIX, DENTAL, Tofflemire

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	(215)737-2870	444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403 Falls Church VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete — 98/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.