INCH-POUND

A-A-54676
19 June 1992

COMMERCIAL ITEM DESCRIPTION

Syringe and Needle, Hypodermic, Disposable, Tuberculin

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers sterile, disposable, hypodermic syringes with and without needles suitable for a single injection.

Salient characteristics:

General. Shall be a 1 cc syringe having a Luer slip tip. Where a needle is required, it shall be a detachable 25 gage, 5/8 inch long, regular bevel having a Luer slip hub.

Material. Needle cannula shall be type 304 stainless steel in accordance with ASTM Designation F899 or AISI. Syringe barrel and plunger shall be rigid polypropylene which meets the USP requirements for Class VI plastics. The plunger tip shall be lightly coated with an inert pharmaceutical or medical grade silicone lubricant. The plunger tip shall be fabricated from a pharmaceutical grade rubber or neoprene which meets the USP requirements for Elastomeric Closures for Injections.

Style and design. Syringe and needle shall be a Luer slip type in accordance with ISO 594. Syringe barrel shall be flanged to provide finger grip and anti-roll feature. The inside of the barrel shall have a smooth, uniform surface, free from pits, grind marks, bubbles, flash, or other imperfections. Plunger shall glide freely and smoothly through the barrel while maintaining a fluid-tight fit. The plunger tip and barrel shall both be beveled or flat-faced and form a flush fit at zero volume. Barrel shall be transparent and the plunger tip shall have visible rings touching the sides of the barrel.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

The outside surface of the barrel shall have main graduations at each 0.1 cc and subgraduations at each 0.01 cc. Markings shall be at 0.1 cc. Syringe shall deliver ± 5 percent of the capacity indicated at each main graduation mark when tested against a calibrated buret. Barrel shall also have the name or registered trademark of the contractor. Actual fluid volume shall be determined by position of the distal ring of the plunger tip. As an alternate the determination of capacity test listed in ISO 7886 is acceptable.

Needle shall be furnished with a removable, protective cap. The tip of the needle point shall be ground to maximum keenness. The heel of the needle point shall be dull to prevent coring.

The needle style shall be as follows:

Gage	O.D. BWG (+ 0.0005 - 0 inch)	I.D. BWG (+ 0.002 inch)	Needle Length (inch)	Bevel (inch)
25	0.020	0.0095	5/8 ± 1/16	0.078 - 0.103

Leakage test. The water pressure test system shall have a water pressure gage and a rubber plug. The needle shall be secured to the Luer tip of the syringe by applying it with a 5 pound axial force and while maintaining this compressive stress twisting the needle hub 45 degrees relative to the Luer tip. The needle of the hypodermic shall be inserted into the plug. The syringe shall be tested for liquid leakage at the syringe piston and syringe nozzle under compression and during aspiration in accordance with ISO 7886.

Needle assembly test. The needle cannula shall not crack or separate from the needle hub when subjected to a 5 pound axial tensile force exerted for one minute. (See Note 3).

Needle bending resistance test. The needle cannula shall not crack or separate from the needle hub when bent 25 degrees from its rest position and subsequently bent 25 degrees in the opposite direction.

Sterility, pyrogenicity, toxicity. The syringe needle shall be nonpyrogenic and nontoxic in accordance with the United States Pharmaceutical (USP). The syringe and needle shall be sterile in accordance with USP or Association for the Advancement of Medical Instrumentation (AAMI) requirements. Sterility assurance level shall be 0.0001.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Instructions. Each unit package shall be provided with instructions describing the proper techniques for dispensing the sterile syringes and needles, as applicable, from the containers and preparing them for use. In addition, information as to methods for destroying syringes and needles, as applicable, to prevent their re-use shall also be shown.

Workmanship. The syringes and needles shall be free from defects which detracts from their appearance or impairs their serviceability.

Unit. Package (PG). One package containing one hundred individually packaged, sterile syringe and needle units of the type and size, as specified, constitutes one unit. The individual package shall be a sealed peel-open type container or rigid plastic tamper-proof container capable of maintaining sterility of contents unless package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified below. The unit of product for test purposes shall be one syringe and needle.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

NOTE (1): The following National Stock Numbers and their respective item identifications are covered by this document:

National Stock Number	Item Identification
6515-00-982-4205	Syringe and Needle, Hypodermic, Disposable, Tuberculin, 25 Gage, $^{5}/_{8}$ inch Needle, 1cc Syringe, 100's
6515-00-982-4206	Syringe, Hypodermic, Tuberculin, 1cc, 100's

NOTE (2): ANSI and ISO standards are available from the American NAtional Standards Institute, Inc., 11 West 42nd Street, New York, NY 10036.

NOTE (3): An alternate test, such as a pull-to-destruction test, is acceptable in lieu of the 5 pound separation test, provided product found acceptable of the pull-to-destruction test also passes the 5 pound separation test.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

A-A-54676

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-5440
Location: ML Disk #2
(SYRINGE2)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54676	2. DOCUMENT DATE (YYMMDD): 29 June 1992

3. DOCUMENT TITLE: Syringe and Needle, Hypodermic, Disposable, Tuberculin

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable) / e. DATE SUBMITTED (YYMMDD)

7. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	215)737-2870	444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.