INCH-POUND

A-A-54686
19 June 1992

COMMERCIAL ITEM DESCRIPTION

PLUGGER, AMALGAM, DENTAL (Markley)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-P-36577A.

This commercial item description covers dental pluggers suitable for use in condensing silver amalgam during dental restorative procedures.

Salient characteristics:

The pluggers shall be of the following styles as specified below:

Style 1. - Markley No. 1
Style 2. - Markley No. 3
Style 3. - Markley No. 4
Style 4. - Markley No. 5

The pluggers shall be double-ended and of three-piece construction.

Material.

Operative ends. The operative ends of the pluggers shall be fabricated entirely of class 4, type 440A martensitic stainless steel as specified in ASTM Designation F899 for chemical composition only.

Handles. The handle shall be type A,B,C, or D as shown in Defense Personnel Support Center (DPSC) Drawing No. 24173. Type A, B and D brass handles shall conform to ASTM Designation B135 for chemical composition only. Type A handle shall be low leaded brass, copper alloy No. C33000 of ASTM Designation B135. Type B handle and type D handle shall be cartridge brass, copper alloy No. C26000 of ASTM Designation B135. Brass handles shall be chromium-over-nickel plated. Type C handle shall be class 3, type 303 austenitic stainless steel as specified in ASTM Designation F899.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Style, design, and dimensions. The pluggers shall be double-ended. The design and dimensions of the pluggers shall conform to Defense Personnel Support Center (DPSC) Drawing No. 24173. The pluggers shall be suitable for use in compacting plastic silver amalgam around reinforcement pins in the mouth of a dental patient.

Construction. The plugger shall be double-ended and of three piece construction, with the shanks of the operative ends secured in the handle. The interface of the shank and handle shall be smoothly blended with no crevices evident. There shall be no movement, loosening, or separation between the operative end and the handle when the operative end is subjected to an 85 pound (minimum) pull out force. Handle A shall be octagonal. Handles B, C and D shall be round. Handles A and B shall be hollow. Handles C and D shall be solid. The finger grip portions of handles B, C, and D shall be diamond knurled. Handle A shall have superficial diamond serrations on seven faces and shall have one smooth face for identification marking as shown on DPSC Drawing No. 24173.

Hardness. After proper stress relieving, the Rockwell hardness of the operative ends shall be C52 to C56. Handles and shanks shall be properly tempered and shall have a hardness suitable for the purpose intended. Hardness shall be determined in accordance with ASTM E18 or E92.

Finish. The operative end and shank shall have a smooth mirror or smooth bright finish. The operative end shall be free of defects such as grind marks, cracks, gouges, burrs, feathers or other imperfections. Handles shall have a mirror or bright finish, except for the diamond knurl. Finger grip or serrated portion of handle shall be free from barbs and shall provide adequate finger grip during use.

Corrosion resistance. The entire instrument shall pass the boil test specified in ASTM F1089. The copper sulfate test requirements of ASTM F1089 shall apply to the entire plugger for handle C (corrosion-resisting steel handle) but only to the operative ends for pluggers with handles A, B, and D (chromium over nickel plated brass).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Identification marking. The smooth side of the octagonal handle and the middle section of the round handles shall be permanently and legibly marked with the name or registered trademark of the manufacturer or contractor, the instrument number ("1M", "3M", "4M", or "5M", as applicable) and the word "stainless". Instruments of foreign manufacture shall bear the name of the country of origin.

Workmanship. The plugger shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One plugger, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the component or material, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one plugger. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, except for dimensions, which shall be conducted at inspection level S-1, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-2. For tests, the acceptance number shall be zero (0) for all sample sizes.

Examination. The amalgam pluggers shall be examined to determine compliance with specification requirements.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

National Stock Number	Item Identification
6520-00-935-7167	PLUGGER, AMALGAM, DENTAL, Markley, No.1
6520-00-935-7168	PLUGGER, AMALGAM, DENTAL, Markley, No.3
6520-00-935-7169	PLUGGER, AMALGAM, DENTAL, Markley, No.4
6520-00-935-7170	PLUGGER, AMALGAM, DENTAL, Markley, No.5

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

A-A-54686

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6520-2517
Location: PL Disk #2
(AMALGAM)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54686	19 June 1992

3. DOCUMENT TITLE
PLUGGER, AMALGAM, DENTAL (Markley)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	(215)737-2870	444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403 Falls Church VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.