INCH-POUND

A-A-54691
25 June 1992

COMMERCIAL ITEM DESCRIPTION

LIGHT, FLOOR

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for W-L-316B which is cancelled.

This Commercial Item Description covers a portable, floor standing operating and examining light, capable of accommodating a standard incandescent lamp bulb.

Salient Characteristics:

The floor light shall be a portable operating and examination light, capable of accommodating an incandescent lamp bulb of sufficient wattage to provide adequate illumination without overheating the reflector, and suitable for use in hospital clinics.

The light shall consist of a floor lamp with weighted base, upright housing, telescopic rod operating within the housing, a flexible arm and socket, and a reflector which is designed to allow air to flow for venting purposes.

The base of the light shall be between 10 and 11-1/2 inches in diameter and shall be finished so as to permit the light to rest on a flat surface without wobbling or tilting. The base shall be of sufficient weight (8 pounds minimum) to provide stability when telescopic rod is fully extended. The base shall be made of pressed steel or it shall be made of cast iron or lead with a sheet metal or heavy-duty plastic cover.

The upright housing shall be not less than 7/8 inch in diameter and the telescoping tubing not less than 5/8 inch in diameter, and shall be brass or electric welded steel tubing. The telescopic rod shall move easily within the upright and

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

shall readily locate the lamp head at any desired height by means of a friction type mechanism. The friction type mechanism shall not have manually operated thumb screws that may easily become lost. The telescoping rod shall permit height adjustments of the lamp head as measured from the floor to the top of the telescoping rod of approximately 35 to 54 inches.

The flexible arm shall be made of steel tubing, and be of conventional design, approximately 14 inches long and 5/8 inch in diameter. It shall be properly bushed to receive the electric cord.

The lamp switch shall be single-pole with non-metallic control and casing, and shall be of the key, turn-knob, or push button type. The lamp socket shall be rated at 250 watts, 250 volts with a brass exterior. It shall be securely fastened to the upper end of the flexible arm or within the reflector.

The reflector shall be designed to allow air to flow through for venting purposes and made of steel, aluminum, or a suitable heavy duty plastic. The reflector outer surface shall not exceed 122 degrees F at any point during operation. The reflector shall be connected to the flexible arm in such a manner to allow adjustment of the light head to any desired direction.

The lamp shall be rated for 80 footcandles at 24 inches using a 60 watt standard bulb. The lamp shall be an Underwriter's Laboratory listed portable lamp meeting UL Standard #153 and shall bear the UL label.

The floor light shall have a 3 x 18 AWG, 60 degree Celsius rated, strain and pullback relieved cord fitted with a hospital grade plug. Minimum length of cord shall be 9 feet.

A name plate shall be securely and firmly attached to the lamp and shall bear the name or registered trade-mark of the manufacturer. The marking shall be clear, distinct, legible, and of a permanent nature.

All exterior metal surfaces shall have a plated, enameled, painted, varnished, lacquered, zinc or cadium coated finish.

Workmanship. The floor light shall be free from defects which detract from its appearance or impair its serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Unit. Each (EA). One floor light, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one floor light.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one floor light.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to, the following:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Operation test. Each light shall be operated in accordance with manufacturer's instructions to determine compliance with requirements herein.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

6530-00-706-5100
6530-01-174-2336

MILITARY INTERESTS

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
GSA - FSS - FCGC
USPHS
VA - OSS

Project Number: 6530-2349

Location: LTFLRC1B.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54691	2. DOCUMENT DATE (YYMMDD): 25 June 1992

3. DOCUMENT TITLE: LIGHT, FLOOR

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	e. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 98/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.