METRIC

A-A-54703
20 July 1992

COMMERCIAL ITEM DESCRIPTION

DEMINERALIZER, WATER, ION EXCHANGE, HAND OPERATED

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-D-37524 which is cancelled.

This Commercial Item Description covers the requirements for hand-operated water demineralizer suitable for use in removing dissolved ionized minerals from water.

Salient characteristics:

Shall be a hand-operated water demineralizer consisting of a bottle and a cartridge.

The bottle shall have a nominal capacity of 473 ml and shall be fabricated of polyethylene. The neck of the bottle shall be constructed so that it securely holds the cartridge. The bottle shall be 171 ± 4 mm high with an overall outside diameter of 70 ± 4 mm.

The cartridge shall be fabricated of rigid plastic and contain ion exchange resins and an indicator. The top of the cartridge shall have a protuberance which fits into the top of the bottle. The overall inside diameter of the protuberance shall be 16.00 ±.13 mm.

The cartridge shall contain enough resins to remove 2.21 grams of sodium chloride from a solution of sodium chloride and water. The ion exchange resins shall be a mixture of cation and anion resins. A blue indicator shall be applied to the anion resin prior to mixing. The change in the color of the indicator shall be apparent as a yellowish brown band. The band shall start at the base of the cartridge and expand upward as the ion exchange capacity of the cartridge is expended. The cartridge shall have a permanent discard mark. The mark shall indicate that the cartridge is to be discarded when the yellowish brown band reaches the mark.

AMSC N/A FSC 4610

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

With a full bottle of water attached to the cartridge and squeezed the water shall pass through the cartridge and there shall be no evidence of the leakage from the cartridge, the bottle, or the area where the bottle and cartridge are connected.

The demineralizer shall produce water which contains no particles and no more than 10 parts per million of sodium chloride.

The demineralizer shall be permanently marked with the name or registered trade mark of the manufacturer.

Workmanship. The water demineralizer shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One water demineralizer, as specified, constitutes one unit.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor Certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:37Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. (Intermediate/exterior package quantities, labeling and marking must be specified in the contract and/or order).

NOTE: The following National Stock Number is covered by this document:

NSN	Item Identification
4610-00-299-9827	DEMINERALIZER, WATER, ION EXCHANGE HAND-OPERATED.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

A-A-54703

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 4610-0127

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54703	20 JULY 1992

3. DOCUMENT TITLE
DEMINERALIZER, WATER, ION EXCHANGE, HAND OPERATED

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME Defense Personnel Support Center ATTN: MSE	b. TELEPHONE (Include Area Code) (1) Commercial: (215)737-2870 (2) AUTOVON: 444-2870
c. ADDRESS (Include Zip Code) 2800 S. 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete. '98

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.