INCH-POUND

A-A-54705
20 July 1992

COMMERCIAL ITEM DESCRIPTION

FUNNEL, COMMON, LABORATORY

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-F-36748 which is cancelled.

This Commercial Item Description covers one type of ribbed polyethylene laboratory funnels for transferring liquid.

Salient characteristics:

Capacity sizes. Shall be laboratory funnels with the following nominal capacity sizes (see ordering data):

2 quart

4 quart

General. Shall be laboratory funnels with external ribbing for air release. The funnels shall be fabricated from polyethylene and shall be capable of withstanding temperatures up to at least 180°F. Each funnel shall be of one piece construction and shall consist of a bowl with a 60° body taper and a stem tip with an angle cut. The top outside perimeter of the bowls shall have a reinforcing bead. The exterior of each funnel shall have at least four longitudinal ribs that are spaced 90° apart around the bowl and include the locations where the stem and body meet. The surfaces of each funnel shall be uniform and free of mold marks. The body shall be uniformly round and the stem shall be perpendicular to the center of the body. The interior may be supplied with ribs.

Material. The material for the funnels shall be polyethylene in accordance with ASTM D1248, type I, class A, grade 3.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Style, design and dimensions. The 2 quart capacity funnel shall have inside diameter at top of 8 in. and stem length of 3-11/16 in. The 4 quart capacity funnel shall have inside diameter at top of 11 in. nominal and stem length of 3-1/2 in. A tolerance of ± 1/2 inch shall apply to all the dimensions.

Identification marking. Each funnel shall be permanently marked with the manufacturer's name or registered trademark.

Workmanship. The funnels shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One funnel, of the size specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination and Testing. Examination and testing shall be conducted on the funnels in accordance with commercial practice. The sampling for examination and testing shall be in accordance with commercial practice.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor Certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Other publications. The following documents form a part of this specification to the extent that is specified herein. Unless otherwise indicated, the effect on date of invitation for bids or requests for proposals shall apply.

ASTM D1248 Polyethylene Plastics Molding and Extrusion Materials

(Application for copies should be addressed to ASTM, 1916 Race Street, Philadelphia, PA 19103.)

Ordering data. (Capacity size required, intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.)

NOTE: The following National Stock Numbers are covered by this document:

NSN	Item Identification
6640-00-058-9146	FUNNEL, COMMON, LABORATORY, POLYETHYLENE, 4 Quart.
6640-00-063-7872	FUNNEL, COMMON, LABORATORY, POLYETHYLENE, 2 Quart.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6640-1751

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2 DOCUMENT DATE (YYMMDD)
	A-A-54705	20 JULY 1992

3. DOCUMENT TITLE

FUNNEL, COMMON, LABORATORY

4 NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed.)*

5 REASON FOR RECOMMENDATION

6. SUBMITTER

a NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME	b TELEPHONE (Include Area Code)	
Defense Personnel Support Center	(1) Commercial (215)737-2870	(2) AUTOVON 444-2870
ADDRESS (Include Zip Code) ATTN: MSE 2800 S. 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT. Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 *Previous editions are obsolete* 198/290

0102 - LF - 009 - 2700

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.