INCH-POUND

A-A-54708
24 July 1992

COMMERCIAL ITEM DESCRIPTION

STERILIZER, SURGICAL INSTRUMENT AND DRESSING (AUTOCLAVE)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a table top sterilizer (autoclave) suitable for sterilizing dental and surgical instruments and dressings.

Salient characteristics:

The unit shall be a table top sterilizer (autoclave), electrically heated, automatically controlled, pressure, single shell, encased in a cabinet, suitable for sterilizing dental and surgical instruments and dressings.

The autoclave shall have a peripheral heater to assure uniform chamber heat. Shall have a metering valve which introduces fresh water for each cycle but no recirculation of used water which causes "autoclave odor".

Shall have the following nominal dimensions:

a. Overall length - 23 inches

b. Overall width - 17-1/2 inches

c. Overall height - 11-1/2 inches

Shall have a one-piece seamless copper and brass chamber for rust-free performance. Chamber dimensions shall be 8 inches diameter by 14 inches deep.

Shall have a solid-state temperature control and an automatic timer. The sterilizer shall be capable of operating on 120 volts, 60 Hz, 1600 watts, 14 amps.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

The sterilizer shall be fabricated of corrosion-resisting metal with the exception of minor parts such as knobs, handles, controls, insulation, etc. As an alternate, the outer cabinet may be constructed of a combination of high temperature polysulfane, heat resistant ABS or a suitable material, finished to resist corrosion.

The unit shall be capable of completing at least ten sterilization cycles, with load, on one filling of the reservoir. It shall further be designed to return the residual steam and condensate in the chamber to the reservoir following a sterilization cycle. The reservoir shall be completely enclosed within the cabinet. The fill opening for the reservoir shall be located on the top of the cabinet. The reservoir shall have a cover to adequately prevent inadvertent spillage of distilled water during pitch and roll movements when installed aboard naval vessels or other transport modes. A drain connection for emptying the reservoir shall be accessible from the exterior of the cabinet.

The operating temperature range shall be 250°F to 270°F. The sterilizer shall be capable of attaining and maintaining a temperature of 270°F for five minutes, and 250°F for 20 minutes. The sterilizer control system shall control chamber temperature to within plus 5°F minus 2°F of the set point.

Operation and Service Data Manuals shall be furnished as follows:

a. Information content:

The contractor shall furnish manuals, handbooks, and/or brochures containing complete operation, installation, and service/maintenance instructions (including pictures or illustrations, as necessary) with complete schematics and wiring diagrams. All manuals, handbooks, and/or brochures shall be written in the English language, and all schematics and wiring diagrams shall use American electrical and electronics symbols. The manuals shall include electrical data and connection diagrams for all applicable utilities. The instructions shall also contain a complete list of all replaceable parts showing part number, name, and quantity required. All service notes, service memos, etc. pertaining to the equipment and issued subsequent to the printing of the original instructions shall also be included.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

These instructions shall be, as a minimum, those furnished to service engineers (or serviceman) who normally install and service the equipment for the company or distributor. When the system being procured includes, or will operate ancillary equipment, the service data shall include complete instructions and drawings which show interfacing of all system components.

b. Quantities, distributions and schedules:

Two complete sets of the above information shall be packaged with each unit of the required equipment. Where multiple quantities of the same unit are to be delivered to the same ultimate consignee, only a total of two service manuals need to be supplied to each consignee. However, two operational manuals still need to be supplied with each unit.

Workmanship. The sterilizer shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One sterilizer, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one sterilizer

Sampling for tests. Sampling for tests (dimensional), shall be conducted in accordance with MIL-STD-105, with an AQL of 2.0 (percent defective) and an inspection level of II. The unit of product for test purposes shall be one sterilizer.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor Certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	Item Identification
6530-01-351-0363	STERILIZER, SURGICAL INSTRUMENTS AND DRESSING (Autoclave).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6530-2367

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.

2 The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER A-A-54708	2 DOCUMENT DATE (YYMMDD) 24 July 1992

3 DOCUMENT TITLE
STERILIZER, SURGICAL INSTRUMENT AND DRESSING (AUTOCLAVE)

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed.)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, Middle Initial)	b ORGANIZATION	
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7 DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME Defense Personnel Support Center	b TELEPHONE (Include Area Code) (1) Commercial (215) 737-2870 (2) AUTOVON 444-2870
c ADDRESS (Include Zip Code) ATTN: MSE 2800 S. 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.