METRIC

A-A-54712
31 July 1992

COMMERCIAL ITEM DESCRIPTION

VACUUM PUMP, ROTARY, POWER DRIVEN, SMALL

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-V-36457 which is cancelled.

This Commercial Item Description covers requirements for a power driven rotary vacuum pump.

Salient characteristics:

Pump assembly. Shall be a small, power driven rotary vacuum pump suitable for producing a vacuum of 0.0003 mm of mercury. The unit shall consist of a motor, pump, drive belt, electrical cord with integral switch and sufficient amount of oil for initial filling of the oil reservoir. The unit shall be mounted on a suitable metal base and accurately aligned. The motor shall be at least 1/4 HP and shall operate on a nominal 110V, 60 AC. The electrical cord shall be a three conductor cord fitted with a three prong male connector meeting the requirements of ANSI C73.11. The pump casing shall have an intake nipple which shall serve as the connection to the equipment to be evacuated. The intake nipple shall be no smaller than 1/4" pipe or 3/8" OD tubing.

The pump shall have a normal operating speed of 350 RPM and free air displacement of at least 10 liters per minute.

Belt Guard. A belt guard shall be provided with each pump. The guard shall be fabricated of metal and shall cover the outside and top perameters of the V-belt and pulley wheels. The guard shall be so constructed and mounted that it does not interfere with the operation of the pump assembly.

AMSC N/A FSC 4310

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Performance.

Reduction of pressure. The unit shall reduce the pressure to 0.0003 millimeter (mm) of mercury or less.

Manuals. The contractor shall furnish, with each unit, two copies of service data containing installation and maintenance Instructions. The service data shall also include a detailed component parts list with wiring diagrams and drawings. The data shall be descriptive by manufacturer's part number so replacement parts can be ordered without reference to model or serial numbers. This service data shall be submitted to the contracting officer on a pre-award basis to allow incorporation by the contractor of required revisions without delaying delivery in the contract. In addition the contractor shall mail 15 copies of the approved service. data to the contracting officer.

Workmanship. The vacuum pump shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA) One vacuum pump, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one vacuum pump.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, with an AQL of 0.0 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one vacuum pump.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to, the following:

Vacuum. Before testing, the unit shall be subjected to a break in procedure as outlined in the contractors service data. After the pump has been connected to a 25 liter container by means of the intake nipple and rubber tubing, the motor shall be turned on and allowed to run for ten minutes. The pressure in the container shall be determined by a pressure gauge which is calibrated and readable at a pressure of 0.0003 millimeters of mercury.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Contractor Certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulation requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	Item Identification
4310-00-438-4100	VACUUM PUMP, ROTARY, POWER DRIVE, SMALL

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 4310-0191

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOS

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document nur letter should be given

2 The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE: This form may not be used to request copies of documents, nor to request waivers, o requirements on current contracts Comments submitted on this form do not constitute or imply waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER	2 DOCUMENT DATE (Y
	A-A-54712	31 JULY 1992

3 DOCUMENT TITLE

VACUUM PUMP, ROTARY, POWER DRIVEN, SMALL

4 NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as n*

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, Middle Initial)	b ORGANIZATION	
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7 D

8 PREPARING ACTIVITY

a NAME	b TELEPHONE (Include Area Code) (1) Commercial	(2) A
Defense Personnel Support Center	(215) 737-2870	4
c ADDRESS (Include Zip Code) ATTN: MSE 2800 S. 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 D Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, Telephone (703) 756-2340 AUTOVON 289-	

DD Form 1426, OCT 89 *Previous editions are obsolete* 0102 - LF - 009 - 2700 U.S. Government Printing Office 19

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.