INCH-POUND

A-A-54716
10 September 1992

COMMERCIAL ITEM DESCRIPTION

BAG, TUBE FEEDING

The General Services Administration has authorized the use of this Commercial Item Description

This Commercial Item Description covers a sterile, disposable, tube feeding bag.

Salient Characteristics:

The tube feeding bag shall be a sterile, disposable, plastic bag with an integral ice pouch having a drain and pre-attached administration set suitable for enteral nutrition support.

The bag shall have a capacity of 1000 milliliters, graduated at least every 100 milliliters. The top or proximal end of the bag shall be equipped for suspending from an IV pole. The top shall also have a leakproof opening for ease in filling. The bag shall be designed to drain completely, leaving no residual formula in the bag.

Clear plastic tubing shall be attached to the bottom of the tube feeding bag. The tubing shall measure at least 120 inches in length and shall be compressible and compatible with the enteral feeding pump.

The tubing shall have a roller flow clamp allowing not more than 4 drops per minute above the set rate. The tubing shall also have an integral transparent drip chamber for flow rate observation. The end of the tubing shall have a male step-wise tapered adapter that is compatible for leakproof connection to funnel-end feeding tubes.

Plastic components shall meet the U.S. Pharmacoepia (USP) requirements for Class II plastics.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Sterility. The entire assembly shall be sterile in accordance with the U.S. Pharmacoepia (USP) or the American Association of Medical Instrumentation (AAMI). Sterility assurance level shall be no greater than 10^{-6}.

Workmanship. The tube feeding bag shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing thirty-six individually packaged sterile tube feeding bags, as specified, constitutes one unit. The individual package shall be a sealed, peel open type container capable of maintaining sterility of contents unless package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105; inspection level S-3, acceptance number zero. The unit of product for test purposes shall be one tube feeding bag.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. The tests shall include, but shall not be limited to, the following:

Security of roller in clamp body test. The roller clamp shall be removed from the set together with a piece of tubing approximately one inch longer than the clamp body. The tubing shall be present in the roller clamp in its normal position with a 1/2 inch length extending from each end of the roller clamp. The clamp shall be held vertically, with the large end down, over a supporting gap so that the roller if forced out of the body is free to fall. A polished steel rod shall be applied to the top of the roller, so that the roller shall be able to rotate in the course of the test, and the vertical downward force required to separate the roller from the clamp body shall be measured. As an alternate, any position of the clamp is acceptable providing the clamp body is not distorted and the load measurement is accurate. The force required to remove the roller from the clamp shall be at least 8.0 pounds.

Security of assembly test. The security of assembly of the tubing to the bag shall be tested by gradually applying a seven-pound weight to the tubing above the drip chamber while suspending the bag from an IV pole. This load shall be maintained for one minute without evidence of separation of the tubing from the bag.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTES: 1. The tube feeding bag is intended to be compatible with NSN 6515-01-256-2556, Enteral Feeding Pump.

2. The following National Stock Number is covered by this document.

6515-01-125-5530

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA - OSS

Project Number: 6515-5265

Location: BAGTBC1A.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54716	2. DOCUMENT DATE (YYMMDD) 10 September 1992

3. DOCUMENT TITLE
BAG, TUBE FEEDING

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	e. DATE SUBMITTED (YYMMDD)

7. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.