INCH-POUND

A-A-54719
10 August 1992

COMMERCIAL ITEM DESCRIPTION

NEBULIZER-MANIFOLD ASSEMBLY, INHALATION THERAPY APPARATUS

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a disposable, single use, manifold assembly with nebulizer to deliver aerosol therapy for use with a Puritan-Bennett MA-1 ventilator.

Salient characteristics:

The nebulizer-manifold shall be a disposable, single use assembly specifically designed for use with the Bennett MA-1 ventilator to deliver aerosol therapy.

The assembly shall consist of the following components:

One manifold nebulizer mounted approximately 39 inches from the patient adapter

Two aerosol hoses, approximately 39 inches in length with an "F" swivel adapter and cap or "Y" adapter and cap

One aerosol hose approximately 39 inches in length with or without a spirometer vial adapter

One aerosol hose, approximately 39 inches in length with two supply tubes

One breathing circuit valve

The aerosol hoses shall be crush and kink resistant to prevent accidental occlusion and shall be in accordance with ASTM F 1100.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Hoses shall have cuttable cuffs every 5 to 6 inches allowing rapid shortening of aerosol hoses or addition of other accessories.

The circuit shall provide for attachment of a spirometer in accordance with ASTM F 1100 providing a 1-1/4 inch male tapered connection.

The nebulizer shall deliver aerosol particles of which 83 percent are in the range of 0.8 to 6.0 microns. The nebulizer shall be in accordance with ANSI Z79.9.

The reservoir for the nebulizer shall be provided with the following characteristics:

* It shall provide high visibility of medication level in the reservoir.

* It shall provide usable volumetric capacity of at least 5 ml or at most 15 ml and shall be provided with appropriate markings indicating medication level which shall include indications of 1 ml increments at least between 2 and 4 ml inclusive.

* A suction tube shall be positioned at the bottom of the reservoir which shall either have a bulb-tip or be cone shaped.

The assembly shall have some type of aerosol hose securing tie to hold the tubes together until time of use.

The patient connector of the F or Y type adapter shall be standard 15/22 mm in accordance with ASTM F 1054 as required by ASTM F 1100. Hose cuffs shall be standard 22 mm ID in accordance with ASTM F 1054.

The nebulizer manifold assembly, including hoses, shall not leak more than 350 ml/minute of gas when pressurized at 1 psig.

The assembly shall include operating/user instructions which are normally provided to the commercial market. Upon **request, testing protocol shall be furnished with measurements and site of measurements. The provided testing procedures shall allow Government testing to verify percentage of particle size.**

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Each nebulizer-manifold assembly shall be cleaned, assembled, and individually sealed in a plastic bag.

Workmanship. The nebulizer-manifold assembly shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing ten individually packaged assemblies, as specified, constitutes one unit. Each nebulizer-manifold assembly shall be cleaned and sealed in an individual plastic bag which identifies the assembly for single patient use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-1. The unit of product for examination purposes shall be one nebulizer-manifold assembly.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one nebulizer-manifold assembly.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:37Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

6515-01-004-8481

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA - OSS

Project Number: 6515-5395

Location: NEBULC12.DOC

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1 2 3 and 8 In block 1 both the document number and revision letter should be given

2 The submitter of this form must complete blocks 4 5 6 and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE This form may not be used to request copies of documents nor to request waivers or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

1. RECOMMEND A CHANGE:	1 DOCUMENT NUMBER A-A-54719	2 DOCUMENT DATE (YYMMDD) 10 August 1992

3 DOCUMENT TITLE

NEBULIZER-MANIFOLD ASSEMBLY, INHALATION THERAPY APPARATUS

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite if possible Attach extra sheets as needed)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, Middle Initial)	b ORGANIZATION	
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7 DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b TELEPHONE (Include Area Code) (1) Commercial AC 215-737-2870	(2) AUTOVON AV 444-2870
c ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403 Falls Church VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426 OCT 89 Previous editions are obsolete 9879/9

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:37Z
Check the source to verify that this is the current version before use.