FED. SUP CLASS
6640



PART NUMBER	LENGTH* INCHES (cm)	PART NUMBER	LENGTH* INCHES (cm)
MS51418-1	1.6 (4)	MS51418-7	36.0 (91)
MS51418-2	2.0 (5)	MS51418-8	48.0 (120)
MS51418-3	6.0 (15)	MS51418-9	60.0 (150)
MS51418-4	12.0 (30)	MS51418-10	72.0 (200)
MS51418-5	18.0 (46)	MS51418-11	96.0 (250)
MS51418-6	24.0 (61)		

* Tolerance: ± .1 inch (3 mm).

MATERIAL: Rods shall be made of hard aluminum.

DESIGN: Rod shall be .50 inch (13 mm) diameter, centerless ground. Ends shall be cleanly cut and chamfered.

USE: Rods are intended for laboratory frame construction.

NOTES:
1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.
2. FSCM: 22527 and 48619.

USER ACTIVITIES: ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 82

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

APPROVED 20 JUL 1983 REVISED

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		ROD, LABORATORY FRAME	MS51418
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE PROJECT NUMBER 6640-1190

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T18:01Z
Check the source to verify that this is the current version before use.