MILITARY SPECIFICATION

CURETTES, PERIODONTAL (Columbia)

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers double-ended periodontal curettes of Columbia design.

1.2 Classification. Curettes shall be of the following numbers, as specified:

Columbia, No. 2L and 2R.
Columbia, No. 4L and 4R.
Columbia, No. 13 and 14.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications and standards. The following specifications and standards form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specification and Standards (DoDISS) and supplement thereto, cited in the solicitation.

AMSC N/A FSC 6520

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
PPP-B-566	Boxes, Folding Paperboard.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Boxes, Shipping, Fiberboard.
PPP-B-676	Boxes, Set-up.

Military

MIL-P-116	- Preservation, Methods of.
MIL-L-10547	- Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.
MIL-STD-147	- Palletized Unit Loads.
MIL-STD-794	- Parts and Equipment, Procedures for Packaging of.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

24411	Curettes, Periodontal, Columbia.

(Copies of specifications, standards, and drawings required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of the documents which are DoD adopted shall be those listed in the issue of the DoDISS specified in the solicitation. Unless otherwise specified, the issues of documents not listed in the DoDISS shall be the issue of the nongovernment documents which is current or the date of the solicitation.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

ASTM A 473 - Specification for Stainless and Heat-Resisting Steel Forgings.
ASTM B 135 - Specification for Seamless Brass Tube.
ASTM E 18 - Tests for Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials.
ASTM E 92 - Test for Vickers Hardness of Metallic Materials.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.)

(Nongovernment standards and other publications are normally available from the organizations which prepare or which distribute the documents. These documents also may be available in or through libraries or other informational services.)

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence. Nothing in this specification, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 General requirements. The curettes shall be in accordance with applicable requirements of GG-I-526 with the options and additional requirements specified herein.

3.2. Bid sample. Twelve bid samples which are representative of the curettes to be supplied shall be furnished complete with protocols and satisfactory report of tests performed in accordance with the specification to the Defense Personnel Support Center for evaluation by the Military Medical Services. When tested by periodontists under conditions of normal usage, the curettes shall be suitable for use in periodontal surgical procedures. The curettes shall be well balanced and the blade shall be finished and contoured to produce a sharp edge around the blade which shall effectively accomplish subgingival curettage.

3.3 Material.

3.3.1 Operative ends. The operative ends shall be class 4, type 440A, 440B, or 440C stainless steel.

3.3.2 Handle. Handle A and D shall be plated brass. Brass shall conform to class 8, copper alloy number 260 or 330, as specified in ASTM B 135 for chemical composition only. The brass handles shall be chromium over nickel plated. Handle B shall be fabricated of martensitic stainless steel, type 416 or 410, as specified in GG-I-526, or ferritic stainless steel conforming to 430F stainless steel in accordance with ASTM A 473 for composition only. Handle C shall be fabricated of austenitic stainless steel, type 303, as specified in GG-I-526. The corrosion tests, as specified in 3.8, shall apply as applicable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

3.4 Hardness. After proper stress relieving, the Rockwell hardness of the operative ends shall be C52 to C56 for type 440A, C54 to C58 for type 440B, and C56 to C60 for type 440C. Handles and shanks shall be of proper temper and shall have a hardness suitable for the purpose intended. Hardness shall be determined as specified in 4.4.1.

3.5 Style, design, and dimensions. The curettes shall be double-ended Columbia curettes. Curettes shall be furnished in the applicable number specified (see 1.2). The curettes shall be fine bladed and suitable for use in subgingival curettage, removal of serumal calculus from root surfaces, and for thoroughly polishing the cementum to assist in physiological tissue reattachment during periodontal treatment. The curettes shall be furnished with the handle and shank design of either Handle A, B, C or D. The style, design, and dimensions of the curettes shall conform to Defense Personnel Support Center (DPSC) Drawing 24411 for the applicable number.

3.6 Construction. The curettes shall be of three piece construction consisting of a handle and an operative end at each end. Handle shall be of solid or hollow construction as specified on the drawing for the respective handle. Knurled or serrated portions of handles shall be as specified on the drawing for the respective handle. The operative ends shall be permanently secured to the handle. Junction of operative end to handle shall be smoothly finished and free of crevices. There shall be no movement, loosening, or separation between the operative end and the handle when the operative end is subjected to an 85 pound (minimun) pull out force, when tested as specified in 4.4.2.

3.6.1 Drawings. The drawings forming a part of this specification are not production drawings. Drawings establish dimensional limitations and control design and configuration. The supplier shall be responsible for preparing his own shop drawings. The supplier shall select dimensions and tolerances within the limitations prescribed on the drawings to provide a product with proper design, configuration, and function. No deviation from the dimensional limitations prescribed on the drawings is permissible without prior approval of the contracting officer.

3.7 Finish. The operative ends, including shanks, shall have a smooth mirror or smooth bright finish, except for the surfaces that form the cutting edges which shall have a finely honed finish. The operative ends (entire blade) shall have a sharp cutting edge and shall be free of defects, such as coarse grind marks, cracks, gouges, burrs, feathers, or other imperfections which affect curettaging performance, when viewed as specified in 4.4.3. Honed portion of operative end shall be blended at junction of shank to form a smooth contour. Finger grip portions of handles shall be serrated or knurled. Handles shall have a mirror or bright finish, except for the serrated or knurled portions. Finger grip portions of handles shall be free from barbs and shall provide adequate finger grip during use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

3.8 Corrosion resistance. The copper sulfate test requirements of GG-I-526 shall apply to the operative ends, martensitic stainless steel handles, and ferritic stainless steel handles. The boil test requirements of GG-I-526 shall apply to the entire curette (operative ends and all handles). The salt spray test specified for brass composition in GG-I-526 shall not apply.

3.9 Identification marking. In addition to the requirements of GG-I-526, each curette shall be permanently and legibly marked with identification marking identifying the curettes as Columbia curettes with the respective Columbia instrument numbers, as applicable. Each instrument shall be marked on the front face. All marking shall be on the smooth portion of the handle. Each curette shall be marked with the name or registered trademark of the manufacturer (and contractor, if other than the manufacturer). The word "stainless", when applicable, shall also be marked on instruments of foreign manufacture. As an alternate to the word "stainless", the use of a registered trade name identifying the steel as stainless shall be acceptable, provided prior approval is received from the cognizant contracting officer.

3.10 Workmanship. In addition to the requirements of GG-I-526, the operative ends shall be free of deep pits, discoloration, and rounded edges.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification is defined as both examination (such as visual and auditory investigation without the use of special laboratory applicances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 202610-21T20:10Z
Check the source to verify that this is the current version before use.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the raw material, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or suppler, the NSN, the Item Identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in tables I, II, III (see 4.2.1, 4.2.2, 4.3), and 4.3.1.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQLs) shall be as indicated in table I. The unit of product for examination shall be one curette.

TABLE I. Sampling for examination.

Type defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	4.0

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQLs) shall be as indicated in table II. The unit of product for test purposes shall be one instrument. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

TABLE II. Sampling for tests.

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL
Entire instrument	Composition	3.3.1, 3.3.2, & GG-I-526	GG-I-526	S-1*	**
Operative end	Hardness	3.4 and GG-I-526	4.4.1	S-2	1.5
Operative end	Dimensions	3.5	SMI	S-2	2.5
Handle, shank, Overall instrument	Dimensions	3.5	SMI ***	S-4	4.0
Operative end and handle	Pull test	3.6	4.4.2	S-2	1.0
Cutting edges	Magnification inspection	3.7	4.4.3	I	**
Operative end and handle (martensitic and ferritic)	Copper sulfate test	3.8	GG-I-526	S-2	1.0
Entire instrument	Boil test	3.8	GG-I-526	S-2	1.0

* A sufficient amount of each lot of metal used in the construction of the curettes may be selected for composition analysis in lieu of sample curettes, provided lot identity has been maintained and can be demonstrated to the Government.

** Acceptance number shall be zero for all sample sizes.

*** Measurements shall be taken utilizing Standard Measuring Instruments (SMI).

4.3 Examination. The curettes shall be examined to determine compliance with specification requirements. Examination shall be conducted in accordance with the following classification of defects. Examination shall include, but shall not be limited to, the defects listed in table III.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects.

Categories	Defects
Major	
101	Style and design not as specified for the applicable number curette.
102	Instrument not constructed as specified.
103	Operative end finish not as specified.
104	Honed portion of operative ends not blended at junction of shank to form a smooth contour.
Minor	
201	Instrument not permanently and legibly marked with name or trademark of the contractor.
202	Instrument identification not present, not properly located, or not complete.
203	Curette not free from dirt, oil, compound, or foreign matter.
204	Handle serrations or knurling missing, not uniform, or not complete.
205	Junction of handle not smoothly finished and free of crevices.

4.3.1 Inspection of packaging. The sampling and inspection of the preservation and interior package marking shall be in accordance with groups A and B quality conformance inspection requirements of MIL-P-116. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with the requirements of this specification (see 4.2.2.).

4.4.1 Hardness. Hardness shall be determined in accordance with ASTM Designation E 18 (Rockwell) or E 92 (Vickers or Diamond Pyramid).

4.4.2 Pull test. The handle and shank of the operative end shall be securely held in separate chucks or other holding devices. A slowly increasing pull out force, increasing from zero to 85 pounds (minimum), shall be applied to the shank of the operative end. The force shall be maintained at the high value for a minimum of 1 second. The instrument shall then be examined for movement, looseness, or separation of the operative end with respect to the handle. The test shall be repeated for each operative end of the double-ended instrument.

Source: https://assist.dla.mil -- Downloaded: 202610-21T20:10Z
Check the source to verify that this is the current version before use.

4.4.3 Magnification inspection. The operative ends (blade portion) shall be inspected under four power (4X) magnification.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C, as specified.

5.1.1 Level A.

5.1.1.1 Cleaning. The curettes shall be cleaned in accordance with MIL-P-116.

5.1.1.2 Drying. The curettes shall be dried in accordance with MIL-P-116.

5.1.1.3 Preservative applications. Preservatives shall not be used.

5.1.1.4 Unit package. Each curette shall be packaged in accordance with MIL-P-116, submethod IC-2. As an alternate, each curette shall be packaged in a rigid plastic container having a friction, screw cap or plug type closure. The curette shall be immobilized within the container and the operative ends protected with a suitable cushioning medium.

5.1.1.5 Intermediate package. Twelve unit packages containing curettes of one style only shall be packaged in a box constructed in accordance with PPP-B-566, PPP-B-636, class domestic, or PPP-B-676. Closure shall be as specified in the applicable box specification.

5.1.2 Level C. The level C preservation for curettes shall conform to the MIL-STD-794 requirements for this level.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.1).

5.2.1 Level A. Twenty-four intermediate packages (288 units) containing curettes of one style only shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type, PPP-B-621, class 2 or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

5.2.2 Level B. Twenty-four intermediate packages (288 units) containing curettes of one style only shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type, PPP-B-621, class 1 or PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

5.2.3 Level C. The level C packing for curettes shall conform to the MIL-STD-794 requirements for this level.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination equal 40 cubic feet or more. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 40 cubic feet need not be unitized. Unitized loads shall be uniform in size and quantities to the greatest extent practicable.

5.2.5.1 Levels A and B. The curettes shall be packed as specified in 5.2.1 and 5.2.2 and shall be unitized on pallets conforming to MIL-STD-147.

5.2.5.2 Level C. The curettes shall be packed as specified in 5.2.3 and shall be unitized as specified in MIL-STD-794.

5.3 Marking.

5.3.1 Levels A, B, and C. Each unit pack, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.3.1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data.

6.1.1 Acquisition requirements. Acquisition documents should specify the following:

a. Title, number and date of this specification.
b. Number required (see 1.2).
c. Selection and applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 National Stock Number (NSN) coverage. This specification covers the following items listed in the Federal Supply Catalog:

National Stock Number	Item Identification
6520-01-046-7334	CURETTE, PERIODONTAL, Columbia University, No. 2L and 2R.
6520-01-046-7335	CURETTE, PERIODONTAL, Columbia University, No. 4L and 4R.
6520-01-047-9801	CURETTE, PERIODONTAL, Columbia University, No. 13 and 14.

6.3 Supersession data. This specification includes the requirements of Specification Data Sheet No. 1, dated 31 May 1979 for NSNs 6520-01-046-7334/7335, and 6520-01-047-9801, CURETTES, PERIODONTAL, Columbia.

6.4 Subject term (key word) listing:
Columbia, No. 2L and 2R
Columbia, No. 4L and 4R
Columbia, No. 13 and 14
Periodontal Curettes

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

Review Activities:

Army - MD
Navy - MS
Air Force - 03

Project No. 6520-1881

Civil Agency Coordinating Activities:

VA-OSS
PHS
FDA-MPQAS

05895

Location: NCR Disk 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
hiladelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 202610-21T20:10Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 202610-21T20:10Z
Check the source to verify that this is the current version before use.