MIL-L-42006A (DM)
24 March 1987
USED IN LIEU OF
MIL-L-42006
4 August 1986

MILITARY SPECIFICATION

LIGHT, DENTAL, OPERATING

This specification is approved for use by the Defense Personnel Support Center based upon currently available technical information but it has not been approved for promulgation as a revision of MIL-L-42006. It is subject to modification. However, pending its promulgation as a coordinated military specification, it may be used in acquisition.

1. SCOPE

1.1 Scope. This specification covers a dental operating light suitable for use in dental procedures.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications and standards. The following specifications and standards form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of these documents are those listed in the issue of the Department of Defense Index of Specifications and Standards (DoDISS) and supplement thereto, cited in the solicitation.

SPECIFICATIONS

MILITARY

MIL-M-7298 - Manual, Technical, Commercial Equipment.

AMSC N/A FSC 6520

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:09Z
Check the source to verify that this is the current version before use.

STANDARDS

FEDERAL

FED-STD-H28 - Screw-thread Standards for Federal Services.

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.

MIL-STD-889 - Dissimilar Metals.

(Copies of specifications, standards, and drawings required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity).

2.2 Non-Government publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of the documents which are DoD adopted are those listed in the issue of the DODISS specified in the solicitation. Unless otherwise specified, the issues of documents not listed in the DODISS are the issues of the non-Government documents which are current on the date of the solicitation.

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION (ISO)

ISO/TC 106/SC 6/WG 1N 32 AUGUST 1985
Technical Committee ISO/TC 106 "Dentistry" Dental Operating Light
New British Proposal

(Application for copies should be addressed to the American National Standards Institute, 1430 Broadway, New York, New York 10018).

UNDERWRITERS LABORATORY, INC. (UL)

UL 544 Standard for Safety Medical and Dental Equipment

(Application for copies should be addressed to the Underwriter's Laboratories, Inc., 207 East Ohio Street, Chicago, IL 60611, 333 Pfingsten Road, Northbrook, IL 60062, 1285 Walt Whitman Road, Melville, LI, NY 11746, 1655 Scott Blvd., Santa Clara, CA 95050).

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification takes precedence. Nothing in this specification, however, supersedes applicable laws and regulations unless a specific exemption has been obtained.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:09Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Materials. Materials shall be as specified herein. When materials are not specifically designated, those used shall be new, be of the same or better quality as used in commercial practice, and entirely suitable for the purpose intended.

3.1.1 Dissimilar metals. Contact between dissimilar metals likely to cause deterioration of parts by galvanic corrosion shall be avoided. Where such contact cannot be avoided, suitable protection against galvanic corrosion shall be applied in accordance with MIL-STD-889.

3.2 Safety requirements.

3.2.1 The light shall meet the applicable requirements of UL-544 and the following requirements specified herein.

3.2.2 All electrical parts shall be located or enclosed to prevent accidental contact with live parts.

3.2.3 Electrical circuits shall be properly fused.

3.2.4 Electrical light bulb shall be shielded to prevent injuries to the operator or patient in the event the bulb should accidentally shatter.

3.2.5 The on/off switch and intensity control shall be easily accessible.

3.2.6 Electrical circuit contained within or about the luminaire and the adjustable arms shall be low voltage (not to exceed 25 volts).

3.2.7 Fire and casualty hazards. The contractor shall provide proof, to the procurement contracting officer, not less than thirty days prior to delivery, that the lights to be delivered under the contract conform to the applicable requirements of Underwriters' Laboratories, Inc. (UL) Standard of Safety 544. A copy of the test report shall be made available to the Government inspection personnel for their use in determining that the lights being submitted for acceptance are identical to the light submitted for approval. Proof shall consist of one of the following:

a. UL Listing and UL Label affixed to the product, provided the contractor agrees to have Underwriters' Laboratories, Inc. forward a copy of the UL report directly to the contracting officer, and the Government inspection personnel, upon request.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:09Z
Check the source to verify that this is the current version before use.

b. A label or equivalent means of identification affixed to the product as evidence that the product design has been examined and tested for design safety compliance by a nationally recognized independent laboratory.

If an independent laboratory is used, it must be adequately equipped and competent to determine compliance with UL Standards. The Government reserves the right to require a facilities report and to conduct a facilities survey. In support of the label/product identification, the independent testing laboratory shall provide surveillance over the product manufacturing system to the degree necessary to insure that production units are manufactured in the same manner and with the same type of components as the test sample evaluated, and upon which the validity of the product identification is based. The testing laboratory shall develop a detailed test report which includes a conclusion as to whether the test sample evaluated is "approved" or "not approved" as complying with all requirements of the standards of safety.

3.3 First article. When specified in the contract or purchase order, a sample shall be subjected to first article inspection (see 4.3 and 6.1.1).

3.4 Standard commercial product. The light shall, as a minimum, be in accordance with the requirements of this specification and shall be the manufacturer's standard commercial product. Additional, or better features which are not specifically prohibited by this specification but which are a part of the manufacturer's standard commercial product shall be included in the light being furnished. A standard commercial product is a product which has sold or is being currently offered for sale on the commercial market through advertisements, or manufacturer's catalogs or brochures, and represents the latest production model.

3.5 Style and design. The light shall be an adjustable operating light suitable for illumination of the oral cavity and surrounding structures at variable intensity settings. The light shall consist of a light head, adjustable arm, controls, and a suitable mounting attachment. The light shall be equipped with handles for ease in positioning. Handles shall facilitate both right and left handed operation. The light shall be stable and free of drift following adjustment. All pivots, bushings, and bearings shall be self lubricating. The light shall include an adjustment mechanism for the articulated joint supporting the light head. All such adjustments shall be simply and easily performed in situ. The light shall be adaptable to chair mounting or unit light post mounting, ceiling mounted-pedestal, ceiling mounted-track, or wall mounted, as specified, (see 6.1.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:09Z
Check the source to verify that this is the current version before use.

3.5.1 Arm lengths. The arm lengths and arm movements shall provide mobility for placement and rotation of the light head for positioning the light head to view the oral cavity during dental procedures.

3.5.2 Light head. The light head shall be light and easily positioned. The light head shall feature a multifaceted reflector or other suitable means of simulating a multi-point light source so as to minimize the effect of blocking shadows.

3.5.2.1 The light head shall be adjustable so as to illuminate all areas of the oral cavity. It shall be capable of being positioned at all points within the base of a spherical segment whose radius is equal to the focal length of the light, whose vertex is equal to a 90 degree solid angle located at the imaginary center of the oral cavity, and whose axis is coincident with a sagittal axis through the imaginary center with the patient reclined in a dental chair with the chair back at 45 degrees when tested as specified in 4.4.3.1.1.

3.5.3 Illumination pattern. The light shall provide a consistent, even, rectangular illumination pattern of at least 6 centimeters high by 12 centimeters wide at the focal distance specified by the manufacturer (see 3.5.5) when measured as specified in 4.4.3.1.2.

3.5.4 Illumination intensity. The light shall feature no less than three (3) intensity settings such that the lowest intensity is 800 to 1200 foot candles and the highest intensity is no less than 2000 foot candles at the specified optimum focal length between 25 and 32 inches from the oral cavity. Intermediate settings shall be at approximately equal increments. As an alternate, an infinitely variable intensity control shall be acceptable.

The greatest illumination intensity shall be in the center of the pattern and the illumination intensity shall decrease toward the pattern edge such that the intensity difference between any two points one (1) centimeter apart within a central 5 centimeters by 10 centimeters pattern shall not be greater than 30 percent of the maximum intensity along the minor axis nor greater than 20 percent along the major axis.

No area of light intensity within a 5 centimeters by 10 centimeters area about the pattern's center shall be less than one-half of the greatest light intensity.

The illumination intensity 6 centimeters and greater above a horizontal line through the center of the pattern shall be not more than 200 foot candles at the highest intensity setting so as to prevent patient discomfort.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:09Z
Check the source to verify that this is the current version before use.

Illumination measurements shall be made in accordance with 4.4.3.1.3.

3.5.5 Focal distance. The focal distance shall be between 25 and 32 inches from the oral cavity. The light shall have a convenient and accurate means to focus the light at the specified focal distance.

3.5.6 Color properties. The light shall be capable of providing a color corrected, natural white light suitable for critical shade discrimination characterized by a correlated color temperature in the range from 3800 to 5500 degrees Kelvin, with a broad spectrum output containing all visible wavelengths. The light emitted shall have a color within an area described by a rectangle on the Commission International De L' Eclairage, CIE 1976 Uniform Chromaticity Diagram, the range of coordinates to describe the rectangle being defined by the following U' and V' values.

U'	V'
0.2350	0.4975
0.2250	0.5225
0.1998	0.4925
0.2158	0.4700

Color properties shall be measured as specified in 4.4.3.1.4.

3.5.7 Temperature rise. The light shall be designed for heat reduction at the patient. The temperature shall not rise more than 5 degrees centigrade above ambient at any point in the working light pattern with the light at the 1500 ± 25 foot candle setting when tested as specified in 4.4.3.1.5.

3.5.8 Overall operation of the light. The light (all operating components, adjustments, and controls) shall operate in an efficient manner and shall show no signs of defects or malfunctioning, when tested as specified in 4.4.3.1.6.

3.5.9 Bulb replacement. The service life of the bulb shall be no less than 2000 hours at the high intensity setting. The light bulb shall be field replaceable with no tools or with tool provided by the manufacturer. If a tool is furnished by the manufacturer, there shall be a means of attaching the tool to the control arm.

3.6 Electrical characteristics. The electrical characteristics shall be as follows, as specified (see 6.1.2).

a. 115 ± 10% volts, 60 ± 2 Hz.
b. 115 ± 10% volts, 50 ± 2 Hz.
c. 230 ± 10% volts, 50 ± 2 Hz.
d. 230 ± 10% volts, 60 ± 2 Hz.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:09Z
Check the source to verify that this is the current version before use.

The light shall be furnished with an electrical power cable, comprised of the power cable of the light and an extension cable of sufficient length to accommodate a light post in a normal operating system.

3.7 Finish. All surfaces shall be clean and free of foreign matter, flux, or welding residue. The finish of components shall be as specified. If no finish is specified the finish shall be entirely suitable for the purpose intended. Finish on edges and surfaces shall be uniform and free from burrs, slivers, sharp edges, cracks, or other defects. All exposed surfaces of the light shall be smooth and made of materials which can withstand repeated wipe down with ADA acceptable disinfectants without deterioration of function or esthetics.

3.8 Plating (coatings). Plating and coatings shall be uniform in thickness and appearance; shall be free from blisters, peeling, visible porosity, or other defects. The base metal to be plated shall be free from flaws or defects which will be detrimental to the final finish and shall be subjected to such undercoating, polishing, cleaning, dipping and plating procedures as are necessary to yield deposits with the desired function, quality, and appearance.

3.9 Workmanship. Workmanship shall be in accordance with best commercial practice. All components shall be clean, well finished, and free from defects which detract from their appearance or impair their serviceability.

3.9.1 Casting. Castings shall be sound and free from repairs, warping, or other defects that may render the casting unsound for use. Repair processes, such as welding, peening, plugging, or filling with cold solder or metallic paste shall not be permitted.

3.9.2 Welding, brazing and soldering. All welds, brazing, and soldering shall be durable and sound in every detail. Welding shall be in accordance with the latest practices recommended by the American Welding Society. Components shall be joined without any resultant distortion and shall be free from twists or strains. All welds shall be produced by qualified welding operators with satisfactory welding equipment using satisfactory fill metal. Surfaces to be welded shall be free from grease, oil, oxides or any other foreign matter. Any cleaning method which will remove these substances and leave the surface clean is satisfactory. All aluminum parts shall be welded within 48 hours after a chemical cleaning treatment. Parts must be chemically recleaned if the 48 hour time period is exceeded. All weld joints shall blend into the adjacent metal in gradual, smooth curves. Beads shall be smooth and free of slag, excessive undercut, or excessive spatter. Sufficient metal shall be added to provide a suitable fillet or reinforcement unless otherwise specified. Welds shall not be oxidized. All welds shall be continuous with no visible voids, pin holes, or cracks. Exposed welds shall be smooth and blend with adjacent surfaces.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:09Z
Check the source to verify that this is the current version before use.

All welds shall be cleaned, discoloration removed, and be free from slag and spatter before finishing. Welds shall be ground, only to the degree necessary to remove sharp edges or projections which may cause defects to adjacent surfaces or can be injurious to the touch.

3.9.3 Machine work. Machining shall be in accordance with best commercial practice. All bolt, machine screw and rivet holes shall be accurate and clean. The holes for screws and bolts shall be sized appropriately for the size screw or bolts used. All bolts, nuts, screws, and fasteners shall be completely fastened and tight.

3.9.4 Metal fabrication. Metals (sheets, plates and tubing) shall be free from dents, folds, or distortions. Material shall be straightened by methods that will not cause injury to the metal. Operations (such as bending, shearing, or cutting) shall be done neatly and accurately. Metal forming shall be made with controlled means to insure uniformity of size and shape.

3.9.5 Hardware. All hardware shall be entirely suitable for the purpose intended and shall be completely compatible with the material to which it is attached or which it joins including plated material. Any fastening device required to be loosened, tightened, or removed for repair, maintenance or adjustment shall not be swaged, peened, staked, or otherwise deformed to prevent loosening of the device. All rotative fasteners shall be correctly torqued. All threaded parts shall comply with the Federal Standard H-28.

3.10 Operation of the light. The assembled light (all operating components, adjustments, switches, and controls) shall operate quietly, smoothly, efficiently and shall show no signs of defects or malfunctioning, when tested as specified in 4.4.3.1.6.

3.11 Color. The color of the light shall be cream white.

3.12 Identification marking. Each light shall have one or more name plates, permanently fastened and reasonably accessible, bearing the name of manufacturer, type and model designation, serial number, and electrical characteristics. The serial number shall be traceable by the manufacturer to the date of manufacture.

3.13 Interchangeability. All parts having the same manufacturer's part number shall be functionally and dimensionally interchangeable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:09Z
Check the source to verify that this is the current version before use.

3.14 Tolerances. Where tolerances are not specified, standard commercial tolerances shall be permissible. Tolerances shall be held to a minimum where accuracy is of paramount importance for proper operation and interchangeability of component parts. Assemblies shall be precision-fitted for durable, smooth, and consistent functioning.

3.15 Service data. The contractor shall furnish, with each unit, two copies each of a technical manual (installation manual and maintenance/repair manual, and an operations manual) which contains complete instructions for installation, operation, and maintenance and lists of component parts. The manual shall comply with the requirements of MIL-M-7298.

The installation manual and maintenance/repair manual shall include the following: Step by step illustrated installation instructions, list of special tools (if required), exploded drawings with parts identified by part number, functional test procedures, electric utility requirements, electrical schematic diagrams, checkout procedures to ensure light is functioning properly, instructions on proper focusing of the light, telephone number for technical assistance and a trouble-shooting diagram.

The operating manual shall include the following: Step by step procedures for operation of the light including illustrations of the location and function of each control, step by step instructions for bulb replacement and focusing of the light, and instructions for adjusting arm counter balance tension and adjustments of the tension of the light head articulation, instructions for the proper cleaning and disinfecting of all exposed surfaces; and a daily functional checklist.

As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the Contracting Officer via the cognizant Government inspector, two preliminary copies of his proposed technical manual, for review and comment by the Directorate of Medical Materiel, Defense Personnel Support Center. The preliminary manuals shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract or purchase order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:09Z
Check the source to verify that this is the current version before use.

In addition, the Contractor shall furnish 15 copies of the approved technical manual to the Contracting Officer for distribution. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of contract and propose to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon request to the Contracting Officer. Waiver of submission of 15 additional copies shall be at the option of the Government.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to ensure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification is defined as both examination (such as visual and auditory investigation of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials may be furnished in lieu of actual performance of such testing by the contractor provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the manufacturer or supplier, the Item Identification, the name of the Contractor, the contract number, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements. Certificates of quality may be furnished for the following: material composition, hardness, tempers, torque requirements on springs, machining and casting requirements (such as flatness, concentricity, and finish), and plating or coating.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:09Z
Check the source to verify that this is the current version before use.

4.2 Classification of inspection. The inspection requirements specified herein are classified as follows:

a First article inspection (see 4.3).

b Quality conformance inspection (see 4.4).

4.3 First article inspection. When a first article is specified in the contract, the first article inspection shall consist of all examinations specified herein, and all testing specified herein and shall apply to the first article sample only (see 6.1.1). The failure of compliance with any examination or test shall be cause for rejection of the first article.

4.4 Quality conformance inspection. Except as otherwise specified herein, sampling for inspection shall be performed in accordance with the provisions set forth in MIL-STD-105.

4.4.1 End item examination.

4.4.1.1 Visual examination. End item examination shall be conducted in accordance with the following classification of defects. Examination shall include, but shall not be limited to, the defects listed in table I. The inspection level shall be II with an acceptable quality level (AQL) of 2.5 for major defects and an AQL of 10 for minor defects. The unit of product shall be the component of an end item or an end item, as applicable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:09Z
Check the source to verify that this is the current version before use.

TABLE I. End item classification of defects.

	Major	Minor
Design and Construction		
Design and construction of light not as specified.	X	
Light head		
Light head components damaged (reflector, lens, etc).	X	
Light shield missing or damaged.	X	
Bulb missing or damaged.	X	
Color		
Color not as specified.	X	
Controls		
Controls broken or damaged.	X	
Hardware		
Components such as pins, screws and fasteners missing, broken, stripped, cracked, or deformed.	X	
Fasteners loose.		X
Finish		
Finish not as specified.	X	
Surfaces not clean, not free of foreign matter, flux, spatters, or other defects.		X
Surfaces of lens, reflectors scratched or chipped.	X	
Damage or defects on exterior or interior surfaces such as burrs, slivers, cracks, sharp edges, which affect function or serviceability.	X	
Plating or coating missing (affecting function).	X	
Plating or coating missing (not affecting function).		X
Plating not uniform in appearance.		X
Plating not free from blisters, peeling,		

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:09Z
Check the source to verify that this is the current version before use.

TABLE I. End item classification of defects (continued).

	Major	Minor
visible porosity or other defects (affecting function or serviceability).	X	
Plating not free from blisters, peeling, visible porosity or other defects (not affecting function or serviceability).		X
Workmanship		
Components not free from coarse machine tool or die marks which affect function or serviceability.	X	
Any component fractured, broken, punctured, torn, bowed, deteriorated, or malformed.	X	
Any component dented or creased.		X
Any component misplaced or not in proper alignment.	X	
Any component not connected or not joined properly.	X	
Fastening device that requires to be loosened or removed is swaged, peened, staked or otherwise permanently fastened.	X	
Components missing.	X	
Components do not fit or mate properly.	X	
Interface fits between components not proper (too loose, too tight, binding).	X	
Components not free from defects which may impair their serviceability.	X	
Components not free from defects which do not impair their serviceability.		X
Removal components cannot be removed and replaced without difficulty.	X	
Components not properly assembled or aligned.	X	
Identification Marking		
Identification marking missing, incomplete, not legible.		X
Identification marking incorrect or not permanent.		X
Caution plates (where applicable) missing, incomplete, not legible.		X
Caution plates (where applicable) not permanently secured.		X
Marking not properly located.		X
Service Data		
Service data not included.	X	

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:09Z
Check the source to verify that this is the current version before use.

4.4.2 Dimensional examination. Dimensional examination shall be conducted for compliance with dimensions specified. Any dimension not within specification tolerance shall be a defect. The inspection level shall be S-3 with an AQL 2.5. Unit of product shall be the component of an end item or an end item, as applicable.

4.4.3 End item testing. End item testing shall be conducted in accordance with MIL-STD-105 and as indicated in table II, except that the acceptance number shall be zero for all sample sizes. A failure during 100 percent inspection shall be cause for rejection of the component or item only (not the entire lot). Unit of product shall be one light.

TABLE II. Sampling for tests

Characteristic	Requirement	Test Procedure	Inspection level
Light head adjustability	3.5.2.1	4.4.3.1.1	S-2 1/
Illumination pattern	3.5.3	4.4.3.1.2	S-2
Illumination intensity	3.5.4	4.4.3.1.3	S-2
Color properties	3.5.6	4.4.3.1.4	S-2
Temperature rise	3.5.7	4.4.3.1.5	S-2
Operation test	3.5.8	4.4.3.1.6	100%

1/ Minimum sample size shall be three (3).

4.4.3.1 Test. Tests shall be conducted to determine compliance with specification requirements. Tests procedures and conditions shall be as specified in proposed ISO/TC 106 except as specified herein. The shadow test shall not apply. All tests and measurements shall be conducted at the voltage specified by the manufacturer, which shall be between 110 volts and 120 volts and at the focal distance specified by the manufacturer, which shall be between 25 and 32 inches. If none are specified, a distance of 70 cm (27.6 inches) and a voltage of 115 volts shall be used.

Voltage shall be a stabilized voltage within plus or minus one (1) percent of specified voltage. The checking of the uniformity of gradients within the central 50 mm diameter circle in accordance with proposed ISO/TC 106 shall be made at 1 centimeter intervals.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:09Z
Check the source to verify that this is the current version before use.

4.4.3.1.1 Light head adjustability. The light (lamp head and arms) shall be mounted on light post (normal dental operating condition) or ceiling or wall mounted (when specified, see 6.1.2). The light head shall be moved and positioned as specified in 3.5.2.1 with the light remaining focused on the center of the oral cavity.

4.4.3.1.2 Illumination pattern. The illumination pattern of the light shall be measured with the light directed on a flat object or screen. Measurement shall be made at each intensity setting.

4.4.3.1.3 Illumination intensity. Except as specified herein, illumination intensity shall be measured in accordance with proposed ISO/TC 106.

4.4.3.1.4 Color properties. Except as specified herein, measurements of color properties shall be made in accordance with proposed ISO/TC 106. Three illuminance setting values (low, medium, high) shall be used.

4.4.3.1.5 Temperature rise. Except as specified herein, measurements of temperature rise shall be made in accordance with proposed ISO/TC 106.

4.4.3.1.6 Operation test. The light shall be operated to determine the proper functioning of all components, adjustments, switches, and controls. Operation shall include, but not limited to, the following:

1. On-off switch.
2. Intensity switch and controls.
3. Arm movements (smooth and drift free).
4. Ease of bulb replacement.
5. Removability of lens shield.
6. Light head movements (smooth).

5. PACKAGING

5.1 Preservation. Preservation shall be level C.

5.1.1 Level C. Each dental operating light shall be packaged and cushioned in accordance with the suppliers commercial practice in such a manner to prevent damage during shipment and storage.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:09Z
Check the source to verify that this is the current version before use.

5.2 Packing. Packing shall be level C.

5.2.1 Level C. The packaged dental operating light shall be packed in a substantial commercial exterior container (export when applicable) which will afford adequate protection during shipment from the supply source to the receiving activity. This pack shall conform to the applicable carrier's rules and regulations.

5.3 Marking.

5.3.1 Interior packages. Each package shall be marked to identify contents and the quantity.

5.3.2 Exterior container. Each exterior container shall be marked as specified in MIL-STD-129.

6. NOTES

6.1 Ordering data.

6.1.1 Aquisition requirements. Aquisition documents should specify the following:

a. Title, number and date of this specification.
b. Options required (see 6.1.2.)
c. First article, when required (see 3.3).
d. Approval of colors (see 3.11.1).

6.1.2 The following options are listed to allow the user to choo items required to ensure the operatory light meets the needs of the individual facility. In each category several options are given. I no option is checked, the first option listed will be the default value.

(a) Electrical requirements:

a. 115 ± 10% volts 60 ± 2 Hz
b. 115 ± 10% volts 50 ± 2 Hz
c. 230 ± 10% volts 50 ± 2 Hz
d. 230 ± 10% volts 60 ± 2 Hz

(b) Type of mount: (1) chair mounted; (2) ceiling mounted-pedestal; (3) ceiling mounted-track; (4) wall mounted.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:09Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force - 03

PREPARING ACTIVITY:

DLA-DM

Agent:

DoD-MB

Project No. 6520-1959

Location: NCR DISK #5

Civil Agency Coordinating Activities:
VA-OSS
PHS

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:09Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4086 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:09Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
MIL-L-42006A(DM)	LIGHT, DENTAL OPERATING

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)* _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 1426 82 MAR

PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:09Z
Check the source to verify that this is the current version before use.