LIMITED MILITARY SPECIFICATION

PIN, THREAD-FORMING, DENTAL

This specification is approved for use within DLA, Defense Personnel Support Center (DM), and is available for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers self-threading pins suitable for pin retention of dental restorative materials.

1.2 Classification. Pins shall be of the following types and sizes, as specified.

Type I - Self-threading

Size 0.024 inch
Size 0.031 inch

Type II - Self-threading, Self-shearing

Size 0.0145 inch
Size 0.019 inch
Size 0.031 inch

Type III - Self-threading, Self-shearing, Two-in-One

Size 0.024 inch
Size 0.031 inch

AMSC N/A FSC 6520

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RSTS, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications and standards. The following specifications and standards form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DoDISS) and supplement thereto, cited in the solicitation.

SPECIFICATIONS

Federal

GG-I-526	- Instruments, Dental and Surgical; General Specifications.
PPP-B-566	- Boxes, Folding, Paperboard.
PPP-B-601	- Boxes, Wood, Cleated-Plywood.
PPP-B-621	- Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	- Boxes, Shipping, Fiberboard.
PPP-B-676	- Boxes, Setup.

Military

MIL-P-116	- Preservation, Methods of.
MIL-L-10547	- Liners, Case and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Federal

FED-STD-151	- Metals; Test Methods.

Military

MIL-STD-105	- Sampling Procedures and Table for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.
MIL-STD-147	- Palletized Unit Loads.
MIL-STD-794	- Parts and Equipment, Procedures for Packaging of.

DRAWING

Defense Personnel Support Center (DPSC)

23205 Pin, Thread-Forming, Dental

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

(Copies of specifications, standards, and drawing required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of the documents which are DoD adopted shall be those listed in the issue of the DoDISS specified in the solicitation. Unless otherwise specified, the issues of the documents not listed in the DoDISS shall be the issue of the nongovernment documents which is current on the date of the solicitation.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

E 92 Standard Test Method for Vickers Hardness of Metallic Materials.

(Application for copies of ASTM publications should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103).

(Nongovernment standards and other publications are normally available from the organizations which prepare or which distribute the documents. These documents also may be available in or through libraries or other informational services.)

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence. Nothing in this specification, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

3. *REQUIREMENTS*

3.1 Material. Pins shall be fabricated from AISI 302 stainless steel. Pins shall be flash plated with elemental 24 karat gold.

3.2 Hardness. The Rockwell hardness of the plated (steel portion) of the pin (threaded portion) shall be Rc 40 to 45 when tested as specified in 4.4.2. Temper and hardness of the finished pin shall allow pin installation without deformation, twisting or fracturing during use and shall permit shearing at shear grooves of self-shearing pins (upon continued application of torque after the pin reaches the drilled hole depth).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

3.3 Style, design and dimensions. Pins shall be right-handed self-threading pins suitable for pin retention of dental restorative materials. Pins shall be self-threading and capable of locking within smooth undersized drilled holes in tooth dentine without twisting, deforming, or fracturing. Pins shall be of the type and size as specified in table 1. Pins shall be suitable for use with the applicable drills, drivers and mandrels as specified in table 1. All pins shall have bilateral flattened head extension on one end capable of engagement with the applicable driver and mandrel specified in table 1.

Type I pins shall be self-threading pirs. Type II pins shall be self-threading and self-shearing pins. Type III pins shall be self-threading, self-shearing and two-in-one configuration.

Type I pins shall be capable of advancing (threading) to the full pre-determined hole depth and locking within the drilled hole (having the diameter and depth as specified in table 2).

Type II pins shall be capable of advancing (threading) to the full predetermined hole depth and locking within the drilled hole (having the diameter and depth as specified in table 2) whereupon continued application of torque shall cause self-shearing at the shear groove.

Type III pins shall consist of two joined pins separated by a shear groove. Type III pins shall be capable of advancing (threading) to the full pre-determined hole depth, shearing at the shear groove, and then permit the remaining pin to be installed as a separate pin.

The plated pins shall be capable of amalgamation bonding and shall insure biocompatibility. The threads shall be pitched so as to advance all pins 2 millimeters in no more than 10 revolutions except for type II size 0.0145 inch which shall advance 1.3 millimeters in no more than 11 revolutions and type II size 0.019 inch which shall advance 1.5 millimeters in no more than 10 revolutions. Style, design, and dimensions of the pins shall conform to DPSC Drawing 23205 for the applicable type and size pin.

Source: https://assist.dla.mil -- Downloaded: 202610-21T20:10Z
Check the source to verify that this is the currentversion before use.

TABLE 1.

TYPE	SIZE (INCH)	FOR USE WITH 1/			
		DRILL	DRIVER	MANDREL	BENDER
I	0.024	6520-00-017-4930 6520-00-149-0571	6520-00-935-9865 6520-01-099-2319	6520-00-935-9873	6520-00-935-9864
I	0.031	6520-00-149-0565 6520-00-420-0510 6520-00-935-1098	6520-00-935-9866 6520-01-099-2319	6520-00-935-9872	6520-00-935-9863
II	0.0145	6520-01-095-3707	6520-01-094-8297	6520-01-094-6141 6520-01-095-1333	6520-00-935-9864
II	0.019	6520-01-036-2699 6520-01-119-6397	6520-00-935-9866 6520-01-099-2319	6520-00-935-9877	6520-00-935-9864
II	0.031	6520-00-420-0510 6520-00-935-1098 6520-00-149-0565	6520-00-935-9866 6520-01-099-2319	6520-00-935-9877	6520-00-935-9863
III	0.024	6520-00-017-4930 6520-00-149-0571 6520-00-420-0511	6520-00-935-9865 6520-01-099-2319	6520-00-935-9873	6520-00-935-9864
III	0.031	6520-00-149-0565 6520-00-420-0510 6520-00-935-1098	6520-00-935-9866 6520-01-099-2319	6520-00-935-9872	6520-00-935-9863

TABLE 2.

TYPE	SIZE (INCH)	HOLE	
		DIAMETER (INCH)	DEPTH (INCH)
I	0.024	0.021 (0.525mm)	0.079 (2mm)
I	0.031	0.027 (0.675mm)	0.079 (2mm)
II	0.0145	0.0135 (0.350mm)	0.051 (1.3mm)
II	0.019	0.017 (0.425mm)	0.059 (1.5mm)
II	0.031	0.27 (0.675mm)	0.079 (2mm)
III	0.024	0.021 (0.525mm)	0.079 (2mm)
III	0.031	0.027 (0.675mm)	0.079 (2mm)

Source: https://assist.dla.mil -- Downloaded: 20260-21120:102
Check the source to verify that this is the current version before use.

3.4 Corrosion-resistance. The unplated pins shall meet the requirements of GG-I-526 for resistance to boil test, when tested as specified in 4.4.3.

3.5 Bending ability. The plated pin (threaded portion) shall be capable of being bent (cold bending) a minimum of 45 degrees without signs of fracture when tested as specified in 4.4.4.

3.6 Shearing ability. Plated self-shearing pins shall shear at the appropriate self-shearing groove without the pin twisting, deforming or fracturing, when tested as specified in 4.4.5 and 4.4.6 (see DPSC drawing for location of shear groove).

3.7 Plating characteristics. The gold plating shall be such that the plating shall not darken nor separate from the basis metal (flake, peel or blister) when tested in accordance with the plating test specified in 4.4.7.

3.8 Finish. Threads shall be well defined. The plated pin shall have a bright gold plated finish. Plating shall be smooth, continuous, adherent, free from defects such as lumps, blisters, peeling and uncoated areas.

3.9 Workmanship. Threads shall be continuous and uniform. Pins shall be free from irregularities, burrs, slivers, cracks, or other defects which impair their serviceability, proper functioning or intended use.

3.10 Bid samples. Bid samples which are representative of the pins to be supplied shall be furnished complete with protocols and satisfactory report of tests performed in accordance with the specification to the Defense Personnel Support Center for evaluation by the Military Medical Services. When tested by dentists under conditions of normal usage, the pins shall be suitable for pin retention of dental restoration materials in the self-threading pin retention technique. The plated pin shall be capable of amalgamation bonding and shall insure biocompatibility. The pins shall be suitable for use with the applicable armamentaria (handpiece, drills, drivers, mandrels and benders), shall be self-threading, and shall lock within smooth appropriate undersized drilled holes in tooth structure without twisting, deforming, or fracturing. Self-shearing pins shall shear at the shear groove after being threaded to a specified pre-determined depth with continued application of torque. The two-in-one pins shall thread to a specified pre-determined depth, lock within the hole, and shear at the shear groove, thereby allowing the remaining pin to be installed as a separate pin. Once in place, the pins shall be capable of being bent without fracture.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in tables III, IV, and V (see 4.2.1, 4.2.2, and 4.3).

4.1.4 Inspection lot. An inspection lot, for purposes of sampling and testing in accordance with MIL-STD-105, shall consist of pins of the same type and size, manufactured under essentially the same conditions.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQLs) shall be as indicated in table III. The unit of product for examination purposes shall be one pin.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

TABLE III. Sampling criteria for examination.

Type defect	Inspection level	AQL (Percent defective)
Major A	II	1.0
Major B	I	1.0
Major C	S-4	*
Minor	II	4.0

*Acceptance number shall be zero (0) for all sample sizes.

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified in table IV. The acceptance number shall be zero for all sample sizes. The unit of product for test purposes shall be one pin.

TABLE IV. Sampling criteria for tests.

Characteristics	Require-ments	Test Procedures	Inspection Level
Chemical composition	3.1	4.4.1	1/
Hardness test	3.2	4.4.2	S-2
Corrosion-resistance 2/	3.4	4.4.3	I
Bending ability	3.5	4.4.4	I
Shearing test (hand operation)	3.6	4.4.5	I
Shearing test (hand-piece operation)	3.6	4.4.6	I
Plating test	3.7	4.4.7	I

1/A sufficient quantity of each lot of metal used in the manufacture of the pins shall be selected for composition analysis provided lot identity has been maintained and can be demonstrated to the Government.

2/Testing shall be conducted on the finished pin prior to plating.

4.3 Examination. The pins shall be examined to determine compliance with specification requirements. Examination shall be conducted in accordance with table III. The pins shall be examined for defects, including but not limited to those listed in table V. Examination shall be conducted under four power (4X) magnification for Major A and Minor classified defects.

Source: https://assist.dla.mil -- Downloaded: 202610-21T20:10Z
Check the source to verify that this is the currentversion before use.

TABLE III. Classification of defects.

Categories	Defects
Major A:	
101	Style and design not as specified for the applicable type and size specified.
102	Finish not as specified.
103	Pins not free from burrs, slivers or cracks.
104	Threads not uniform and symmetrical.
105	Pins not plated.
106	Plating not free from lumps, blisters, peeling and uncoated areas.
107	Threads stripped or damaged affecting serviceability.
108	Threads missing.
109	Plating discolored.
Major B:	
151	Dimensions not as specified (other than threads per inch).
Major C:	
161	Number of threads per inch (TPI) not as specified.
Minor:	
201	Flattened portion of pin grossly irregular.
202	Threads damaged, not affecting serviceability.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Chemical composition. In the absence of certificates of quality for the raw material, specimens of materials used to fabricate the pins, or taken from the pins, shall be subjected to analysis in accordance with methods 111 or 112 of Fed. Test Method Std. No. 151. In case of dispute, Method 111 shall be used for referee analysis.

4.4.2 Hardness test. Hardness shall be determined in accordance with ASTM Designation E92 (Vickers or Diamond Pyramid).

4.4.3 Boil test. The boil test as specified for instruments in GG-I-526 shall be conducted on the pins.

4.4.4 Bend test. The threaded portion (free end) of the pin (2mm length for all type pins except 1.5mm length for size 0.0145 inch and size 0.019 inch pin) shall be held in a chuck or other holding device. The pin shall then be bent a minimum of 45 degrees using the bender specified in table I for the type and size pin being tested. The pin shall be examined at seven power (7X) magnification for signs of fracture.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

4.4.5 Shearing test (hand operation). The threaded portion (free end) of the self-shearing pins (2 mm length for all Type II and III sizes except 1.5 mm length for size 0.0145 inch) shall be securely held in a chuck or other holding device. The head of the pin shall then be slowly rotated using the driver specified in table 1 for the type and size pin being tested until the pin shears. The pin shall be examined at seven power (7X) magnification to determine if the pin sheared at the shear groove and for pin twisting, deforming, or fracturing.

4.4.6 Shearing test (handpiece operation). The test as specified in 4.4.5 shall be conducted using a low speed handpiece (100-500 rpm) 6520-00-019-9267 and the mandrel specified in table 1 for the type and size pin being tested.

4.4.7 Plating test. The plated pin shall be placed in a pre-heated oven at 450° ± 5°C degrees for 3 minutes. After removal and cooling, the surface of the pins shall be examined at four power (4X) magnification for darkening, flaking, peeling and blistering.

4.5 Inspection of packaging. The sampling and inspection of the preservation and interior package marking shall be in accordance with groups A and B quality conformance inspection requirements of MIL-P-116. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C, as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Unit package. One hundred Type I or Type III pins of one size only or one hundred Type II, size 0.031 pins or fifty Type II, size 0.0145 or 0.019 pins shall be packaged in a clear plastic vial having a friction plug type closure. As an alternate, the pins shall be packaged in a plastic box having a hinged cover and friction type closure. The vial or box shall be further packaged in accordance with MIL-P-116, Method IC-1.

5.1.1.2 Intermediate package. Twelve unit packages containing pins of one type and size only shall be packaged in a box conforming to PPP-B-566 or PPP-B-676. Closure shall be as specified in the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 202640-21T20:10Z
Check the source to verify that this is the currentversion before use.

5.1.2 Level C. The level C preservation for dental pins shall conform to the MIL-STD-794 requirements for this level.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.1).

5.2.1 Level A. Thirty-six intermediate packages (432 units) containing pins of one type and size only shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type, PPP-B-621, class 2 or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. Thirty-six intermediate packages (432 units) containing pins of one type and size only shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type, PPP-B-621, class 1 or PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

5.2.3 Level C. The level C packing for dental pins shall conform to the MIL-STD-794 requirements for this level.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination equal 40 cubic feet or more. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 40 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 43 inches in height, 48 inches in width, and 40 inches in length. Pallet shall conform to Type IV or Type V.

5.2.5.1 Levels A and B. Dental pins packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets as specified in 5.2.5.

5.2.5.2 Level C. Dental pins packed as specified in 5.2.3 shall be unitized as specified in MIL-STD-794.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Levels A, B and C. Each vial or plastic box shall be marked in accordance with commercial practice. Each unit pack, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.5.

6. NOTES

6.1 Ordering data.

6.1.1 Acquisition requirements. Acquisition documents should specify the following:

(a) Title, number, and date of this specification.
(b) Type and size required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 National Stock Number (NSN) coverage. This specification covers the following items appearing in the Federal Supply Catalog:

NSN	Item Identification	Specification Designation	
		Type	Size
6520-00-420-0508	PIN, THREAD-FORMING, DENTAL, Self-Shearing, 2 in 1, 100s.	III	0.031
6520-00-420-0509	PIN, THREAD-FORMING, DENTAL, Self-Shearing, 2 in 1, 100s.	III	0.024
6520-00-935-9867	PIN, THREAD-FORMING, DENTAL, Self-Shearing, 100s.	II	0.031
6520-00-935-9868	PIN, THREAD-FORMING, DENTAL, 100s.	I	0.024
6520-00-935-9869	PIN, THREAD-FORMING, DENTAL, 100s.	I	0.031
6520-01-042-9335	PIN, THREAD-FORMING, DENTAL, 50s.	II	0.019
6520-01-095-3708	PIN, THREAD-FORMING, DENTAL, 50s.	II	0.0145

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

6.3 Subject term (key word) listing.

Dental Pin
Self-Shearing, Self-threading Pin

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air-Force - 03

Civil Agency Coordinating Activity
VA-OSS

05400

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Project No. 6520-1964

Location: NCR Disk 6

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEPARTMENT OF THE ARMY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 12062 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEPARTMENT OF THE ARMY

Commander
Defense Personnel Support Center
Directorate of Medical Materiel
ATTN: DPSC-RST (9-3-F)
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
MIL-P-42011 (DM)	PIN, SELF-THREADING, SELF-SHEARING, DENTAL

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*: _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:10Z
Check the source to verify that this is the current version before use.