MIL-S-42002 (DM)
10 December 1985

MILITARY SPECIFICATION

SUCTION APPARATUS, DENTAL

This specification is approved for use within DLA, Defense Personnel Support Center (DM), and is available for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a heavy duty dental suction apparatus (aspirator) suitable for surgical purposes.

1.2 Classification.

1.2.1 Suction apparatus, complete. The suction apparatus shall be furnished complete with all component parts, as specified (see 3.3.1).

1.2.2 Replacement parts. Replacement parts for the suction apparatus shall be the following parts, as specified (see 3.3.2).

Part a. Tip handle
Part b. Tip No. 2, Coupland type
Part c. Tip No. 4, Coupland type
Part d. Tube and bottle cap assembly
Part e. Bacteria filter

AMSC N/A FSC 6520

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or or by letter.

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:08Z
Check the source to verify that this is the current version before use.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications and standards. The following specifications and standards form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DoDISS) and supplement thereto, cited in the solicitation.

SPECIFICATIONS

FEDERAL

CC-M-1807 - Motors, Alternating Current, Fractional and Integral Horsepower (500 HP and smaller).
GG-I-526 - Instruments, Dental and Surgical; General Specifications.
PPP-B-566 - Boxes, Folding, Paperboard.
PPP-B-636 - Boxes, Shipping, Fiberboard.
PPP-B-676 - Boxes, Setup.

MILITARY

MIL-M-7298 - Manual, Technical Commercial Equipment.
MIL-P-116 - Methods of Preservation.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.
MIL-STD-147 - Palletized Unit Loads.
MIL-STD-794 - Parts and Equipment, Procedures for Packaging of.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

1143 - Handle and Tips, Dental Suction Apparatus,

(Copies of specifications, standards, and drawings required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity).

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise

Source: https://faesist.dla.mil -- Downloaded: 2026-10-21T20:08Z
Check the source to verify that this is the current version before use.

specified, the issues of the documents which are DoD adopted shall be those listed in the issue of the DoDISS specified in the solicitation. Unless otherwise specified, the issue of documents not listed in the DoDISS shall be the issue of the nongovernment document which is current on the date of the solicitation.

AMERICAN NATIONAL STANDARDS INSTITUTE (ANSI)

ANSI C73.11 - Electric Plugs and Receptacles - General Purpose, 125 Volts, 15 Amperes, 2 Pole, 3 Wire, Grounding Type.

(Application for copies should be addressed to the American National Standards Institute, 1430 Broadway, New York, NY 10018.)

UNDERWRITER'S LABORATORIES, INC. (UL)

UL 544 Standard for Electric Medical and Dental Equipment.

(Application for copies should be addressed to the Underwriter's Laboratories, Inc., 207 East Ohio Street, Chicago, IL 60611; 333 Pfingsten Road, Northbrook, IL 60062; 1285 Walt Whitman Road, Melville, LI, NY 11746; 1655 Scott Boulevard, Santa Clara, CA 95050).

(Nongovernment standards and other publications are normally available from the organizations which prepare or which distribute the documents. These documents also may be available in or through libraries or other informational services.)

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence. Nothing in this specification, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 Fire and casualty hazards.

3.1.1 Proof of compliance. The contractor shall provide proof, to the procurement contracting officer, not less than thirty days prior to delivery, that the items to be delivered under the contract conform to the applicable requirements of Underwriter's Laboratories, Inc. (UL) Standard of Safety 544. A copy of the test report shall be made available to the Government inspection personnel for their use in determining that the units being submitted for acceptance are identical to the unit submitted for approval. Proof shall consist of one of the following:

a. UL listing and UL label affixed to the product, provided the

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:08Z
Check the source to verify that this is the current version before use.

contractor agrees to have Underwriters' Laboratories, Inc. forward a copy of the UL report directly to the contracting officer, and the Government inspection personnel, upon request.

b. A label or equivalent means of identification affixed to the product as evidence that the product design has been examined and tested for design safety compliance by a nationally recognized independent laboratory. If an independent laboratory is used, it must be adequately equipped and competent to determine compliance with UL standards. The Government reserves the right to require a facilities report and to conduct a facilities survey. In support of the label/product identification, the independent testing laboratory shall provide surveillance over the product manufacturing system to the degree necessary to insure that production units are manufactured in the same manner and with the same type of components as the test sample evaluated and upon which the validity of the product identification is based. The testing laboratory shall develop a detailed test report which includes a conclusion as to whether the test sample evaluated is "approved" or "not approved" as complying with all requirements of the standards of safety.

3.2 Materials. Materials not specifically designated shall be of high commercial quality throughout, and shall be entirely suitable for the purpose intended.

3.3 Style and design. The dental suction apparatus shall be mobile pedestal type suitable for operation on 115 volts, 60 hertz, AC.

3.3.1 Suction apparatus, complete. Shall consist of a motor driven pump, a complete suction or aspirating system, properly connected and provided with a 32-ounce bottle, cap, adjusting regulator, 6-foot and 15-inch piece of disposable tubing, compactly mounted on a sturdily built pedestal type stand with a full base and a metal box type top. The unit shall be furnished with the following accessories: Tip handle, Tip No. 2, Tip No. 4, one additional 32-ounce bottle, without cap and a package of 3-bacteria filters.

3.3.2 Replacement parts. The style, design, and dimensions of part a, Tip handle; part b, Tip No. 2, Coupland type; and part c, Tip No. 4, Coupland type; shall conform to DPSC drawing 1143. Part a, b and c shall be chrome or nickel plated brass and shall meet the boil test requirements of GG-I-526 for plated brass. The tips shall be secured and held to the threaded (distal) end of the handle by means of a knurled, internally threaded, tightening sleeve. Part d, Tube and bottle cap assembly, shall consist of one rubber cap and two lengths of tubing. Part e shall consist of three bacteria filters. Each part shall conform to requirements specified herein.

3.3.3 Stand and components. The suction apparatus shall weigh 53 pounds, plus or minus 5 pounds. The top shall be painted 16 gauge

Source: https://assist.dla.mil -- Downloaded: 202610-21T20:08Z
Check the source to verify that this is the current version before use.

CRS, 14 inches wide, 12 inches deep, 8½ inches high with ventilation louvers. A front control panel fabricated of painted 16 gauge CRS shall be riveted to the top. The top shall be mounted to an upright and base, constructed of the same material, and shall be 32½ inches from the floor to the top of the stand. The stand top cover shall be 18-gauge stainless steel. The casters shall be double-wheeled, rubber tread, fender type, oilless with top ball bearings. There shall be two wheels per caster, each 2-3/8 inch diameter; ½ inch wide (2-inch total width). The 32-ounce bottle shall be held by a wire formed bottle bracket secured to the right side of the stand top. The bottle bracket shall hold the bottle in a protective manner.

3.3.4 Motor and pump assembly. Motor and pump assembly shall be 7 3/4 inches long, 7½ inches high and 5 inches wide. The pump shall be mounted on the end of the motor, the shaft being common to both.

3.3.4.1 Motor The pump drive motor shall be shaded pole, continuous duty, vented, fan cooled, and shall contain a built-in overload circuit breaker. The motor shall be rated at 1/8 horsepower, 1575 revolutions per minute, 115 volts, and 60 cycles. The starting current for starting shall not exceed 16 amperes and shall operate at 4.0 amperes maximum, with a maximum temperature rise of 40 degrees Centigrade, and shall be in accordance with CC-M-1807.

3.3.4.2 Cord. An electric cord of proper length with "hospital grade" plug shall be provided to permit connection of aspirator to receptacle in the dental operating unit. The cord shall be secured to the stand with a power cable strain relief. The suction apparatus shall withstand the dielectric strength test specified in 4.4.4. The "on-off" power switch shall be mounted in the front panel and shall light when the unit is energized.

3.3.4.3 Pump. The pump shall be direct drive, single head, diaphragm compressor, capable of developing negative pressure of no less than 22 inches of mercury. The pump shall have sufficient capacity to evacuate the air from a 32-ounce bottle to a vacuum of 22 inches of mercury in 6 seconds. Three filter elements, within the pump head, shall filter air passing thru the inlet and outlet ports. Inlet and outlet valves shall be stainless steel flapper type secured in position with retainer plates and machine screws. An exterior muffler on the outlet port shall be provided to assure quiet operation. The pump shall be capable of 8 hours of continuous operation with a minimum of 25 liters per minute free air flow without requiring servicing.

3.3.4.4 Lubrication. The motor and pump bearings shall be permanently lubricated at the factory and shall need no further lubrication.

3.3.5 Positive and negative inlet. The muffler in the pump outlet port shall be designed to permit easy removal and replacement of the filter material when necessary. The negative or suction system

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:08Z
Check the source to verify that this is the current version before use.

shall include a 30-inch capacity vacuum gauge with a 2½-inch dial (which shall be read both inches and millimeters of mercury). A vacuum gauge shall be accurate to plus or minus 1-inch of mercury. The vacuum regulator fabricated of plastic shall provide adjustable levels. The pump shall be protected from damage from overfilled suction bottle and contamination by a hydrophobic, microporous bacteria filter, part e, connected between, and in series with the pump and the vacuum bottle. The bacteria filter shall be contained in a polypropylene housing and screwed into a bulkhead fitting provided on the stand and shall have a hose barb connector to accept ¼-inch to ½-inch flexible tubing. The filter shall be capable of being autoclaved three times.

3.3.6 Bottle assembly. A 32-ounce capacity glass bottle shall be supplied in the vacuum system. The bottle shall be 8-3/4 inches in height and 3-3/4 inches in diameter, with a neck 2 inches in diameter. The bottle shall be properly labeled, including the maximum filling line. The bottle shall withstand a vacuum of 25 inches of mercury (see 4.4.1).

3.3.6.1 Cap. A molded, overlapping, oilproof, chemical-resistant, synthetic rubber cap shall be furnished with the bottle. The cap shall have two holes for the reception of the inlet and outlet tubes, and shall be such that it will form a seal around the lip as well as the top of the bottle.

3.3.6.2 Fittings. Inlet and outlet tubes shall be provided for attachment of patient tubing and pump tubing. The tubes shall be fabricated of brass, chromium plated, and/or stainless steel, be a slip type connector and provide a proper seal when installed to cap specified in 3.3.6.1. The inlet tube shall extend no less than 2-5/8 inches farther into the bottle than the outlet tube, in order to prevent splashing. The tubing shall be clear, plastic tubing with plastic connectors to adapt to metal parts. The tubing shall be clean and packaged as a unit consisting of one 6-foot piece and one 15-inch piece, ready for use. The tubing shall be disposable, single patient type.

3.4 Tolerances. Unless otherwise specified, tolerances on dimensions shall be ± 5 percent, provided fit, function, and first quality workmanship are not sacrificed thereby.

3.5 Caution and warning notices.

3.5.1 Power supply card. Tag attached to power supply cord shall be permanently and legibly marked as follows:

CAUTION

GROUNDING RELIABILITY CAN ONLY
BE ACHIEVED WHEN THE EQUIPMENT
IS CONNECTED TO AN EQUIVALENT
RECEPTACLE MARKED "HOSPITAL GRADE"

Source: https://assist.dla.mil -- Downloaded: 202610-21T20:08Z
Check the source to verify that this is the current version before use.

3.5.2 Servicing notice. A suitable notice shall be provided on the aspirator where readily visible during any attempt to perform servicing:

CAUTION

ELECTRIC SHOCK HAZARD
DO NOT REMOVE COVER (OR BACK).
REFER SERVICING TO QUALIFIED
SERVICE PERSONNEL.

3.5.3 Flammable anesthetics. Aspirators shall be provided with a readily visible warning silkscreened on the stand front as follows:

DANGER

POSSIBLE EXPLOSION HAZARD
IF USED IN PRESENCE OF
FLAMMABLE ANESTHETICS

3.6 Identification marking. Each suction apparatus unit shall be permanently and legibly marked with the name or registered trademark of the manufacturer, model designation, serial number, and electrical characteristics, such as voltage, cycle, and amperage.

3.7 Service data. The contractor shall furnish, with each suction apparatus unit, two copies each of technical manual which contains complete instructions for installation, operation and maintenance, and lists of component parts. The manual shall comply with the requirements of MIL-M-7298. As soon as praticable after award of the contract or purchase order, the contractor shall furnish to the Contracting Officer, via the cognizant Government inspector, two preliminary copies of his proposed technical manual for review and comment by the Directorate of Medical Material, Defense Personnel Support Center. The preliminary manuals shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract or purchase order. In addition, the Contractor shall furnish 15 copies of the approved technical manual to the Contracting Officer for distribution. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of the contract and propose to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon request to the Contracting officer. Waiver of submission of drafts and submission of 15 additional copies shall be at the option of the Government.

3.8 Workmanship. Items shall be free from defects which detract from their appearance or impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:08Z
Check the source to verify that this is the current version before use.

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material,the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQLs) shall be as indicated in table I.

TABLE I. Sampling for examination.

Type defect	Inspection level	AQL (Percent defective)
Major	II	1.0
Minor A	II	2.5
Minor B	S-2	4.0

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:08Z
Check the source to verify that this is the current version before use.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in table II. The unit of product shall be one dental suction apparatus. Acceptance number shall be zero for all sample sizes.

TABLE II. Sample for tests.

Component	Characteristic	Requirement	Test Procedure	Inspection level
Handle and Tips	Boil Test	3.3.2	GG-I-526	S-2
Suction bottle	Vacuum	3.3.6	4.4.1	S-2
Pressure gage	Accuracy	3.3.5	4.4.2	S-2
End item	Performance	3.3.4.3	4.4.3	100 percent
End item	Dielectric strength	3.3.4.2	4.4.4	100 percent

4.2.3 Sample unit usage. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The dental suction apparatus shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted as specified in 4.2.1. Any unit in the sample containing one or more defects shall be rejected, and, if the number of defective units in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected. Inspection shall not be restricted to the defects listed in the table III.

TABLE III. Classification of defects in accordance with MIL-STD-105.

Categories	Defects
Major:	
101	Any component part missing.
102	Warning plate missing or not as specified.
103	Exposed edges and corners not free of burrs and sharpness.

Source: https://assist.dla.mil -- Downloaded: 202610-21T20:08Z
Check the source to verify that this is the currentversion before use.

104	Any weld omitted, not fused; not free of burns, cracks, and fractures.
105	Any threaded part stripped, or does not function smoothly.
106	Any bottle not free of cracks, breaks, chips, and sharp and deformed areas.
Minor A:	
201	Unit not clean; not free of dirt, oil, and grease.
202	Blistering, checking, or peeling of the finish.
203	Screws not drawn up tight; heads not flush with surface.
204	Service data not included.
205	Identification marking missing, incomplete, illegible,or not properly attached.
206	Fastening members not properly installed.
Minor B:	
251	Dimensions not within tolerances specified.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Vacuum (suction) bottle. The suction bottle shall be subjected to an internal vacuum of 25 inches of mercury for a period of 1 minute without failure (see 3.3.6).

4.4.2 Vacuum gage. Gages shall be examined to determine that the calibrations conform to requirements. The gages shall be measured for accuracy using mercury manometer, and shall be accurate to plus or minus 1 inch of mercury (see 3.3.5).

4.4.3 Performance. The unit shall be operated at maximum vacuum (22 inches of mercury) in accordance with manufacturer's instructions for a minimum of eight hours. The vacuum shall be easily variable from 0 to 22 inches of mercury by means of a needle and seat type of vacuum control valve. The minimum time required for suction apparatus to reach 22 inches of mercury shall be noted (see 3.3.4.3).

4.4.4 Dielectric strength. A dielectric test in accordance with UL requirements shall be performed after the pump has been warmed by operation (see 3.3.4.2).

4.5 Inspection of packaging. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:08Z
Check the source to verify that this is the current version before use.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C, as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Unit package.

5.1.1.1.1 Suction apparatus complete. The bottle, tip handle, tips #2 and #4, instructions, the 3 replacement filters and the additional bottle shall be wrapped in tissue paper or other suitable material and cushioned in one or more fiberboard boxes of a suitable size and design. Closure shall be effected by use of tape, gluing or staples. The suction apparatus along with the boxed components shall be packed in a fiberboard box conforming to PPP-B-636, class domestic and having a minimum bursting strength of not less than 350 pounds. The base of the unit shall be blocked in such a manner that the weight of the unit is distributed over the entire bottom of the box and not on the casters. The unit shall also be cushioned within the box to prevent shifting and damage to the unit or the container. Closure of the box shall be as specified in the box specification.

5.1.1.1.2 Replacement parts.

5.1.1.1.2.1 Tip handle, tip No. 2 and tip No. 4. One tip handle, tip No. 2 or tip No. 4 shall be wrapped in tissue paper or other suitable material and packaged in accordance with MIL-P-116, method IC-1.

5.1.1.1.2.2 Tube and bottle cap assembly. Each unit shall have the tubing neatly coiled and the complete unit wrapped in kraft paper or other suitable material and cushioned within a box conforming to PPP-B-636, class domestic. Closure shall be as specified in the box specification.

5.1.1.2 Intermediate package. Twelve unit packages containing tip handles, tips No. 2 or tips No. 4 shall be packaged in a box conforming to PPP-B-566, PPP-B-636, class domestic or PPP-B-676. Closure shall be as specified in the applicable box specification. An intermediate package shall not be required for the tube and bottle cap assembly.

5.1.2 Level C. The level C preservation for the complete suction apparatus or replacement parts shall conform to the MIL-STD-794 requirements for this level.

5.2 Packing Packing shall be level A, B,or C, as specified (see 6.1).

5.2.1 Level A.

Source: https://assist.dla.mil -- Downloaded: 202610-21T20:08Z
Check the source to verify that this is the currentversion before use.

5.2.1.1 Suction apparatus complete. Each unit packaged as specified in 5.1.1.1.1 shall require no further packing except that the exterior container shall conform to PPP-B-636, class weather-resistant. The box shall be closed, strapped, and waterproofed as specified in the box specification.

5.2.1.2 Replacement parts. Twelve intermediate packages containing tip handles, tips No. 2, tips No.4 or six unit packages containing tube and bottle cap assemblies shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, class weather-resistant. Closure, strapping and waterproofing shall be as specified in the box specification.

5.2.2 Level B.

5.2.2.1 Suction apparatus complete. Each unit packaged as specified in 5.1.1.1.1 shall require no further packing.

5.2.2.2 Replacement parts. Twelve intermediate packages containing tip handles, tips No. 2, tips No. 4 or six unit packages containing tube and bottle cap assemblies shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the box specification.

5.2.3 Level C. The level C packing for the complete suction apparatus or replacement parts shall conform to the MIL-STD-794 requirements for this level.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination equal 40 cubic feet or more. Loads shall be unitized on pallets. Quantities less than 40 cubic feet need not be unitized. Unitized loads shall be uniform in size and quantities to the greatest extent practicable.

5.2.5.1 Levels A and B. Complete suction apparatus and replacement parts packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets conforming to MIL-STD-147.

5.2.5.2 Level C. Complete suction apparatus and replacement parts packed as specified in 5.2.3 shall be unitized as specified in MIL-STD-794.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:08Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Levels A, B and C. Each unit pack, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container container (see 5.2.1, 5.2.2 and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.5.

6. NOTES

6.1 Ordering data.

6.1.1 Acquisition requirements. Acquisition documents should specify the following:

a. Title, number and date of this specification.
b. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 National Stock Number (NSN) coverage. This specification covers the following item listed in the Federal Supply Catalog.

National Stock Number	Item Identification
6520-01-206-0299	SUCTION APPARATUS, DENTAL Mobile, 115 volt, 60 Hz, AC.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force - 03

Review Activities:

Army-MD
Navy-MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 650-1905

Location: NCR

Source: https://assist.dla.mil - Downloaded: 2026-10-21T20:08Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:08Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Defense Personnel Support Center
Directorate of Medical Materiel
DPSC-ATT
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:08Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
MIL-S-42002	Suction Apparatus, Dental

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 202640-21T20:08Z
Check this source to verify that this is the current version before use.