MIL-G-23337 (WP)
AMENDMENT-3
15 MAY 1965
SUPERSEDING
AMENDMENT-2
1 NOVEMBER 1963

MILITARY SPECIFICATION

GAGES, PRESSURE, DIAL INDICATING

This amendment forms a part of Military Specification MIL-G-23337(WEP) dated 15 July 1962 and has been approved by the Bureau of Naval Weapons, Department of the Navy.

Page 2, paragraph 2.1, under SPECIFICATIONS Federal: Delete

"DD-G-451 Glass, Flat and Corrugated, for Glazing, Mirrors and Other Uses"

Page 2, paragraph 2.1, under SPECIFICATIONS Military: Add

"MIL-P-5425 Plastic, Sheet, Acrylic, Heat Resistant"

Page 2, paragraph 2.1, under STANDARDS Federal: Delete

"FED-STD-1 Standard for Laboratory Atmospheric Conditions for Testing"

*Page 4, paragraph 3.4.3: Add to end of first sentence -

"with 0.750 hex nominal across-flats".

Page 5, paragraph 3.5.3: Delete and Substitute

"3.5.3 Cover Lens - The lens shall be in accordance with Specification MIL-P-5425."

Page 5, paragraph 3.5.4: Delete and Substitute

"3.5.4 Cover Lens Mounting - The distance between the front surface of the case and the outside surface of the cover lens shall not exceed .01 inch. The distance between the dial plate and the inside surface of the cover lens shall be the minimum practicable and shall not exceed 0.156 inch. The cover lens shall be properly sealed."

Page 5, paragraph 3.5.5.1, last line: Delete "shall not be less than 1/64th inch"

and substitute "shall be adequate for legibility."

FSC 6685

Source: https://assist.dla.mil -- Downloaded: 2026-10-19T16:27Z
Check the source to verify that this is the current version before use.

*Page 5, paragraph 3.5.6: Delete and Substitute

"3.5.6 Pointer - The pointer shall be as light as practicable and firmly attached to the mechanism. The length of the pointer shall be such that the tip will extend into the scale a distance of 1/3 to 2/3 of the length of the shortest graduation. The pointer shall not mar or chafe the dial markings during operation."

Page 9, paragraph 4.5.1: Delete and Substitute

"4.5.1 Standard Conditions - Unless otherwise specified, all tests required by this specification shall be made under the following conditions:

Temperature	Room ambient 25° C + 5°C
Pressure	Normal atmospheric (approximately 29.92 inches Hg.)
Humidity	Room ambient up to 90% relative humidity"

Page 10, paragraph 4.5.3: Delete

Page 11, paragraph 4.6.6: Delete and substitute

"4.6.6 Vibration Test - The gage shall be tested in accordance with procedure XII of Specification MIL-E-5272 except that the G level shall be 20 G up to a frequency of 2000 cps. Half of each resonant and cycling period shall be conducted at room temperature and the other half at 95° C (199°F). At the end of the test the instrument shall be subjected to and meet the requirements specified in the following tests:

(1) Scale error at room temperature (See Table II)
(2) Friction test
(3) Leakage test."

*An asterisk preceding a page number denotes a change to be made to the specification which did not appear in the previous amendment.

Source: https://assist.dla.mil -- Downloaded: 2026-10-19T16:27Z
Check the source to verify that this is the current version before use.