INCH - POUND

A-A-30056E
October 13, 1992
SUPERSEDING
A-A-30056D
July 28, 1989

COMMERCIAL ITEM DESCRIPTION

CUP, SPECIMEN (Sterile and Nonsterile)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description describes a specimen container with metal and plastic lids, used for the collection of urine specimens.

Salient characteristics. Size - A - The primary container shall be round, disposable, translucent, plastic polypropylene. The lid shall provide a leakproof seal when tested as follows: The cup shall be filled with 4 oz. (118 ml.) of water. With the lid placed firmly on the container, the container shall be held upside down and then placed on its side on a flat surface for 5 minutes. Any evidence of leakage will constitute failure. The container shall be stable on a flat surface, and shall be free from objectionable odor, sharp edges, deformations, cracks, burrs, holes, or thin areas. There shall be an equal number of cups and caps and (labels if applicable) per case. These specimen cups shall be of the following sizes, types and styles, as specified.

Style I - With plastic lid.

Type I - Nonsterile, 4 to 4-1/4 oz. (118 to 133 ml) capacity, with graduations of oz. and ml/cc in 1/4 oz. (7ml.) and 10 ml/cc increments molded into the container. White lid (not translucent), the lid shall have provisions for patient identification, either a label or write-on surface, that accepts, at a minimum, the patient's name and room number. The containers shall be designed to nest, not to exceed 25 per tube or sleeve. Packed to remain clean until time of use.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Type II - Sterile; 4 to 4-1/2 oz. (118 to 133 ml.) capacity, with graduations of oz. and ml/cc in 1/4 oz. (7 ml.) and 10 ml/cc increments molded into the container. Peel open pouch (O.R. pack). White lid not acceptable. The lid shall have provisions for patient identification, either a label or written surface, that accepts, at a minimum, the patient's name, date and room number. The OR pouch shall permit aseptic handling of and delivery of the cup, uncontaminated, onto a sterile field. A tear-open type package shall not be acceptable. The label of the immediate container shall bear the word "sterile".

Style II - With metal lid.

Type I - Nonsterile - 4-1/2 to 5 oz. (133 to 148 ml.) capacity, with graduations of oz. and ml/cc in 1/4 oz. (7 ml.) and 10 ml/cc increments molded into the container. The lid shall have full circumference screw rings; a gasket-like seal or liner to provide a waterproof seal; an indicator to assure proper alignment. The lid shall be inscribed or labeled "please tighten cap securely" or similar words. The lid shall prevent leakage of the contents regardless of the position of the container. The container shall be designed to accept a label which adheres under refrigeration. The label must include the patient's name, date and room number. The label shall be placed on the container, to ensure proper identification.

Type II - Sterile - 4-1/2 to 5 oz. (133 to 148 ml.) . capacity, with graduations of oz. and ml/cc in 1/4 oz. (7 ml.) and 10 ml/cc increments molded into the container. Peel open pouch (O.R. pack). The lid shall have full circumference screw rings; a gasket-like seal or liner to provide a leakproof seal; an indicator to assure proper alignment. The lid shall be inscribed or labeled "please tighten lid securely" or similar words. The lid shall prevent leakage of the contents regardless of the position of the container. The container shall be designed to accept a label which adheres under refrigeration. The label must include the patient's name, date and room number. The label shall be placed on the container, to ensure proper identification. The OR pouch shall permit aseptic handling of the cup and delivery of the cup, uncontaminated, onto a sterile

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

field. A tear-open type package shall not be acceptable. The label of the immediate container shall bear the word "sterile".

Salient characteristics. Size - B The specimen cup shall be a sterile, disposable, translucent, polystyrene or polypropylene container of 4-1/2 ounces plus or minus 1/4 ounce capacity, suitable for the collection of laboratory specimens. The cup shall be supplied with a screw cap which shall form a leakproof closure with the cup when tested as specified. The full cup shall be stable when placed on a flat, level surface.

The cup shall be circular in section, a nominal 2-1/2 inches in diameter and a nominal 2-1/2 inches in height. The cup shall be graduated in 1/4 ounce steps over a minimum range of 20 ml to 110 ml. Graduation numbers and lines shall be cleanly defined and easily read. The cup shall be free from waviness that obliterates the lines of the numerals.

The cup shall be threaded to mate with the threads on the cup. The screw cap shall fit the cup securely, shall be capable of being easily installed and removed from the cup, and shall form a tight seal to prevent spillage of contents.

The cup and cap shall be free from holes, pits, scratches, burrs, and abrasions. They shall be free from objectionable odor, water drops or dried stains, oil or grease smears, or dirt or loose particles, on either inside or outside.

One cup and one cap together shall be sealed in an individual clear plastic container or envelope of a peel-open type. This plastic container or envelope shall maintain the sterility of the contents until manually opened. It shall also permit aseptic handling of the cup and delivery of the cup uncontaminated, onto a sterile field. A tear-open type package shall not be acceptable. The label of the immediate container shall bear the word "sterile".

The plastic container or envelope shall be properly sealed to maintain sterility of contents. It shall be free of cuts or openings which permit entry of dust or foreign matter and strong enough to avoid holes, perforations, or tears during shipment and handling.

Labeling shall be that usually furnished by the supplier as commercial practice. Addition to these labels shall be limited to application of the following to the unit of issue (strip labeling is acceptable):

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

National Stock Number (NSN)
Date of packaging (month and year, i.e.,4/75). Final sterilization shall be on or after packaging date.

When the unit package is the shipping container, additional labeling may be applied.

Workmanship. The specimen container shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG) for size B. One package containing one hundred specimen cups, as specified, constitutes one unit. Package for size A will appear in the solicitation for bid.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government. The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Regards of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspections, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 4.0 (percent defective) and an inspection level of S-2. The unit of product for examination purposes shall be one specimen container as specified in this description.

Defects

Cup or cap not free of any deformation.

Scale numerals or lines not cleanly defined and easily read.

Cup not free from waviness that obliterates the lines or numerals.

Cup or cap not free from holes, pits, scratches, burrs and abrasions.

Cup not stable when placed on a flat surface.

Individual package not properly sealed to maintain sterility of contents.

Individual package not free of cuts or openings which permit entry of dust or foreign matter.

Cup or cap not free from objectionable odor, water drops or dried stains, oil or grease smears, or dirt or loose particles, on either inside or outside.

Sampling for tests. Sampling for tests including dimensional test, shall be conducted in accordance with MIL-STD 105, with an AQL OF 2.5 (percent defective) and an inspection level of S-2) The unit of product for test purposes shall be one specimen cup.

Tests

Sterility. Sterilization testing shall be in accordance with FDA's current "Guideline on General Principles of Process Validation". Failure to be in accordance shall constitute a defect.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Water-tightness (Size B only). The specimen cup shall be filled with tap water at room temperature to the top of its graduations. A cap shall then be firmly secured on the cup, and then the cup shall be inverted. The inverted cup shall rest on a sheet of paper which is placed on a flat surface. After fifteen minutes the sealed cup and paper shall be inspected for evidence of leakage, any evidence of leakage shall constitute a defect.

Capacity (Size B only). The cup shall be filled with water to within 1/8 inch (3.175 mm) of the top with water from a known graduated source. A capacity less than 4-1/4 ounces (126 ml) or greater than 4-3/4 ounces (140 ml) shall constitute a defect.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Numbers are covered by this document:

NSN	SIZE	TYPE	STYLE
6530-01-051-9344	A	I	I
6530-01-163-3704	A	II	I
6530-01-023-1160	A	I	II
6530-01-160-5915	A	II	II
6530-00-837-7472	B		

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description
(b) Type, style and size

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VANAC (904)
Agent VA-OSS VACO (92A)

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.