A-A-51654B
24 January 1992
SUPERSEDING
A-A-51654A
20 May 1987

COMMERCIAL ITEM DESCRIPTION

TUBE, TRACHEOSTOMY, DISPOSABLE
(Low Pressure Cuff, Shiley)

The General Services Administration has authorized the use of this comercial item description.

This commercial item description covers sterile, disposable, soft plastic, low pressure cuff tracheostomy tube, Shiley style, suitable for long term ventilation of tracheostomy patients who do not require high ventilation pressures.

Salient Characteristics:

General. The tube shall be either Type I or II. The Type I tracheostomy tube shall consist of an outer cannula with a cuff connected by tubing to a pilot balloon, an inner cannula and an obturator. The Type II tube shall be of Type I construction except being equipped with a pressure regulating valve.

Material. The outer cannula shall be fabricated of semi-rigid polyvinylchloride with the cuff, tubing and pilot balloon fabricated from polyvinylchloride. The inner cannula shall be fabricated from polypropylene and the obturator and flange from polycarbonate. All material shall be medical grade.

Outer cannula. The outer cannula shall have a soft, thin walled, high volume, low pressure cuff that uniformly distributes pressure against the tracheal wall. The cannula shall have a one-way Luer valve pressure retention system and a swivel neck plate for comfort and conformity to neck anatomy with the size clearly imprinted on the front outer surface of the neck plate. The outer cannula shall be radiopaque. The cuff shall be inflated through plastic tubing connected to a pilot balloon containing the one-way Luer valve. The cuff shall be designed to reach tracheal occlusion at internal cuff pressures of less than 25 mm Hg. The cannula shall be suitably curved to allow access to the trachea.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Sizes. The following sizes of Shiley tracheostomy tubes are covered by this document:

Size	O.D.	I.D.	Length
4	8.5 mm (26.0 Fr)	5.0 mm	67 mm
6	10.0 mm (30.0 Fr)	7.0 mm	78 mm
8	12.0 mm (36.0 Fr)	8.5 mm	84 mm
10	13.0 mm (39.0 Fr)	9.0 mm	84 mm

Inner cannula. The tracheostomy tube shall have a removable inner cannula with a twist lock and integral 15 mm connector. The inner cannula shall be easily removed and cleaned to maintain a patent airway without removing the outer tube. Inner cannula proximal end design shall conform to American National Standard Specification ANSI Z79.2.

Obturator. Item shall be provided with a smooth tapered obturator to simplify insertion and minimize trauma.

Sterility. The tracheostomy tube shall be sterile in accordance with the USP or AAMI requirements. Sterility assurance level shall be 0.000001.

Pyrogenicity and toxcity. The tracheostomy tube shall be free from pyrogens, non-toxic, and in accordance with the USP for Class IV plastics.

Workmanship. The tracheostomy tube shall be free from defects which detract from its appearance or impair its serviceability. Components shall be free from cuts, tears, weak sections, flash, dust, grease, or other foreign matter.

Unit.

Sizes 4, 6, and 8.

Each (EA). One tracheostomy tube, of one size, and type only as specifed, constitutes one unit. The individual package shall be a sealed peel-open type container capable of maintaining sterility of contents unless the package is damaged or opened.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Size 10.

Package (PG). One package containing ten individually packaged sterile tracheostomy tubes, as specified, constitutes one unit. The individual package shall be a sealed peel-open type container capable of maintaining sterility of contents unless the package is damaged or opened.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one tracheostomy tube. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one tracheostomy tube. Sampling shall be inspection level shall be S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Pressure Test. The cuff, tubing, ony-way valve and pilot balloon shall not burst, sustain any damage or leak when subjected to a pressure of 1.5 psi for 5 minutes. Inspection test level shall be S-3 with an AQL of 1.0.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers'own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers, their respective Item Identifications and Types are covered by this document:

NSN	ITEM IDENTIFICATION	TYPE
6515-01-147-0204	TUBE, TRACHEOSTOMY, Disposable, Sterile, Shiley, Surgical, Low Pressure Cuff, 7.0 mm Inside Diameter by 78 mm Long, 30.0 Fr, Size 6.	I
6515-00-001-9180	TUBE, TRACHEOSTOMY, Disposable, Sterile, Shiley, Surgical, Low Pressure Cuff, 9.0 mm Inside Diameter by 84 mm Long, 39.0 Fr, Size 10.	I

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

NSN	ITEM IDENTIFICATION	TYPE
6515-01-124-9780	TUBE, TRACHEOSTOMY, Disposable, Sterile, Shiley, Surgical, Low Pressure Cuff, 5.0 mm Inside Diameter by 67 mm Long, 26.0 Fr, Size 4.	II
6515-01-124-9781	TUBE, TRACHEOSTOMY, Disposable, Sterile, Shiley, Surgical, Low Pressure Cuff, 8.5 mm Inside Diameter by 84 mm Long, 38.0 Fr, Size 8.	II
6515-01-124-9782	TUBE, TRACHEOSTOMY, Disposable, Sterile, Shiley, Surgical, Low Pressure Cuff, 7.0 mm Inside Diameter by 78 mm Long, 30.0 Fr, Size 6.	II
6515-01-184-6706	TUBE, TRACHEOSTOMY, Disposable, Sterile, Shiley, Surgical, Low Pressure Cuff, 9.0 mm Inside Diameter by 84 mm Long, 39.0 Fr, Size 10.	II

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC SF010

Project No. 6515-4982
Location: AA51654B/S11

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-51654B	24 January 1992

3. DOCUMENT TITLE

TUBE, TRACHEOSTOMY, DISPOSABLE (Low Pressure Cuff, Shiley)

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center ATTN: DPSC-MSE	(215)7372870	444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 Previous editions are obsolete. 198/290

Source: https://assist.dla.mil/online/start/ as of 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.