A-A-50871A
28 July 1989
SUPERSEDING
A-A-50871
24 MAY 1983

COMMERCIAL ITEM DESCRIPTION

PREVENTIVE DENTISTRY KIT, PATIENT

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Interim Federal Specification GG-P-001428A (DSA-DM) which is cancelled.

This Commercial Item Description covers two sizes of patient preventive dentistry kits.

Salient characteristics:

General. The two kit sizes shall be child and adult. The child kit shall be suitable for professional instruction of pedodontic (child) patients and the adult kit for professional instruction of adult patients in the proper procedures for oral hygiene consistent with a complete preventive dentistry program. The child and adult kits shall contain dentifrice, (1) toothbrush, (2) plaque disclosing tablet, pamphlet/instructions and floss (adult kit only) as specified herein, in a suitable disposable container.

Dentifrice. The kits shall contain one (1) squeeze, collapsible tube containing 0.8 to 1.5 ounces of a palatable, flouride-containing dentifrice which has been classified as an accepted flouride dentifrice by the Council on Dental Therapeutics of the American Dental Association (ADA). The adult dentifrice shall be a tartar control dentifrice. Flavoring of the dentifrice is optional. The optional flavoring for the adult kit shall be a mint or wintergreen flavoring. The pH of the dentifrice shall be 6.8 to 8.3 when tested as specified in test section under pH. Consistency of the dentifice shall be 18 $\pm$ 6 Brookfield units when tested as specified in test section under consistency.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release, distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

The side wall of the tube shall be a paper-aluminum foil-polyolefin laminate structure which shall be side seamed and sealed (to prevent leakage) to form a tube. The high density polyethylene head and neck shall be injection molded on the sidewall sleeve and shall hold a urea disk. The disk shall provide rigidity and favorable barrier properties. Polypropylene screwcap shall be properly threaded to insure adequate closure of the dentifrice tube. Dentifrice shall be free from foreign matter and shall have a pleasant taste and odor.

Marking of the exterior of the dentifrice tube shall include the brand identification, net weight of contents, and name and address of manufacturer. Each dentifrice tube shall be marked with the lot or batch number, or suitably code dated to relate to the lot or batch number.

The tube shall have a clearly marked label.

Toothbrush. The kits shall contain one (1) toothbrush. The brush block shall be molded of Polypropylene in conformance with ASTM D4101, or cellulose acetate propionate in conformance with grade H-3 of ASTM D1562.

Materials and dyes used shall be nontoxic and suitable in all respects for the purpose intended. A dye shall be considered acceptable for use in toothbrushes provided its concentration is less than 0.3 percent, and the dye has been used for this application for a minimum of 3 years without toxic effects.

Brush block of the toothbrush shall consist of a head, neck, and a handle gripping section. Toothbrush (block) shall be of straight line design and shall have at least one flat surface to facilitate thorough tactile sense, the orienting of the bristles in the desired direction and the maintenance of the proper position.

As an alternative to straight sides, sides may be cut out to form a decorative design.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

The toothbrush block shall be any color to include clear (colorless). Block shall have a polished finish, free from flash, rough mold marks and sharp projections. The toothbrush bristles shall be fabricated of natural (uncolored) nylon 610 or 612 filaments conforming to class N, Type II (Polyamide, level) of H-S-951.

Filament ends shall be splayed no more than 15 degrees from the perpendicular to the axis (handle) of the toothbrush when tested as specified in test section under Tuft angle.

Filaments shall be free from sharp points and rough edges and shall be straight, flat, even-trimmed, and end-rounded to approximate hemispheric shape to preclude trauma to the hard and soft tissues of the mouth.

Filaments shall have a bright, smooth finish.

Tufts of bristles shall be firmly secured in holes with a staple. The staple shall be a corrosion-resisting nickel-silver flattened strip of metal. Tufts shall have a flat trim height profile.

The toothbrushes shall conform to the requirements of table I. Tolerance on filament diameters shall be ±8 percent. The bristles shall be arranged in aligned rows (4 rows for the Adult brush, and 3 rows for the Child brush) with respect to the long axis of the head, except for the tufts at the lead (free) end which may be staggered.

TABLE I.

	Size 1 (Adult)	Size 2 (Child)
Overall Length (inches)	6-1/4 ± 3/8	5-1/4, +3/8, -1/8
Head Length (inches)	1-1/4, +1/16, -5/32	1-1/16, +5/32, -1/8
Head Width, Widest Section (inches)	1/2, +1/64, -3/16	5/16, +1/8, -1/32
Filament Ends per Tuft	30 to 44 /1, /2, /3	28 to 34 /4, /5
Bristle Trim Height (inches)	7/16, +1/16, -1/32	3/8 to 7/16

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

TABLE I (cont)

	Size 1 (Adult)	Size 2 (Child)
Number of Tufts	41 to 52 /1, /2, /3	27 to 30 /4, /5
Filament Diameter (inch)	.008 /1, /2, /3	.008 /4, /5

/1 As an alternate, a brush with 51 tufts, 48 to 52 filament ends per tuft, having .007 inch (or a combination of .007 inch and up to 25 percent of .006 inch) diameter filaments shall be acceptable.

/2 As an alternate, a brush with 43 tufts, .007 inch diameter filaments, and 42 to 44 filament ends per tuft shall be acceptable.

/3 As an alternate, a brush with 41 tufts, .007 inch diameter filaments, and 44 to 48 filament ends per tuft shall be acceptable.

/4 As an alternate, a brush with 30 tufts, .007 inch diameter filaments, and 48 to 52 filament ends per tuft shall be acceptable.

/5 As an alternate, a brush with 27 tufts, .007 inch diameter filaments, and 44 to 58 filament ends per tuft shall be acceptable.

Toothbrushes shall be marked with the name or trademark of the contracter or manufacturer and the name of the country of origin. The use of an identification code for marking is permissable provided approval is obtained from the contracting officer prior to the award.

The toothbrush shall be individually wrapped and sealed in cellophane.

An adult brush shall be supplied in the adult kit. A child brush shall be supplied in the child kit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Plaque Disclosing Tablets. The kit shall contain a minimum of two(2) plaque disclosing tablets. The tablets shall be palatable, chewable and dissolve readily in mouth fluids. Each tablet shall contain 5 mg of certified D+C Red 28 (Disodium Xanthine). This dye shall be capable of staining food deposits and plaque not removed in brushing the teeth. Tablets shall comply with the Content Uniformity Test requirements as specified in the USP. Tablet shall have a hardness of not less than 4.0 and not more than 10.0 Stray Cobb Units. Tablets shall assay to contain not less than 90.0 percent and not more than 110.0 percent when assayed by a suitable, accurate and reproducible method. Material used in the manufacture of the tablets shall be of a pharmaceutical grade. Tablets shall not liquefy, soften or disintegrate at storage temperatures up to 140 degrees Fahrenheit (^{0}F). The tablets shall be individually sealed in commercially available tear-off blister or strip packaging which dispenses one (1) tablet at a time.

Pamphlet/Instructions The kits shall contain one (1) Dental Health Pamphlet with instructions in patient oral hygiene and use of disclosing tablets. As an alternate, instructions may be printed on the individual kit container.

Floss. The adult kit shall contain a minimum of ten yards of waxed dental floss in a disposable roll type dispenser with a cutter. Floss is an optional component in the child kit. The floss shall be in accordance with Type I and Shape I of Commercial Item Description A-A-53652 with the length and put up as specified herein.

Container Marking. Kit container shall be marked with the contents, dental health rules, directions for use of the disclosing tablets to check brushing technique and other applicable markings as required by USP. Marking with additional information is optional.

Workmanship. Each kit component shall be free of foreign matter, tears, damage, leakage of its components or any defect which detracts from its appearance or impairs its serviceability.

Unit. Package (PG), One package containing thirty six Preventive Dentistry Kits, of the size specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one kit.

Examination. Sampling for examination shall be conducted at inspection level S-2, AQL 2.5. Kits shall be examined for defects including, but not limited to those listed below.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Defects

Wrong sized toothbrush furnished for applicable kit.

Dentifrice tube not completely crimped at seal to prevent leakage.

Tube not properly threaded; stripped threads on tube or cap.

Cap loose on tube.

Foreign matter in dentifrice.

One disclosing tablet damaged to the extent to prevent normal usage.

Tablet container not properly sealed, closure inadequate.

Toothbrush filaments crimped.

Cutting device missing on floss dispenser in adult kit.

No floss in adult kit.

Tests. Sampling for tests shall be conducted at inspection level S-1, AQL 1.0.

pH. pH of the dentifrice shall be tested in accordance with ASTM E70. A 3 to 1 water to dentifrice slurry shall be used.

Consistency. Consistency of the dentifrice shall be measured using the "E" spindle on a Brookfield viscosimeter. Consistency shall be determined in accordance with method 4287 of FED-STD-141, at $85^{0}F \pm 1^{0}F$ and at 2 revolutions per minute (r. p. m.).

Tuft angle. The toothbrush shall be placed on a level surface with the head resting on the surface. By means of a bevel protractor, determine the angularity of the tufts at the end and side rows.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Certificates of compliance. Certificates of compliance, supplied by the manufacturer of the toothbrushes and floss, may be furnished in lieu of actual performance of such testing by the contractor. When certificates of compliance are submitted, the Government reserves the right to check test such items to determine the validity of the certification.

Contracter certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Goverment reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended,and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

NOTE 1: The following National Stock Numbers are covered by this document:

NSN	Item Identification	SIZE
6520-00-086-6554	PREVENTIVE DENTISTRY KIT, PATIENT Child, 36s.	Child
6520-00-890-2080	PREVENTIVE DENTISTRY KIT, PATIENT Adult, 36s.	Adult

NOTE 2: The following National Stock Number represents the plaque disclosing tablet:

NSN	Item Identification
6520-01-261-6633	PLAQUE DISCLOSING TABLETS, DENTAL, 5 mg, Individually Sealed, 1000s

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent:

DLA - DM

Civil Agency Coordinating Activities:

VA-OSS
PHS
FDA-MPQAS

Project No. 6520-2115

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-50871A	PREVENTIVE DENISTRY KIT, PATIENT

a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil - Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.