COMMERCIAL ITEM DESCRIPTION

DRAPE, SURGICAL, SPLIT SHEET, NONWOVEN FABRIC

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a disposable surgical drape.

Salient characteristics.

General. Shall be a sterile, sealed pack containing one disposable split drape and one outer wrap, suitable for use in surgical operating rooms of Military treatment facilities. Drape and outerwrap shall be made of a non-woven fabric. The drape shall be constructed of fully bleached pulp tissue bonded to nylon or rayon reinforced yarns and treated for moisture repellency sufficiently to be repellent to low surface tension fluids. Tissue of the drape shall be approximately 40 pounds basic weight (18 kg). The drape shall be approximately 120 inches (305 cm) long by 87 inches (221 cm) wide. An aperture approximately 44 inches (112 cm) long, and not less than 1-1/2 inches, nor more than 4 inches (10 cm) wide, shall be centered at the tail or foot end.

The aperture shall be reinforced with an absorbent but impervious laminate panel approximately 36 inches (91 cm) long by 18 inches (46 cm) wide. The reinforcement panel shall border approximately 14 inches (36 cm) along aperture's length when measured from the inner end.

Tensile strength. The tensile strength shall be 10 lbs per inch (1.8 kg per cm) minimum in the machine direction. The tensile strength in the cross direction shall be 8 lbs per inch (1.44 kg per cm) minimum.

Weight. The drape shall weigh 2 ounces per square yard (67 g per square meter).

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Hand of fabric. Hand of fabric shall be 25 grams in the machine direction and 35 grams in the cross direction when tested in accordance with MIL-P-36901A.

Electrostatic. The surface resistivity of this fabric shall be 5X10" OHMS per square inch at 5% relative humidity.

Bursting strength. Shall be 18 lbs, (8 kg) minimum.

Color. Colors shall be pastel shades of blue, green, or grey on minimum of one glarefree face.

Outerwrap. The drape shall be enclosed within a resin bonded, nonwoven fabric outerwrap. The minimum tensile strength of outerwrap fabric shall be 4.5 lbs per inch (0.8 kg per cm) machine direction and 1.5 lbs per inch (0.28 kg per cm) cross direction. The outerwrap shall be white or pastel shades of blue, green, or grey.

The drape and outerwrap shall be functionally folded for optimum aseptic technique. Folded drape shall be clearly marked with the following or similar identifications: Split sheet, tails (with arrow pointing in direction of tails).

The drape and outerwrap shall be abrasive resistant and free of lint, objective noises when used and objective odors. Shall have no irritating or toxic elements present in the finished fabric. Drape and outerwrap shall be enclosed in a sealed paper or plastic intermediate container capable of maintaining sterility of contents until manually opened. If the sealed container is paper, it shall be further enclosed in a plastic bag. The pack components shall meet the flammability requirements outlined in method 5908 Federal Test Method Standard 191 with a minimum resistance time of 3.5 seconds and shall satisfy other applicable requirements of the Federal Flammable Fabrics Act. The components shall comply with National Fire Protection Association Code Pamphlet and No. 56.

Labeling. Labeling shall comply with the Federal Food, Drug and Cosmetic Act. The label of the container on which the Federal Stock /Number appears shall bear the following: Date of packaging (month and year). Final sterilization shall be on or after indicated packaging date.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:22Z
Check the source to verify that this is the currentversion before use.

Sterilization. Sterilization shall comply with FDA requirements.

Workmanship. The surgical drape shall be free from from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing fifteen sealed packaged sterile drapes with an outerwrap for each drape, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting of rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6530-01-262-8681	DRAPE, SURGICAL, SPLIT SHEET, NONWOVEN FABRIC Disposable, 87 By 120 Inches (221 By 305 cm), 15s.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-2012

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
PHILADELPHIA, PENNSYLVANIA 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53662	DRAPE, SURGICAL, SPLIT SHEET, NONWOVEN FABRIC

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION (Mark one)
	☐ VENDOR
b ADDRESS (Street, City, State, ZIP Code)	☐ USER
	☐ MANUFACTURER
	☐ OTHER (Specify)

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER (Last, First, MI) – Optional	b WORK TELEPHONE NUMBER (Incl Area Code) – Optional
c MAILING ADDRESS (Street, City, State, ZIP Code) – Optional	8 DATE OF SUBMISSION (YYMMDD)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 202610-26109:22Z
Check the source to verify that this is the current version before use.