A-A-53663
1 September 1988

COMMERCIAL ITEM DESCRIPTION

COUNTING APPARATUS, BACTERIAL COLONY

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-C-37760 which is cancelled.

This commercial item description covers a Quebec type bacterial counting apparatus suitable for counting bacterial colonies.

Salient characteristics:

The counting apparatus shall consist of a housing, a lens, a 40 Watt lamp or equivalent fluorescent tube , an electric connecting cord and a Wolffhuegel counting plate.

The counting apparatus shall be compatible with the replacement Wolffhuegel counting plate.

The lens shall have a magnification of 1.5X and shall be approximately the same diameter as the counting plate. The lens shall be mounted on a sliding rod for focusing by the user.

The housing shall be inclined at an angle and shall contain an opening large enough in diameter through which the counting plate can be seen. An auxiliary tilting device shall be provided for changing the front panel slope as required.

A centering device for the petri dish shall be provided on the surface of the case, and provisions for easily removing the light source shall be made. Shall be capable of operating on 115V, 60Hz supply lines.

Shall be in compliance with all applicable requirements of the Bureau of Medical Devices and Diagnostics.

Workmanship. The colony counting apparatus shall be free from defects which detract from or impair its serviceability.

Unit. Each (EA). One counting apparatus, as specified, constitutes one unit.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Aquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

6640-01-034-0479 COUNTING APPARATUS, BACTERIAL COLONY

The following is the replacement Wolffuegel Counting Plate:

6640-00-418-8205 PLATE, COUNTING, BACTERIOLOGICAL

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.)

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6640-1490

Location: Disk # 6

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
PHILADELPHIA,PENNSYLVANIA 19101-8419

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:22Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER
A-A-53663

2 DOCUMENT TITLE
COUNTING APPARTUS, BACTERIAL COLONY

3a. NAME OF SUBMITTING ORGANIZATION

b ADDRESS *(Street, City, State, ZIP Code)*

4 TYPE OF ORGANIZATION *(Mark one)*

- ☐ VENDOR
- ☐ USER
- ☐ MANUFACTURER
- ☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b Recommended Wording:

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional

b WORK TELEPHONE NUMBER *(Incl Code)* – Optional

c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional

8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:22Z
Check the source to verify that this is the current version before use.