INTERIM COMMERCIAL ITEM DESCRIPTION

BANDAGE, ELASTIC, RUBBER WARP THREADS, SELF-ADHERING

The General Services Administration has authorized the use of this Interim Commercial Item Description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This Interim Commercial Item Description covers cotton, self-adhering, elastic bandages.

Salient characteristics:

Shall be self-adhering cotton elastic bandages reinforced with rubber warp threads suitable for use in the treatment of phlebitis and for controlling varicosities of the lower extremities.

Bandages shall be of the following sizes:

SIZE	WIDTH (INCHES +/- 1/8 INCHES)
1	4
2	6

The bandage shall be woven of 30 cotton yarns per inch in the warp direction and 10 cotton covered heat resistant natural rubber strands (size 90 to 100) per inch in the warp direction and 72 cotton yarns per inch (no. 20/2 cotton yarns) in the filling direction. The yarns and strands per inch shall be counted in accordance with test method 5050 of Federal Test Method Standard 191. Both edges of the bandages shall be selvaged. The bandage material shall have a nominal weight of 8 ounces per square yard (in the relaxed condition) when measured by method 5041 of Federal Test Method Standard 191. The bandage material shall have a minimum tensile strength of 20 pounds in the warp direction and 50 pounds in the filling

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:23Z
Check the source to verify that this is the current version before use.

direction when tested in accordance with method 5100 of Federal Test Method Standard 191. The bandages shall be of flesh color and the finished bandage shall show good colorfastness to water when tested in accordance with method 5630 of Federal Test Method Standard 191.

The bandages shall be coated with a self-adhering substance on both sides. The self-adhering substance shall be a latex cohesive substance of the following composition:

1.	Dry rubber as latex	100 parts
2.	Potassium salt of oleic acid	1 part per hundred
3.	2, 2'- methylenebis (4-ethyl-6-tert-butylphenol)	2 parts per hundred
4.	Emulsified silicone oils in hydrocarbon solvents	0.05 parts per hundred
5.	Sodium polyacrylate	0.5 parts per hundred

The latex coating substance shall be a minimum of 5% to a maximum of 15 percent of the finished bandage weight and shall be applied so as not to fill the interstices between the yarns. The self-adhering properties of the bandages shall be the result of the application of the latex coating substance to both sides of the elastic fabric:

The finished bandages shall have an unstretched length not less than 54 inches long and shall have a minimum elasticity of 140 percent and shall have a return capacity to within 110 percent of its original length when measured as follows. Using a bandage not previously stretched, mark off a ten inch portion in the relaxed condition. Stretch the ten inch portion to 24 inches for one hour, then release. The marked off portion of the bandage shall be measured again without tension.

The finished bandage shall adhere to itself, shall not adhere to skin and shall not adhere to overdressings. The finished bandage shall be capable of maintaining constant pressure as applied and shall remain in place as applied.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

The finished bandages shall be capable of withstanding sterilization by steam without loss of elasticity and without loss of self-adhering properties and shall remain colorfast during steam sterilization when tested by autoclaving at 250°F for not less than 30 minutes. Bandages shall be tested for loss of elasticity and loss of self-adhering properties. Bandages shall be tested for colorfastness by wrapping bandages in white gauze sponges and examining both the gauze sponge and bandages after autoclauing. Any visual change in color of the sponges shall be a test failure.

The bandages shall be one continuous piece, evenly and neatly rolled on a paperboard core without tension or gapping. The finished bandages shall not be irritating to the skin and shall be nontoxic. The finished bandages shall be clean and shall be free of cuts, ravelings, loose threads and wrinkles.

Instructions for use shall be printed on each individual bandage package or shall be supplied as an insert.

Workmanship. The bandages shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). Each bandage shall be of one continuous piece, evenly and neatly rolled on a paperboard core, without tension or gapping and shall be suitably wrapped. One package containing 10 wrapped bandages, as specified, constitutes one unit.

Quality Assurance Provisions

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:23Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Testing and examination shall be conducted to determine compliance with all specification requirements. Test methods shall be suitable, accurate and reproducible.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. the unit of product for examination shall be one bandage. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5 maximum percent defective.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:23Z
Check the source to verify that this is the currentversion before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	SIZE
6510-00-103-9749	BANDAGE, ELASTIC, RUBBER WARP THREADS, 4 inches by 1½ yards, 10s	1
6510-01-097-2199	BANDAGE, ELASTIC, RUBBER WARP THREADS, 6 inches by 1½ yards, 10s	2

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 05100

Project No. 6510-0764

☆U.S. GOVERNMENT PRINTING OFFICE: 1989 - 604-033/10860

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.