A-A-53689
15 September 1988

COMMERCIAL ITEM DESCRIPTION

LUBRICANT, SILICATE CEMENT, DENTAL

The General Services Administration has authorized the use of this Commercial Item Description in preference to Federal Specification P-L-790(1) which is being cancelled.

This Commercial Item Description covers a prepared dental silicate cement lubricant.

Salient characteristics:

General. Shall be a dental lubricating compound suitable for use to protect freshly inserted silicate restorations against moisture, and to lubricate strips and disks during finishing procedures.

The immediate container shall be a collapsible metal tube. The neck of the tube and inner face of cap shall be designed to form an effective seal which will permit reclosure and prevent loss of contents.

Material. The lubricant shall be of the following composition:

White petrolatum, U.S.P.	.40 percent
Hydrogenated vegetable oil	.50 percent
Theobroma oil, U.S.P.	5 percent
White wax, U.S.P.	5 percent
Flavoring ingredients, U.S.P.	0.5 part (by formula)

The white petrolatum, theobroma oil, white wax and flavoring ingredients shall be in accordance with the tests, standards, and requirments of the Pharmacopeia of the United States of America (U.S.P.).

Any ingredients entering into the lubricant which are not included in the (U.S.P.) or National Formulary (N.F.) shall be of quality suitable for pharmaceutical use.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release, distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

Physical properties.

Extrusion after softening. The lubricating compound shall not liquefy when tested as specified in test section under Extrusion after softening.

Extrusion after chilling. Excessive pressure shall not be required to extrude material from tube after testing as specified in test section under Extrusion after chilling.

Freezing and thawing effects. The material shall show no adverse effects such as graininess, separation, or loss of smooth texture when tested as specified in test section under Freezing and thawing effects.

Performance. The compound shall be of such consistency that, when applied to the teeth in the usual manner, it will not liquefy and run on the teeth.

Delivery schedule. No more than 2 months shall have elapsed from date of manufacture of the lubricant to date of delivery to the Government.

Irritation. Unless otherwise specified, prior to delivery, the manufacturer shall furnish a statement directly to the Directorate of Medical Materiel, Techical Services Branch, DPSC-RST, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101-8419 stating that the material does not normally irritate oral tissues and does not contain poisonous ingredients in sufficient concentration to be harmful to human beings when used as directed.

Workmanship. The lubricant shall be smooth and homogeneous and shall be free from defects which may detract from its appearance or may impair its serviceability.

Unit. Tube (TU). One tube containing a minimum of 21 grams, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the raw materials and packaging materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Inspection of raw materials. The supplier is responsible for insuring that raw materials used were manufactured, tested and inspected in accordance with the requirements of referenced subsidiary specifications to the extent specified, or, if none, in accordance with this specification. In the event of conflict, this specification shall govern.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more characteristics. The unit of product for inspection purposes shall be one tube of lubricant. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests shall be conducted at inspection level S-1, acceptance level shall be zero for all sample sizes.

Examination. Lubricant shall be examined to determine compliance with specification requirements. The lubricant shall be examined for defects, including but not limited to, those listed below.

Defects

Amount of contents in immediate container not as specified.

Splits, rips, tears, or cracks present in immediate container.

Cap does not fit tightly over neck of tube.

Foreign material present in lubricant.

Tube leaks.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Extrusion after softening. A tube of the paste shall be submerged in a water bath at 37 ± 0.1°C for 30 minutes. The tube shall then be removed from the bath and a portion (no less than 1/2 inch in length) shall be extruded onto a glass plate measuring approximately 4 inches by 5 inches.

Extrusion after chilling. A tube of paste shall be submerged in a water bath at 15 ± 0.2°C for 30 minutes. The tube shall then be removed from the bath and a portion extruded as described in extrusion after softening test.

Freezing and thawing effects. A tube of paste shall be refrigerated at -4 ± 0.2°C. for 24 hours. The tube shall then be allowed to thaw for 12 hours. The lubricant shall then be examined for graininess, separation, and texture.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the the salient characteristics of this description and conform to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

Stock Number	Item Identification
6520-00-539-2150	LUBRICANT, SILICATE, DENTAL, 21 grams

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent:

DLA - DM

Civil Agency Coordinating Activities:

VA-OSS
PHS
FDA-MPQAS

Project No. 6520-2151

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53689	LUBRICANT, SILICATE CEMENT, DENTAL

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.