A-A-53690
15 September 1988

COMMERCIAL ITEM DESCRIPTION

VEST ASSEMBLY, TRACTION APPARATUS

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a traction apparatus vest assembly used in conjunction with a cervical halo unit to stabilize cervical spine fractures.

Salient characteristics.

The unit shall be cervical immobilization, traction apparatus, vest assembly containing all items required for use with a cervical halo assembly. The vest shall be provided in the sizes indicated below.

Size	Chest measurement
Small	26 to 33 inches (66 to 84 cm)
Medium	33 to 40 inches (84 to 102 cm)
Large	40 to 46 inches (102 to 117 cm)

The assembly shall consist of at least the following items:

Halo vest with liner.
Halo support structure.
Support attachments.
Tools for application and adjustment.

Vest. Based on the circumference at the xyphoid process, the vest shall fit a male or female with chest measurements within the specified range. The vest shall include an outer shell, a liner, preattached chest and back bars, straps and buckles. The vest shall have a two piece construction.

AMSC/NA FSC 6515
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

The shell of the vest shall be fabricated of a rigid, light weight and durable plastic material. The liner shall be fabricated of washable lambs wool for patient comfort, cleanliness and hygiene. The liner shall attach to the shell by hook and pile closures. The liner shall be slightly larger than the shell to protect the patient's skin from abrasion. Straps and buckles shall be permanently attached and shall be adjustable to accomodate different chest sizes. Strap and buckle shall be provided for each side and shoulder. Chest and back bars shall be permanently secured to the vest to support the halo structure. Mounting hardware shall not contact patient skin.

Support structure. Support structure shall consist of all hardware necessary to support cervical halo units used with patients from 17 to 26 inch head circumference range. Support structure shall be fabricated of aluminum which has been properly hardened and passivated. The structure shall be secured to the halo via 1/4" - 28 thread halo holes. All hardware shall be fabricated of corrosion resistant material.

Workmanship. The vest assembly shall be from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One vest assembly with completed instructions for use, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105. Sampling shall be conducted at inspection level II with an AQL of 2.5.

Tests. The vest assembly shall be assembled and mounted on mannequins of the maximum and minimum size limits of the vest under test. The vest shall be examined for fit and compatibility with halo.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE 1: This Commercial Item Description covers the following items:

NSN	ITEM IDENTIFICATION
6515-01-277-1555	VEST ASSEMBLY, TRACTION APPARATUS, Cervical Immobilization, Small, 26 to 33 inch
6515-01-277-1556	VEST ASSEMBLY, TRACTION APPARATUS, Cervical Immobilization, Medium, 33 to 40 inch
6515-01-277-1554	VEST ASSEMBLY, TRACTION APPARATUS, Cervical Immobilization, Large, 40 to 46 inch

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

NOTE 2: The vest assembly described herein is used in conjunction with the following items:

NSN	ITEM IDENTIFICATION
6515-01-277-1557	CERVICAL UNIT, TRACTION APPARATUS, Halo, Small, 17 to 20 inch
6515-01-277-1558	CERVICAL UNIT, TRACTION APPARATUS, Halo, Medium, 20 to 23 inch
6515-01-277-1559	CERVICAL UNIT, TRACTION APPARATUS, Halo, Large, 23 to 26 inch

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4439

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53690	VEST ASSEMBLY, TRACTION APPARATUS

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

(FOLD DETACH THIS SHEET AS INDICATED)

(FOLD ALONG THIS LINE)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (DO NOT STAPLE) and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4956 WASHINGTON D.C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-
2300 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.