A-A-53692
16 September 1988

COMMERCIAL ITEM DESCRIPTION

ELECTROCARDIOGRAPH, GENERATOR/CALIBRATOR

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a portable, battery operated generator/calibrator suitable for use in calibrating and testing electrocardiograph (ECG) instruments.

Salient characteristics.

The calibrator shall be a portable, battery operated electrocardiograph calibrator suitable for use with three, five and twelve lead electrocardiograph instruments. The calibrator shall include all necessary switches, controls, and connectors required for use. The unit shall be supplied with a carrying case.

The calibrator shall produce PQRST complex ECG waveforms. The unit shall provide signals to enable an ECG to produce 12 graphs indicating signal output for leads I thru V6 (I, II, III, AVR, AVL, AVF, and VI thru V6). The calibrator shall have a minimum of two fixed rate outputs of 30 and 60 BPM ± 2 percent and a continuous variable output of at least 30 to 240 BPM ± 5 percent or at least 2 additional fixed rate outputs between 90 and 240 BPM ± 2 percent. The output amplitude shall be selectable and include at least 0.5 and 1.0 millivolt (mv) and at least one additional mv output with an accuracy of ± 2 percent a peak. All controls, switches and connectors shall be clearly marked.

The unit shall be provided with and operate from an internal battery which shall provide at least 50 hours of continuous operation. The unit shall also provide a battery check function. The unit shall be capable of operating in an ambient temperature range of 15^{0}C to 35^{0}C without affecting calibration accuracy.

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

The unit and any related accessories shall be contained in a reuseable carrying case. The case may be an integral part of the units design. The case shall be capable of protecting the unit during transport and storage.

Each unit shall be supplied with instruction manual(s). The manual shall include technical specifications and operating instructions.

The calibrator dimensions (without carrying case) shall not exceed 0.1 ft^3. The weight (without carrying case) shall not exceed 2 pounds.

Workmanship. The electrocardiograph generator/calibrator shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One electrocardiograph generator/calibrator, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defines as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

Sampling. Sampling for testing shall be conducted in accordance with MIL-STD-105 and as specified. The unit of product for sampling purposes shall be one ECG generator/calibrator. Inspection shall be at inspection level II with an AQL of 2.5 defects per hundred units.

Tests. Tests shall be conducted to determine compliance with specified requirements. Testing shall incorporate the use of a 12 lead ECG. Testing shall include the following:

Verify signal output is within specified tolerances for each of the 12 lead output signals.

Verify battery check is functioning properly and battery is fully charged (per manufacturers specs).

Verify unit is capable of producing PQRST complex waveforms.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: This Commercial Item Description is covered by the following item:

NSN	ITEM IDENTIFICATION
6515-01-049-9449	ELECTROCARDIOGRAPH, CALIBRATOR-GENERATOR, Portable, Battery Operated

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPOAS

Project No. 6515-4431

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53692	ELECTROCARDIOGRAPH, GENERATOR/CALIBRATOR

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street City State ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last First MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil — Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

STRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in bmitting comments and suggestions for improvements All users of military standardization documents are invited to provide ggestions This form may be detached, folded along the lines indicated, taped along the loose edge (DO NOT STAPLE), and ailed In block 5, be as specific as possible about particular problem areas such as wording which required interpretation was so rigid, restrictive loose ambiguous or was incompatible, and give proposed wording changes which would alleviate the roblems Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out an knowledgement will be mailed to you within 30 days to let you know that your comments were received and are being onsidered

OTE This form may not be used to request copies of documents nor to request waivers, deviations, or clarification of ecification requirements on current contracts Comments submitted on this form do not constitute or imply authorization waive any portion of the referenced document(s) or to amend contractual requirements

(Fold along this line)

(Fold along this line)

EFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

IFFICIAL BUSINESS
ENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.