A-A-53693
16 September 1988

COMMERCIAL ITEM DESCRIPTION

CERVICAL UNIT, TRACTION APPARATUS, HALO

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a cervical halo unit used in conjunction with vest assemblies to stabilize spine fractures.

Salient characteristics.

The unit shall be a cervical halo unit containing all items necessary to apply a cervical halo on a patient. The unit shall be be provided in the sizes indicated below:

Size	Head Circumference Range
Small	17 to 20 inch (43 to 51 cm)
Medium	20 to 23 inch (50.8 to 58.4 cm)
Large	23 to 26 inch (58.4 to 66 cm)

The unit shall consist of at least the following items:

QTY	Descriptions
1	Halo ring, aluminum
3	Position pads
5	Skull pins
1	Combination wrench
1	Torque screwdriver

All metal items shall be fabricated of corrosion resistant alloy properly hardened and passivated.

AMSC/NA FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

Halo. The lateral sides of the halo ring shall have multiple holes through which the skull positioning pins and skull halo pins can be inserted. The halo shall be fabricated with stainless steel inserts to form the halo holes. Halo holes shall be size 1/4 inch-28 thread. The location of the holes on one side of the halo ring shall be spaced identically to those on the opposite side. Sufficient numbers of holes shall be provided for pins and mounting to halo vest. The halo shall have one hole at the center front for proper alignment over the bridge of the nose. The occipital position of the halo shall form an upward curve. The halo shall be suitable for use with a cervical vest assembly.

Position pins. The unit shall contain 3 position pins with nylon pads to aid in the alignment of the halo ring before the skull pins are inserted. Position pins shall be size 1/4 inch-28 thread.

Skull pins. The unit shall contain 5 titanium skull pins with lock nuts and removable plastic protectors for the tips of the pins. Pin shall be size 1/4 inch-28 thread. Pins and locknuts shall be rustproof. Pin shall be sterilizeable up to 270^{0}F. The point of the pin shall be designed to draw the skin under it without causing bleeding. The pins and halo shall be able to withstand at least 6 inch/pounds of torque. The assembly shall meet applicable FDA requirements for safety and toxicity for invasive/implantable devices.

Tools. A 6 inch/pound torque screwdriver shall be provided for adjustment of the skull pins. A 7/16 inch combination wrench shall be provided for locking the skull pins.

Workmanship. The cervical unit shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One cervical unit with completed instructions for use, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract of purchase order, the contractor may use his own or nay facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspection set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defines as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for testing shall be conducted in accordance with MIL-STD-105. Sampling shall be conducted at inspection level 2.5 with an AQL of 2.5 defects per hundred. The unit of product for sampling purposes shall be one cervical unit.

Tests. The halo ring shall be assembled and mounted on mannequins of the maximum size limits of the halo under test. The halo shall be inspected for clearance all around the head.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

NOTE 1: This Commercial Item Description covers the following items:

NSN	ITEM IDENTIFICATION
6515-01-277-1557	CERVICAL UNIT, TRACTION APPARATUS, Halo, Small, 17 to 20 inch.
6515-01-277-1558	CERVICAL UNIT, TRACTION APPARATUS, Halo, Medium, 20 to 23 inch.
6515-01-277-1559	CERVICAL UNIT, TRACTION APPARATUS, Halo, Large, 23 to 26 inch.

NOTE 2: The cervical unit described herein is used in conjunction with the following items:

NSN	ITEM IDENTIFICATION
6515-01-277-1555	VEST ASSEMBLY, TRACTION APPARATUS, Cervical Immobilization, Small, 26 to 33 inch.
6515-01-277-1556	VEST ASSEMBLY, TRACTION APPARATUS, Cervical Immobilization, Medium, 33 to 40 inch.
6515-01-277-1554	VEST ASSEMBLY, TRACTION APPARATUS, Cervical Immobilization, Large, 40 to 46 inch.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4438

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53693	CERVICAL UNIT, TRACTION APPARATUS, HALO

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street City State ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *Last First MI* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (*DO NOT STAPLE*) and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation was too rigid, restrictive, loose, ambiguous, or was incompatible and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN. DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.