MIL-A-36495A(DSA-DM)
28 October 1975
SUPERSEDING
MIL-A-36495
27 June 1966

# MILITARY SPECIFICATION

# AUDIOMETER, GENERAL DIAGNOSTIC

This specification is approved for use by all Departments and Agencies of the Department of Defense.

## 1. SCOPE

1.1 This specification covers one type of single channel, pure tone and speech audiometers suitable for use in measuring the sensitivity of hearing.

## 2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

| | |
|---|---|
| PPP-B-601 | Boxes, Wood, Cleated-Plywood. |
| PPP-B-636 | Boxes, Shipping, Fiberboard. |

Military

| | |
|---|---|
| MIL-P-116 | Preservation, Methods of. |

STANDARDS

Military

| | |
|---|---|
| MIL-STD-105 | Sampling Procedures and Tables for Inspection by Attributes. |
| MIL-STD-129 | Marking for Shipment and Storage. |

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:41Z
Check the source to verify that this is the current version before use.

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

AMERICAN NATIONAL STANDARDS INSTITUTE (ANSI)

| | |
|---|---|
| C73.11-1966 | Plugs and Receptacles: General Purpose, 125 Volts, 15 Amperes, 2 Pole, 3 Wire, Grounding Type. |
| S3.6 - 1969 | Specifications for Audiometers. |

(Application for copies should be addressed to American National Standards Institute, 1430 Broadway, New York, NY 10018.)

HEARING AID INDUSTRY CONFERENCE (HAIC)

Interim Bone Conduction, Thresholds for Audiometry.

(Application for copies should be addressed to Hearing Aid Industry Conference, 1001 Connecticut Ave., N.W., Washington, DC 20036.)

NATIONAL BUREAU OF STANDARDS (NBS)

Close Coupler Type 9A

(Application for copies should be addressed to Superintendent of Documents, Washington, DC 20025.)



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:41Z
Check the source to verify that this is the current version before use.

UNDERWRITERS LABORATORIES, INC. (UL)
UL 498 Electrical Attachment Plugs and Receptacles

(Application for copies should be addressed to Underwriters Laboratories, Inc. 207 East Ohio Street, Chicago, IL 60611, 333 Pfingsten Rd., Northbrook, IL 60062, 1285 Walt Whitman Rd., Melville, L.I., NY 11746, 1655 Scott Blvd., Santa Clara, CA 95050.)

3. REQUIREMENTS

3.1 Fire and casualty hazards.

3.1.1 The contractor shall submit proof, prior to delivery, that the items to be delivered under the contract conform to the applicable requirements of Underwriters Laboratories (UL) Inc. Standards of Safety. Proof shall consist of one of the following:

A. UL label affixed to the product provided the contractor agrees to have Underwriters Laboratories, Inc. forward a copy of the UL report directly to the Contracting Officer, upon the latter's request.

B. Laboratory report from a nationally recognized independent laboratory. If an independent laboratory is used, it must be adequately equipped and competent to determine compliance with UL standards. The Government reserves the right to require a facilities report and to conduct a facilities survey. Any laboratory report submitted shall include actual test methods, data taken, and results of tests. Detailed findings of examinations necessary to determine compliance shall also be included. The test report shall also have a section that describes the test sample in such a way that the report can be used to insure identity of production items with the test sample. This section shall be supplemented with construction information such as materials used, spacings between electrical components, types of bushings, photographs, wiring diagrams, sketches, and parts list, all appropriately identified. Compliance with this paragraph does not excuse compliance with all other requirements of the contract and specifications.

3.1.2 Compliance with Underwriters' Laboratories, Inc. requirements regarding fire and casualty hazards does not absolve the contractor from complete compliance with the requirements of this specification in order to secure the acceptance of his commodity.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:41Z
Check the source to verify that this is the current version before use.

3.2 Material. Material shall be of the highest quality throughout, durable, and suitable in all respects for the purpose intended.

3.3 Compliance with American National Standards Institute. Audiometer shall be in accordance with the requirements of American National Standards Institute (ANSI) S3.6. Compliance with these requirements does not absolve the contractor from complete compliance with the requirements of this specification in order to secure acceptance of his commodity. In the event of differences between the requirements of this specification and other specifications referenced herein, the requirements of this specification shall govern.

3.4 Construction. The audiometer unit shall be contained in a shock-resisting wood cabinet, covered to give a hard, smooth, walnut grained surface. The cabinet shall have a corrosion-resisting metal control panel, a protectively lined storage compartment, and a dust cover. The cabinet shall provide a fuse holder with a slow-blow type fuse, and phono, tape, microphone, phone, and monitor phone jacks, suitably identified and readily accessible to the operator. Also provided on the cabinet shall be the necessary patchcord connections to extend receivers and patient signal cords from jacks on the audiometer to the audiometric test booth, to enable two room operation. A reusable hinged shipping container of such size to house the audiometer and all accessories shall be furnished with each unit.

3.4.1 Control panel. The control panel shall be located on the front of the cabinet and shall be provided with at least the following: power ON-OFF, input, output, talkback, and tone interrupter switches, a variable attenuator of 5 decibels (db.) steps or less, a self-aligning frequency selector, a loudness balance control, a masking control, an input gain control with power-ON, tone-ON and patient signaling indicating lights and a V.U. meter.

3.4.2 Storage compartment. The protectively lined storage compartment shall contain mountings for securing six replacement fuses and shall be capable of housing all accessories included in the unit.



Source: https://assist.dla.mil -- Downloaded: 202610-21T08:41Z
Check the source to verify that this is the currentversion before use.

3.4.3 Accessories. The following items shall be included in the unit of issue for the audiometer:

Two air conduction receivers with cord.
A bone conduction receiver with two headbands and cord.
Six replacement fuses (slow-blow type).
Patchcord set (to permit use with Audiometer Examination Booth. )
Patient signaling device with cord.
Microphone and stand.
100 Audiogram charts.
Service manuals.
A two prong adapter plug with grounding pigtail.

3.4.4 Receivers. The earphone cups shall be fabricated of insulated type plastic. Each cup shall hold layers of sound insulating material and shall be provided with a large contour, non-liquid, ear pad, manufactured of material which does not take a permanent set. The pad shall completely enclose the ear when used. Each cup shall enclose a dynamic moving coil type speaker.

3.5 Design. The audiometer shall be a fixed frequency oscillator audiometer designed to furnish readings which give an individual's hearing threshold in terms of hearing loss in decibels. The audiometer shall have a zero reference level. The unit shall be capable of performing the following tests, without the use of accessories not currently included as component parts of the unit:

a. Pure tone air conduction
b. Pure tone air conduction with white noise masking
c. Pure tone bone conduction
d. Pure tone bone conduction with masking
e. Either automatic or manual Alternate Binaural
f. Stenger
g. Lombard
h. Audiometer Weber
i. Automatic Pulse Tone

The audiometer unit, with appropriate accessories (not currently components of the unit), shall be capable of performing the following additional tests:

j. Delayed speech (requires a tape recorder)
k. Speech (requires either a microphone, tape, or disk playback)
l. Narrow Band Masking (requires narrow band masking generator)
m. Short Increment Sensitivity Index (SISI) (requires SISI attachment)



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:41Z
Check the source to verify that this is the current version before use.

3.5.1 Attenuation. The audiometer shall be calibrated with 5 db. steps or less. The accuracy of the attenuator shall be within plus or minus 1.5 db. per five db. steps and plus or minus five db. cumulative over the entire intensity range. The audiometer shall be calibrated to the ANSI Standard S3.6.

3.5.2 Wave form. The purity of the tone in the air conduction receiver shall be such that the harmonics at any particular frequency shall be at a level not less than 25 db. below the fundamental tone as measured by acoustic output from receiver with a National Bureau of Standards (NBS) Coupler 9-A or a coupler having equivalent acoustical characteristics. Accessory noises such as line hum, click of interrupting switches shall not in any way interfere with the test tone audible at intensity settings as high as 85 db.

3.5.3 Frequency. The audiometer shall generate test tones of at least the following identified fixed frequencies with a 3 percent accuracy: 125, 250, 500, 1000, 2000, 3000, 4000, 6000, and 8000 hertz for air conduction measurements. The audiometer shall generate test tones of at least the following identified fixed frequencies within a 3 percent accuracy: 125, 250, 500, 1000, 2000, 3000 and 4000 hertz for bone conduction measurements. The frequency selector shall be of the self-aligning type and shall be supplied with a large dial numerically marked to indicate at least the test frequencies listed above.

3.5.4 Air conduction receivers. The test receivers shall be provided with the left and right receivers of different colors or distinctly marked for rapid identification. The two units shall be joined by a suitable adjustable headband that exerts a pressure of 1000 to 2000 grams laterally. Earphones shall be of the dynamic moving coil type and shall be fitted with type MX-41/AR cushions as indicated in ANSI S3.6. The audiometer shall employ individual trimpots for calibration at each frequency. Insulated type audiometric enclosures having layers of sound insulating material and/or sound absorbent blown plastic, manufactured of materials that will not take a permanent set and which completely enclose the ear when in use shall be provided if so specified (see 6.1). The test receivers shall provide for purity of sound production as indicated in ANSI S3.6.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:41Z
Check the source to verify that this is the current version before use.

3.5.5 Bone conduction. The audiometer unit shall be provided with a small, lightweight, hearing aid type bone conduction receiver of not less than 1 square centimeter in size, designed to minimize air conduction radiation so that radiated sound reaching the tympanum through the auditory meatus shall not influence the validity of the bone conduction measurement. The conductor and headband shall be so designed that, when in use, the static force of the conductor shall be adequate to prevent the dynamic force from effecting a physical beat of the conductor against the area under test. The bone conduction receiver shall be calibrated in accordance with the Hearing Aid Industrial Conference (HAIC) Interim Bone Conduction Thresholds for Audiometry.

* 3.5.6 Tone interrupter. The tone interrupting device shall be a lever type switch mounted on the control panel which enables reception of tones which are normally ON - manually OFF, or normally OFF - manually ON. The switch shall activate a "tone ON" indicator light on the panel. The operation of the switch shall produce no intensity overshoot and no extraneous or transient frequencies. Rise/delay times shall not be less than 30 milliseconds nor more than 70 milliseconds, respectively.

* 3.5.7 Microphone speech circuit. A speech circuit shall be provided with each unit and shall have a gain control and V.U. meter, for accurately monitoring a live voice. The circuit shall obtain its speech input from a high quality, stand-mounted microphone with an essentially flat response throughout the bandpass. The selection of such an input shall be accomplished by the input selector switch mounted on the control panel. The speech circuit shall be calibrated at 20 db. sound pressure level (SPL) for 0 db. hearing loss.

3.5.8 Masking device. The unit shall be provided with a masking device to prevent carrying over of sound to the ear opposite the one under test when making bone and air conduction tests. The selector and control for white masking shall be calibrated to cover a range of at least 45 db. to 110 db., but not to exceed 120 db. SPL. The control which regulates the intensity of the masking tone shall indicate steps of 5 db. When used, the masking tone shall be automatically transmitted to the ear opposite the one under test.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:41Z
Check the source to verify that this is the current version before use.

3.5.9 Talk back circuit. An internal amplifier circuit shall be provided to permit the patient in an acoustic testing room to talk to the operator. The control switch for this circuit shall be located on the control panel.

* 3.5.10 Patient signal switch, cord, and light. Each apparatus shall be provided with a patient signal switch, cord, and light. The switch shall be in the form of a pushbutton which shall be conveniently connected to a lamp on the face of the instrument by means of an insulated two-conductor cord not less than 6 feet long. The operation of this assembly shall be such that when the patient pushes the button, the lamp on the face of the machine shall light. The operating voltage of this circuit shall be less than 10 volts.

* 3.5.11 Audiogram cards. One hundred audiogram cards suitable for listing the results of examination shall be furnished with each audiometer. Audiogram cards shall contain as a minimum, provisions for annotating the name, date, test frequency and hearing threshold level (db.) for both the right and left ears at all required frequencies.

* 3.5.12 Electric power source and cord. The audiometer shall be designed to operate on an electric power source of 105-125 volt or 210-250 volt, 50-60 Hertz as specified in the contract or purchase order (see 6.1). The audiometer shall have an internal voltage regulation circuit so designed that when the line voltage varies within the ranges of 105-125 volts or from 210-250 volts, as applicable, the acoustic output level at 60 db. shall not drift more than ± 2.5 db. from its value at the nominal line voltage of 115 or 230 volts, as applicable. The electric power cord shall be a 3-wire grounding type, having a minimum length of 6 feet, fitted with a Hospital Grade male plug conforming to American National Standards Institute (ANSI) Standard C73.11 and Underwriters' Laboratories (UL) Standard 498. A two-prong adapter plug with grounding pigtail for use with two-conductor electrical outlets shall be provided with each unit.

3.6 Identification marking. Each audiometer shall be permanently and legibly marked with at least the following information: Manufacturer's name or registered trademark, model and serial number and electrical ratings (volts, amperes, watts and Hertz). Identification marking shall be secured to unit.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:41Z
Check the source to verify that this is the current version before use.

3.7 Service data. The contractor shall furnish, with each unit, two copies each of a technical manual which contains complete instructions for installation, operation, and maintenance and lists of component parts. The manual shall comply with the requirements of MIL-M-7298. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the Contracting Officer, via the cognizant Government inspector, two preliminary copies of his proposed technical manual for review and comment by the Directorate of Medical Materiel, Defense Personnel Support Center. The preliminary manuals shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract, or purchase order. In addition, the Contractor shall furnish 15 copies of the approved technical manual to the Contracting Officer for distribution. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of the contract and propose to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon request to the Contracting Officer. Waiver of submission of drafts and submission of 15 additional copies shall be at the option of the Government.

3.8 Workmanship. Workmanship shall be first class throughout. The finished audiometer shall be free of defects which detract from the appearance or impair the serviceability of the unit.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:41Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the component or material may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and Table I.

Table I. Sampling for Examination

| For visual examination | Inspection level | AQL (defects per hundred units) |
|---|---|---|
| Major defects | II | 1.0 |
| Total (major and minor combined) | II | 4.0 |

4.2.1 For tests. Sampling for tests shall be conducted in accordance with Table II and MIL-STD-105. The acceptance level shall be zero. Inspection level shall be S-1. Certification establishing proof that testing required in 4.4.6 has successfully been passed with an audiometer of the same design and construction as that being supplied under the current contract shall be considered acceptable providing such test was conducted within two years prior to the date of the contract or purchase order.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:41Z
Check the source to verify that this is the current version before use.

Table II. Tests

| Tests | Para. | Requirements | Para. |
|---|---|---|---|
| Accuracy of tone frequencies | 4.4.1 | Frequency | 3.5.3 |
| Purity of tone | 4.4.2 | Wave form | 3.5.2 |
| Accuracy of attenuator | 4.4.3 | Attenuation | 3.5.1 |
| Accuracy of sound pressure levels | 4.4.4 | Air conduction receivers | 3.5.4 |
| Stability with respect to variations in supply voltage | 4.4.5 | Electric power source and cord | 3.5.12 |
| Sound from test earphones | 4.4.6.1 | Wave form | 3.5.2 |
| Sound from second earphone | 4.4.6.2 | Masking device | 3.5.8 |
| Other unwanted sound | 4.4.6.3 | Bone conduction | 3.5.5 |

4.3 Examination. The audiometer shall be examined to determine compliance with all the requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in Table III. Examination shall not be restricted to the classified possible defects listed below.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:41Z
Check the source to verify that this is the current version before use.

Table III. Classification of Defects

| Categories and defects | |
|---|---|
| **Major** | |
| 101 | Any component part missing, broken, or otherwise unusable. |
| 102 | Exposed edges not free of burs or sharpness. |
| 103 | Accessory storage compartment lining not suitably attached. |
| 104 | Any control dial or switch, loose or does not operate smoothly. |
| 105 | Any indicator light does not operate properly. |
| 106 | Service data missing, not complete, or not as specified. |
| 107 | Headbands do not exert proper lateral pressure. |
| **Minor** | |
| 201 | Identification marking not complete, permanent or legible. |
| 202 | Finish not free of chipped areas, cracks, scratches, or roughness. |
| 203 | Any component part not securely held by mounts when not in use. |

* 4.3.2 Examination for preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of section 5. Defects shall be scored as specified in Table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:41Z
Check the source to verify that this is the current version before use.

Table IV. Examination for preparation for delivery

| Examination | Defect |
|---|---|
| Containers | Not as specified. |
| Quantity | Quantity in interior packages and shipping containers not as specified. |
| Packaging & packing materials | Any component missing or damaged. |
| Workmanship | Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted. |
| Markings | Omitted, incorrect, illegible, improper location, sequence, or method of application. |

4.4 Tests. Tests shall be conducted to determine compliance with the requirements of this specification.

4.4.1 Accuracy of tone frequencies. The accuracy of the tone frequencies generated shall be determined by using a frequency counter which is accurate within ± 3 percent at the frequency recorded.

4.4.2 Purity of tone. The sound pressure levels of the fundamental frequency and its harmonics shall be measured with the earphone coupled to a National Bureau of Standards Coupler 9-A. Measurements shall be made at all frequencies and at the 60 db. hearing loss setting. The difference between the sound pressure levels of the fundamental and its harmonics shall be recorded.

4.4.3 Accuracy of attenuator. Measurement of attenuator accuracy shall be made electrically at the input to the earphone terminated in a dummy load of 100 ohms.

4.4.4 Accuracy of sound pressure levels. The accuracy of sound pressure levels at each frequency shall be determined with the earphone coupled to a National Bureau of Standards 9A Coupler. The accuracy of the sound pressure levels at each frequency shall be measured after one hour of operation. The sound pressure levels obtained shall conform to requirements as indicated in American National Standards Institute (ANSI) 3.6.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:41Z
Check the source to verify that this is the current version before use.

4.4.5 Stability with respect to variations in supply voltage. The audiometer shall be tested over the full range of supply voltages to determine compliance with requirements. The voltage during the test shall be varied from 105 to 125 VAC for 115 volt, 50-60 Hertz operation and 210 to 250 VAC for 230 volt, 50-60 Hertz operation. The audiometer shall meet the accuracy of tone frequencies and purity of tone and also the acoustic output of the audiometer shall meet the accuracy of sound pressure levels and shall not vary through a total range of more than 2.5 db. Measurements shall be made at the 60 db. hearing threshold level setting and the 1000 Hz frequency setting to determine the stability with respect to variations in supply voltage.

4.4.6 Noise during test.

4.4.6.1 Sound from test earphones. The rms weighted sound pressure, due to all frequencies except the signal frequency and its harmonics, shall be determined with the earphone coupled to a National Bureau of Standards 9A Coupler.

4.4.6.2 Sound from second earphone. The sound pressure level in the earphone not being used for test purposes shall be measured to determine compliance with requirements. This test shall be performed by measuring the electrical signal across a dummy load of 100 ohms which simulates the impedance of the earphone.

4.4.6.3 Other unwanted sound. A subject having normal hearing shall be located in the normal test position wearing a pair of disconnected earphones. The electrical output of the audiometer shall be absorbed in a 100 ohm resistor while the test is being performed. The audiometer shall be operated through a normal test cycle at a hearing loss setting of 50 db. The subject shall listen for any sound emanating from the audiometer which could furnish a clue that could influence the hearing test results. As an alternate the test may be performed as indicated in American National Standards Institute (ANSI) S3.6. This test is not intended to apply to sounds from a mechanism as the hand-operated switch, since the subject knows when he presses this switch. In the event of a conflict of test results between the two test procedures, the test results using ANSI S3.6 shall govern.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:41Z
Check the source to verify that this is the current version before use.

## 5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Unit package. Each audiometer shall be packaged in accordance with MIL-L-116, method IA-15 or IC-2 except that a preservative shall not be used.

5.1.2 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.1).

5.2.1 Level A. Each audiometer shall be packed in an exterior container designed for a type 1 load and constructed in accordance with PPP-B-601, overseas type or PPP-B-636, type CF, class weather-resistant, variety DW. Wood boxes shall be lined with fiberboard or heavy weight kraft paper to prevent chafing of the waterproof barrier. Closure, and strapping shall be as specified in the appendix of the applicable box specification.

5.2.2 Level B. Each audiometer shall be packed as specified in 5.2.1 except that the exterior containers shall be domestic types and strapping shall not be required. Closure of fiberboard boxes shall conform to method II as specified in PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Unit package. The unit package shall be marked in accordance with MIL-STD-129. Contract or purchase order number and name of contractor shall be shown.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:41Z
Check the source to verify that this is the current version before use.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

a. Number and date of this specification.
b. Whether insulated type audiometric enclosures are required (see 3.5.4).
c. When accessories, not currently included as component parts of the unit, are required (see 3.5).
d. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
e. Whether 115 or 230 volt, 50 or 60 Hertz operation is required (see 3.5.12).

6.2 The audiometric booth for this purpose is identified as NSN 6515-00-065-9600 Booth, Audiometric Examination. An acceptable commercial booth is Standard Model No. 400-A with Forced Ventilation fabricated by the Industrial Acoustics Corp.

6.3 This specification covers the following item appearing in the Federal Supply Catalog:

| National Stock Number | Item Identification |
|---|---|
| 6515-00-782-2646 | AUDIOMETER, General Diagnostic, 115-230 volt, 50-60 Hz, AC |

6.4 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

6.5 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1966



Source: https://assist.dla.mil -- Downloaded: 202610-21T08:41Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

**INSTRUCTIONS:** The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-A-36495A AUDIOMETER, GENERAL DIAGNOSTIC

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

**DD** FORM 1 JAN 72 **1426**

REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

★ GPO: 1974— 603-680/6126 2-1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:41Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE f
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:41Z
Check the source to verify that this is the current version before use.