MIL-C-36501B
19 August 1976
SUPERSEDING
MIL-C-36501A
18 September 1973

MILITARY SPECIFICATION

CHART, RECORDING AUDIOMETER

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

* 1.1 Scope. This specification covers one type of chart suitable for use with Automatic Audiometers, NSNs 6515-00-584-3057 and 6515-00-533-6644; calibrated to ISO Standard R389/ANSI Standard S3.6.

2. APPLICABLE DOCUMENTS

* 2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of the specification to the extent specified herein:

SPECIFICATIONS

FEDERAL

TT-I-563	- Ink, Writing, for Fountain and Dip Pens.
PPP-B-636	- Boxes, Shipping, Fiberboard.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, Code ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:42Z
Check the source to verify that this is the current version before use.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.

MIL-STD-129 - Marking for Shipment and Storage.

DRAWINGS

Defense Personnel Support Center (DPSC)

24152 - Chart, Recording Audiometer.

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

* 2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

American National Standards Institute (ANSI)

S3.6 Specification for Audiometers.

International Organization for Standardization (Standards) ISO

R389 Standard Reference Zero for the Calibration of Pure-Tone Audiometers.

(Application for copies should be addressed to the American National Standards Institute, 1430 Broadway, New York, NY 10018.)

Joint Committee on Printing, Congress of the United States

Government Paper Specification Standards - Part 2

(Application for copies should be addressed to the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:42Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Materials.

3.1.1 Paper. The paper shall be white tag, 100% bleached sulphate. No fluorescent (optical) brightners shall be added to the pulp or paper during manufacture, but minimal trace amounts are permissible. The paper surface shall be free from lint, fuzz, or any particles which will pick, lift, fluff or pile under normal conditions. Paper acidity shall be not less than a pH of 4.4. The paper weight, 500 sheets, 24 x 36 inches, shall be 125 pounds + 5%. The paper thickness shall be 0.0092 inch + 5%. The paper shall be tested as specified in 4.4.2. When tested as specified in 4.4.3, the paper shall have a good printing and writing surface and shall show no feathering.

3.1.2 Ink. The ink shall be black lithographic ink of either air drying oxidizing type, or flash drying type utilizing a heater for drying. Ink shall bind to paper so that, when thoroughly dry and ready for shipment, there shall be no smudges when the chart is subjected to the rubbing test in 4.4.1.

* 3.2 Design. The audiogram chart shall be lithographed in black ink, cut to size, and have two punched holes. The reverse side of the audiogram chart shall be blank. Design and construction shall be in accordance with DPSC drawing 24152.

3.3 Identification marking. Each chart shall be clearly and legibly marked with the manufacturer's name or registered trademark, and part number if applicable.

3.4 Workmanship. Workmanship shall be first class throughout. Charts shall be cut square and edges shall be parallel with the lines printed thereon. Charts shall be free from defects which detract from their appearance and which may impair their intended use and serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:42Z
Check the source to verify that this is the current version before use.

inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

* 4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

* 4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I. The unit of product for sampling purposes shall be on one chart.

TABLE I. Sampling for examination

	Inspection level	AQL (defect per hundred units)
For visual examination		
Major defects	S-1	1.0
Total defects (Major and Minor combined)	S-1	4.0
For dimensional examination	25 Charts	0

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:42Z
Check the source to verify that this is the current version before use.

* 4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II. The unit of product for sampling purposes shall be one chart. Acceptance number shall be zero.

TABLE II. Sampling for tests

Characteristic	Requirement	Test procedure	AQL (defect per hundred units)
Paper	3.1.1	4.4.2	(See 4.4.2)
Paper Surface	3.1.1	4.4.3	S-1
Ink	3.1.2	4.4.1	S-1

* 4.3 Examination. The chart shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table III. Examination shall be not be restricted to the classified possible defects listed below.

TABLE III. Classification of defects

Category	Defect
Major	
101	Chart values incorrect.
102	Specified printing missing, incomplete, illegible.
Minor	
201	Holes not cleanly punched.
202	Grid line coloring not as specified.
203	Chart not clean.

* 4.3.2 Dimensional examination. The charts shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:42Z
Check the source to verify that this is the current version before use.

4.4 Tests.

4.4.1 Rubbing test. A clean piece of bond paper shall be manually rubbed four times over the selected sample. The bond paper and chart shall then be visually examined for any smudge.

4.4.2 Paper. The paper shall be tested for the applicable characteristics for each lot presented for examination. The characteristics tested, number of determinations, acceptance criteria and applicable test procedures shall be as specified in Government Paper Specification Standards Part 2. For those quality characteristics applicable to the sample unit, all sample units must pass for the lot to be considered acceptable.

4.4.3 Surface. Lines or characters shall be drawn or written on both sides of the chart with a smooth medium-point dip pen using an ink conforming to TT-I-563. To be considered satisfactory, there shall be no feathering of the inked lines or characters.

* 4.5 Packaging inspection. The inspection of the unit, intermediate package, packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

* 5. PACKAGING

5.1 Packaging. Packaging shall be Level A or C as specified (see 6.2).

5.1.1 Unit of issue. One package, containing 1000 charts, as specified, constitutes one unit of issue.

5.1.2 Level A.

* 5.1.2.1 Unit package. One thousand charts shall be packaged in a box conforming to PPP-B-636, type CF, class domestic. Closure shall be as specified in the appendix of the box specification.

5.1.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:42Z
Check the source to verify that this is the current version before use.

5.2 Packing. Packing shall be Level A, B or C as specified (see 6.2).

* 5.2.1 Level A. Four unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure strapping and waterproofing shall be as specified in the appendix of the box specification.

* 5.2.2 Level B. Four unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

* 5.3.1 Unit packages. Each unit package shall be marked in accordance with MIL-STD-129. Contract or purchase order number and name of contractor shall also be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

* 6.1 Intended use. The chart covered by this specification is intended for use with Bekesy-type recording Audiometers, NSNs 6515-00-584-3057 and 6515-00-533-6644. The Model ARJ-4A manufactured by Tracor, Inc., has been furnished under NSN 6515-00-584-3057. The Model ARJ-4B manufactured by Tracor Inc. has been furnished under NSN 6515-00-533-6644. Not all Bekesy type recording audiometers will accept the chart design specified herein.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:42Z
Check the source to verify that this is the current version before use.

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents.

(a) Title, number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6515-00-113-8332	CHART, RECORDING INSTRUMENT, Audiometer

6.4 Changes from previous issue. The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
DSA-DM

Project No. 6515-1999

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:42Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4066 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
hiladelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:42Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*: _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:42Z
Check the source to verify that this is the current version before use.